

## Water Treatment – Daily Checks Form

		Dialysis Centre:			Page No :		
<b>Date:</b>							
S/No	Description of the Checks	Mon	Tue	Wed	Thu	Fri	Sat
1.	HARDNESS TEST (a) Soft Water from the bottom of 5 Micron filter						
2.	CHLORAMINE TEST (a) Soft Water (from bottom of 5 micron filter)						
3.	Inlet Water Pressure						
5.	Flow at Flow Meter (100-1000litres/hr)						
7.	Conductivity (0-30 MicroSiemens)						
8.	Piping Flow (200-2000litres/hr) (PRODUCT) [momentary throughput]						
9.	Any error message displayed?						
10.	Name of the Staff:						

Checked by :

Date:

Hardness and Chloramine testing must be carried out prior to commencement of patient treatment and Bicarbonate preparation.