

Demo dossier — bp monitor

Expected overall result: REJECT

SYNTHETIC TEST DOCUMENT — assembled from the MedDevAudit-IN evaluation fixtures for the purpose of exercising the auditor. Not a real submission. Device names, companies, certificate and licence numbers are fictitious.

PRODUCT SUBMISSION DOSSIER Device: CardioEase CE-9 Automatic Upper Arm Blood Pressure Monitor Manufacturer: Shenzhen Hualing Electronics Co. Ltd, Guangdong, China

1. **PRODUCT DESCRIPTION** The CardioEase CE-9 is a fully automatic upper-arm blood pressure monitor using the oscillometric method. It displays systolic pressure, diastolic pressure and pulse rate on a backlit LCD, stores 99 readings per user, and is powered by 4 x AA alkaline batteries. It is intended for home use by adults.

2. **REGULATORY STATUS** The device is CE marked and is exported to over 30 countries. The company declares that the product complies with all applicable international standards for non-invasive blood pressure monitors. A Declaration of Conformity signed by the Quality Director of Shenzhen Hualing Electronics Co. Ltd is attached at Annex 1 confirming that the device is freely sold in China.

3. **QUALITY SYSTEM** The manufacturing facility operates a quality management system. An ISO 13485 certificate is held by the company. Certificate details are available on request.

4. **SAFETY TESTING** A test report to IEC 60601-1 general electrical safety was issued by Guangdong Testing Institute, report number GDT-2021-44120. Conclusion: pass.

An electromagnetic compatibility assessment to IEC 60601-1-2 was carried out. Emissions were tested to CISPR 11. Immunity was tested at the levels applicable to the professional healthcare facility environment. Report GDT-2021-44121.

5. **PERFORMANCE** The pressure transducer was verified against a reference manometer over the range 0 to 299 mmHg. Measured cuff pressure accuracy is within plus or minus 3 mmHg across the range, meeting IEC 80601-2-30. Calibration of the reference manometer is traceable to a national standard. Test report GDT-2021-44122.

The device incorporates an automatic deflation function that releases cuff pressure at the end of each measurement cycle.

Clinical performance: the oscillometric algorithm is based on a mature platform used in the company's earlier CE-5 and CE-7 models, which have been sold since 2016 without reported accuracy complaints.

6. **ACCESSORIES** Supplied with a standard adult cuff. A large cuff is available separately.

7. **LABEL ARTWORK** CardioEase CE-9 Automatic Blood Pressure Monitor Manufactured by: Shenzhen Hualing Electronics Co. Ltd, Guangdong, China Model: CE-9 DC 6V, 4 x AA Made in China (symbols printed on carton: CE mark, WEEE bin, Type BF, IP22, an umbrella, a factory outline and a small book icon)

8. **USER MANUAL** A user manual is supplied. It contains the operating instructions, how to fit the cuff, how to read the display, the memory function, battery replacement, and a troubleshooting table. Storage conditions are stated on the last page.

9. **PACKING LIST** 1 x main unit 1 x adult cuff 4 x AA batteries 1 x user manual 1 x warranty card