

Submission dossier — digital thermometer

Fixture thermometer-mixed-minor · expected overall result: MINOR_IMPROVEMENT

SYNTHETIC TEST DOCUMENT — assembled from the MedDevAudit-IN evaluation fixtures for the purpose of exercising the auditor. Not a real submission. Device names, companies, certificate and licence numbers are fictitious.

2.1 Import licence. The device is imported under import licence Form MD-15 number IMP/MD/2024/004518, granted by the Central Licensing Authority, CDSCO, on 14 March 2024 and valid until 13 March 2029. A copy of the licence is placed at Annex 2. The licence number is reproduced verbatim on the carton label (Annex 8) and on the Bill of Entry used for customs clearance, so that the three documents can be reconciled at the port.

2.2 Application and authorised agent. The application was made in Form MD-14 dated 02 January 2024 by Sundaram Healthcare Distributors Pvt Ltd, 41 Anna Salai, Chennai 600002, Tamil Nadu, India, holder of wholesale licence 20B/21B No. TN/CH/2019/3312, acting as the Indian authorised agent of Meridian Medical GmbH. The notarised power of attorney dated 18 December 2023 by which Meridian Medical GmbH appoints Sundaram Healthcare Distributors Pvt Ltd for the purposes of Rule 35 is at Annex 3.

2.3 Free Sale Certificate. A Free Sale Certificate for the FS-200 Fingertip Pulse Oximeter was issued by the Bundesinstitut für Arzneimittel und Medizinprodukte (BfArM), the national regulatory authority of Germany, the country of origin, under certificate number 2023/FSC/88214 dated 05 November 2023, confirming that the device is freely sold in Germany. The certificate is valid for two years and is placed at Annex 4 with an apostille.

3.1 Quality management system. The manufacturer holds ISO 13485:2016 certificate number DE-QMS-2019-51120 issued by TÜV SÜD Product Service GmbH for the design and manufacture of non-invasive patient monitoring equipment at Industriestrasse 12, Fürstentfeldbruck. Date of issue 12 April 2019; valid until 11 April 2022. Recertification audit has been scheduled and the renewed certificate will be forwarded to the licensing authority upon receipt.

4.1 Risk management. Risk management has been carried out in accordance with ISO 14971:2019 across the device lifecycle. The risk management file at Annex 6 comprises the risk management plan RMP-FS200-01, the hazard identification and risk analysis worksheet (design FMEA, 64 identified hazards), the risk evaluation and risk control table with verification of effectiveness for each control, the residual risk evaluation, and the signed risk management report RMR-FS200-03, which concludes that the overall residual risk is acceptable in relation to the clinical benefit.

5.1 Electrical safety. Conformity with IEC 60601-1:2005 + A1:2012 + A2:2020 (general requirements for basic safety and essential performance) is demonstrated by test report number TR-60601-1-22884 issued by TÜV SÜD Product Service GmbH, an ILAC-MRA accredited test laboratory, dated 19 September 2023. The applied part is classified Type BF. The report covers dielectric strength, leakage currents in normal and single-fault condition, mechanical strength and marking, and concludes PASS on all applicable clauses. Full report at Annex 7.

5.2 Electromagnetic compatibility. Conformity with IEC 60601-1-2:2014 + A1:2020 is demonstrated by EMC test report number TR-EMC-22891 (TÜV SÜD Product Service GmbH, 22 September 2023). Emissions were tested to CISPR 11 Group 1 Class B. Immunity was tested at the levels specified for the home healthcare environment, including ESD ± 8 kV contact / ± 15 kV air, radiated RF 10 V/m, and proximity fields from wireless communications equipment. The declared electromagnetic environment is home healthcare. The manufacturer's EMC declaration tables are reproduced in section 8 of the Instructions for Use. Report at Annex 7.

8.1 Label — manufacturer. The device label and carton artwork (Annex 8) carry the manufacturer symbol followed by "Manufactured by: Meridian Medical GmbH, Germany" and, beneath it, "Distributed in India by Sundaram Healthcare Distributors Pvt Ltd, 41 Anna Salai, Chennai 600002". The full postal

address of the manufacturer is stated in the Instructions for Use.

8.2 Label — importer and licence. The carton artwork (Annex 8) carries, in addition to the manufacturer block: "Imported and marketed by: Sundaram Healthcare Distributors Pvt Ltd, 41 Anna Salai, Chennai 600002, India" and "Import licence number: IMP/MD/2024/004518". The licence number matches the MD-15 licence at Annex 2.

8.3 Label — language. The FS-200 is intended for home use. The carton artwork (Annex 8) carries the statement "For Sale in India" in English with the Hindi (Devanagari) rendering "इसका उपयोग भारत में ही करना चाहिए" and the essential particulars — device name, intended use, manufacturer and importer — are presented bilingually in Hindi and English in the same panel.

8.4 Label — symbols. Symbols used on the artwork conform to ISO 15223-1:2021. The symbol glossary reproduced in section 1 of the Instructions for Use lists each symbol with its meaning: manufacturer, date of manufacture, batch code (LOT), serial number (SN), catalogue number (REF), consult instructions for use, Type BF applied part, temperature and humidity limits, and "keep dry". Annex 8 marks the position of each symbol on the device label and carton.

9. Instructions for Use. The IFU (IFU-FS200-EN rev 5, Annex 9) is supplied in English and contains the intended use and intended user, operating instructions, contraindications, warnings and precautions, residual risks, cleaning and maintenance instructions, storage and transport conditions and the expected service life. Adverse events may be reported to the manufacturer at vigilance@meridian-medical.de or to the competent authority in the user's country.

1. Device Master File — index. This Device Master File is submitted in the format of the Fourth Schedule, Part II. Table of contents: 1 executive summary; 2 device description and intended use; 3 manufacturing process and site details; 4 list of applied standards; 5 design verification and validation data; 6 clinical accuracy study; 7 stability and shelf-life data; 8 labelling and packaging set; 9 Instructions for Use; 10 post-market surveillance plan and vigilance procedure. Each numbered section is present in this bundle and every annex referenced in the sections is attached (Annexes 1–12).

6.1 Performance. Test report TR-80601-2-56-31177 (SGS Fimko, Helsinki, ILAC-accredited, 8 May 2023) to IEC 80601-2-56:2017+A1 covers the DT-10 clinical thermometer. Laboratory accuracy in a stirred water bath was ± 0.1 °C across the measurement range 35.0 to 42.0 °C and ± 0.2 °C from 32.0 to 34.9 °C and 42.1 to 43.0 °C. Response time to within 0.1 °C of final reading: 8 s (predictive mode). Clinical accuracy study (n = 112, axillary site, reference: calibrated mercury-in-glass): clinical bias " 0.04 °C, limits of agreement " 0.28 to +0.20 °C, clinical repeatability 0.09 °C. Report at Annex 11.

8.5 Label and IFU — measurement site. The DT-10 is a digital thermometer for measuring body temperature and may be used orally, axillary or rectally, as described in the Instructions for Use. Measurement accuracy is ± 0.1 °C. Mode of operation and the reference body site are as per IEC 80601-2-56. The label carries the wording "Digital clinical thermometer — body temperature".

5.3 Internal power source. The DT-10 is internally powered by one LR41 alkaline button cell, 1.5 V, 32 mAh. The cell manufacturer's IEC 62133-2:2017 certificate (Energizer, certificate CB-DK-15221, Annex 7B) is enclosed. Section 15.4.3 of the IEC 60601-1 report (TR-60601-1-31160) assesses the internal electrical power source, including protection against excessive temperature and electrolyte leakage. Because the cell is a button cell and the device may be used on children, the IFU carries the button-cell ingestion warning in section 2.

7.1 Biocompatibility. The probe tip (stainless steel 304) and the single-use polyethylene probe cover are surface-contacting devices with limited ("d24 h) contact duration. The biological evaluation plan BEP-DT10-01 and biological evaluation report BER-DT10-02 (Annex 12) are per ISO 10993-1:2018 and include cytotoxicity (ISO 10993-5, MTT, L929: non-cytotoxic), skin irritation (ISO 10993-23, reconstructed human epidermis: non-irritant) and skin sensitisation (ISO 10993-10, GPMT: non-sensitiser), performed by Nelson Labs, report numbers 1240221-1 to -3.