

NeuroPulse AI: Electromyography (EMG) Profile

Yellow = cleaned signal | Green = muscle strength | Red = muscle active shadow


Patient: Patient Name | **ID:** NP-20260609-6079 | **Date:** 09-06-2026 13:06

Avg Strength 0.0	Active Time 0.0%	Major Spikes 0	Signal Quality Idle / No contact
---------------------	---------------------	-------------------	-------------------------------------

1. Clinical Summary

Session Score	0/100
Symmetry Score	0.0%
Side Deficit	100.0% (Right lower)
Lowest Activity Channel	Channel 1
Fatigue Drop	Low (0.0% decline)
Signal Quality	Idle / No contact

2. 30-Second Graph Snapshot



3. Muscle-wise Quantitative Summary

Muscle / Channel	Avg	Peak	Active %	Quality
Channel 1	0.0	0.0	0.0%	No contact
Channel 2	0.0	0.0	0.0%	No contact
Channel 3	0.0	0.0	0.0%	No contact
Channel 4	0.0	0.0	0.0%	No contact

4. Local Data-Based EMG Findings

Signal Reasoning (Quantitative Text Analysis):

Awaiting live EMG data...

Recorded Session Findings:

- * Recording summary: 1 EMG samples captured over approximately 60 seconds.
- * Session Score: 0/100. Signal Quality: Idle / No contact.
- * Electrode/sensor contact was not strong enough for reliable EMG interpretation.
- * Recorded values are treated as idle baseline/noise, so side deficit, weakest muscle, fatigue, and symmetry findings are suppressed.
- * Action required: attach electrodes with correct reference/ground, wait for the signal to settle, then press CLEAR before recording.
- * Clinical note: Affected side was not selected; interpretation is limited to channel activity and symmetry. This is a non-diagnostic rehab monitoring prototype that provides quantitative EMG feedback for physiotherapy support.

5. Clinical Observations & Notes

Summary Notes: N/A

Comparative Analysis:

- Left Side Avg: 0.0
- Right Side Avg: 0.0
- Symmetry Ratio (L/R): 0.0%
- Most Dynamic Channel: Channel 1