

NeuroPulseAI ECG Analysis Report

Powered by NeuroPulseAI

Patient: Aditya Kumar Singh Age: N/A Blood Group: N/A

Generated: 23-04-2026 13:08:58

Analysis Metrics

Duration (sec): 60.0
Sampling Rate (Hz): 250.0
Mean BPM: 114.55
Min BPM: 47.47
Max BPM: 127.12
SDNN (ms): 129.55
R-Peak Count: 105
Signal Quality: Good

Observations

- Recording duration 60.0 sec at approx 250.0 Hz sampling.
- Detected 105 R-peaks with signal quality marked as good.
- Average BPM 114.5, range 47.5 to 127.1.
- Estimated short-term variability (SDNN) 129.6 ms.

Diagnostics

- Signal morphology is stable enough for engineering-level ECG review.
- Detected heart rate is above 100 BPM in this window.
- Short-term variability appears high or peak detection may be unstable.
- These are engineering diagnostics, not a clinical diagnosis.

AI Analyst Summary

Recording duration 60.0 sec at approx 250.0 Hz sampling. Detected 105 R-peaks with signal quality marked as good. Average BPM 114.5, range 47.5 to 127.1. Estimated short-term variability (SDNN) 129.6 ms. Signal morphology is stable enough for engineering-level ECG review. Detected heart rate is above 100 BPM in this window. Short-term variability appears high or peak detection may be unstable. These are engineering diagnostics, not a clinical diagnosis. AI Analyst note: this report supports engineering interpretation and is not a certified medical diagnosis.

Best ECG Graph Snapshot

