

Signal validation: sensor reading vs lab sample

Equipment tag:	KO-3202
Instrument:	Water-in-oil sensor WIO-TEST-02
Date:	14-Oct-2026
Sensor reading:	410 ppm
Lab sample:	425 ppm (within tolerance)
Signed:	Lab analyst C (signature on file)

Validation of the KO-3202 sensor.

SYNTHETIC TEST DOCUMENT