

GxP IT SYSTEM AUDIT EVIDENCE DOSSIER

System: **Validated Laboratory Information Management System (SYS-LIMS-001)** | Status: **PRE-OPERATIONAL**
Release Decision: **HOLD / DEFER - DO NOT RELEASE** | Readiness: **82%**

1. Executive Summary & Verification Posture

This dossier summarizes the compliance evaluation of Validated Laboratory Information Management System. Deterministic readiness index: **82%**. System release is currently **HELD / DEFERRED** due to incomplete intended-use verification (OV/PfV/UAT not performed), unrated residual risks across 49 working high requirements, and unsatisfied lifecycle gates G5 and G6.

2. Lifecycle Phase Gate Status (G1 - G6)

Gate	Name	Status	Prerequisites / Blocking Reason
G1	Project Inception & Regulatory Scope	MET	Satisfied
G2	Requirements Baseline Approval	MET	Satisfied
G3	Detailed Design & Specification Freeze	MET	Satisfied
G4	Technical Verification & IQ	MET	Technical interfaces and IQ passed; intended-use testing pending.
G5	Release Readiness & Residual Risk Acceptance	NOT_MET	Blocked by missing intended-use verification (OV/PfV/UAT not performed), unrated residual risks across 49 working high requirements, and deferred VSR.
G6	Operational Handover & SLA Activation	NOT_MET	Blocked by incomplete shopfloor training (0 of 250 operators qualified) and unactivated SLA.

3. Active Compliance Gaps & Release Blockers

- [HIGH]** QA Approval Missing from URS Baseline: Operating system requirement baseline without Quality Unit authorization violates 21 CFR Part 11 and EU Annex 11.
- [MEDIUM]** Periodic Review Required: SOP_Change_Control.docx: SOP review date is due in less than 30 days.
- [MEDIUM]** Data Archival Policy Ambiguity: Retention period for electronic records lacks explicit ALCOA+ enduring backup duration.

4. Cryptographic SHA-256 Audit Trail Verification

Timestamp (UTC)	Actor / Agent	Action	SHA-256 Hash
2026-09-04 08:15:30	USER: None	Human Approved GxP Workflow & Executed ServiceNow Task	301ad0a2de...829d56
2026-09-04 08:15:30	AGENT: audit_engine	AUDIT_CHECKLIST_EXECUTED	2c35dff522...40054a
2026-09-04 08:15:26	USER: supervisor_agent	RAG_QUERY	274b37e811...522f82
2026-09-03 19:32:15	SYSTEM: continuous_compliance_monitor	Triggered Periodic Review Expiration Event	f8d8176173...c775da
2026-09-03 19:32:15	AGENT: audit_evidence_agent	EVIDENCE_PACK_GENERATED	2ca34f4406...dd3d33
2026-09-03 19:32:15	USER: None	Human Approved GxP Workflow & Executed ServiceNow Task	1d1f800a03...134c8b
2026-09-03 19:32:15	AGENT: audit_engine	AUDIT_CHECKLIST_EXECUTED	1a070014dc...0dcac2

2026-09-03 19:32:11	USER: supervisor_agent	RAG_QUERY	33836d39e5...66babcb
2026-09-03 19:32:05	AGENT: audit_engine	AUDIT_CHECKLIST_EXECUTED	5481a2463c...1d89ca
2026-09-03 19:31:17	AGENT: audit_engine	AUDIT_CHECKLIST_EXECUTED	2e0d4412b2...26eb00