

GxP IT SYSTEM AUDIT EVIDENCE DOSSIER

System: **Validated Laboratory Information Management System (SYS-LIMS-001)** | Status: **PRE-OPERATIONAL**
Release Decision: **HOLD / DEFER - DO NOT RELEASE** | Readiness: **82%**

1. Executive Summary & Verification Posture

This dossier summarizes the compliance evaluation of Validated Laboratory Information Management System. Deterministic readiness index: **82%**. System release is currently **HELD / DEFERRED** due to incomplete intended-use verification (OV/PfV/UAT not performed), unrated residual risks across 49 working high requirements, and unsatisfied lifecycle gates G5 and G6.

2. Lifecycle Phase Gate Status (G1 - G6)

Gate	Name	Status	Prerequisites / Blocking Reason
G1	Project Inception & Regulatory Scope	MET	Satisfied
G2	Requirements Baseline Approval	MET	Satisfied
G3	Detailed Design & Specification Freeze	MET	Satisfied
G4	Technical Verification & IQ	MET	Technical interfaces and IQ passed; intended-use testing pending.
G5	Release Readiness & Residual Risk Acceptance	NOT_MET	Blocked by missing intended-use verification (OV/PfV/UAT not performed), unrated residual risks across 49 working high requirements, and deferred VSR.
G6	Operational Handover & SLA Activation	NOT_MET	Blocked by incomplete shopfloor training (0 of 250 operators qualified) and unactivated SLA.

3. Active Compliance Gaps & Release Blockers

- [HIGH]** QA Approval Missing from URS Baseline: Operating system requirement baseline without Quality Unit authorization violates 21 CFR Part 11 and EU Annex 11.
- [MEDIUM]** Periodic Review Required: `SOP_Change_Control.docx`: SOP review date is due in less than 30 days.
- [MEDIUM]** Data Archival Policy Ambiguity: Retention period for electronic records lacks explicit ALCOA+ enduring backup duration.

4. Cryptographic SHA-256 Audit Trail Verification

Timestamp (UTC)	Actor / Agent	Action	SHA-256 Hash
2026-09-03 18:31:45	USER: None	Human Approved GxP Workflow & Executed ServiceNow Task	9f9f4841f5...e259c9
2026-09-03 18:31:43	USER: supervisor_agent	RAG_QUERY	f66e0ef22a...1abfb8
2026-09-03 18:30:23	SYSTEM: continuous_compliance_monitor	Triggered Periodic Review Expiration Event	0f2be803d6...e2a566
2026-09-03 18:30:23	AGENT: audit_evidence_agent	EVIDENCE_PACK_GENERATED	6af54dc69b...4aecc7
2026-09-03 18:30:23	USER: None	Human Approved GxP Workflow & Executed ServiceNow Task	7c520047f9...59dd73
2026-09-03 18:30:20	USER: supervisor_agent	RAG_QUERY	7b561fccdb...3ca5d9
2026-09-03 18:29:44	SYSTEM: continuous_compliance_monitor	Triggered Periodic Review Expiration Event	f499132f27...008a1a

2026-09-03 18:29:44	AGENT: audit_evidence_agent	EVIDENCE_PACK_GENERATED	7f65e32b69...e6e13d
2026-09-03 18:29:43	USER: None	Human Approved GxP Workflow & Executed ServiceNow Task	37dab02703...87bd5f
2026-09-03 18:29:40	USER: supervisor_agent	RAG_QUERY	cd999b8b0f...6b0755