



DISTRIBUTED BY :

PT. BERITO JAYA MEDIKA

Jl. Achmad Jais No.34 Ruko Grand Achmad
Jais Blok C-2, Peneleh, Kec. Genteng,
Surabaya, Jawa Timur 60274

KEMENKES RI AKD 20502620011

TOO CARE PATIENT MONITOR 5 PARAMETER

TCPM-05



40%
TKDN

Tingkat Komponen Dalam Negeri



DETAIL LAYAR 1



DETAIL LAYAR 2



DETAIL LAYAR 3



DETAIL LAYAR 4

TOO CARE PATIENT MONITOR 5 PARAMETER



DESCRIPTION

EKG

- Mode utama : 5 Leads (R, L, F, N, C or RA, LA, LL, RL, V)
- Pemilihan lead : I, II, III, avR, avL, avF, V
- Bentuk gelombang : 2 ch
- Mode lead : 3 Leads (R, L, F or RA, LA, LL)
- Pemilihan lead : I, II, III
- Penguatan : x 2.5mm/mV, x 5.0mm/mV, x 10mm/mV, x 20mm/mV, otomatis
- Interval pengukuran : Neo/Ped 15 ~ 350 bpm
- Akurasi : $\pm 1\%$ or ± 1 bpm, which great
- Resolusi : 1 bpm
- Sensivitas : > 200 (μ V P.P)
- Jenis : ASYSTOLE, VFIB/VTAC, COUPLET, IGENINY, TRIGEMINY, R ON T, VT > 2 , PVC, TACHY, BRADY, MISSED BEATS, PNP, PNC
- Impendensi Input Diferensial : > 5 MO
- Monitor : > 105 dB
- Operation : > 105 dB
- Diagnosa : > 85 dB
- Potensi Offset : $+ 300$ mV
- Elektrada
- Arus kebocoran : < 10 uA
- Pemulihan dasar : < 3 S After Defi
- Rentang Sinyal EKG : $+ 8$ m V (Vp-p)
- Pembedahan : 1 ~ 15 Hz
- Monitor : 0.5 ~ 35 Hz
- Diagnostik : 0.05 ~ 100 Hz
- Sinyal Kalibrasi : 1 (mV p-p) Accuracy : $\pm 5\%$
- Pengukuran dari Alarm : -20 ~ $+ mV$
- Jenis : ASYSTOLE VAB/VTAC COUPLE IGENINY TRIGEMINY, RON TVT > 2 PVC, TACHY, BRADY MISSED BEATS, PNP, PNC
- Alarm : Tersedia
- Review : Tersedia

PEMULIHAN

- Metode : Impedance between R-F (RA-LL)
- Impedansi Input Diferensial : > 2.5 M Ω
- Mengukur Rentang Impedansi : 03 ~ 5.00
- Rentang Impedansi garis dasar : 0 ~ 2.5 K Ω
- Bandwidth: 0.3 ~ 2.5 Hz
- Tingkat pernafasan : Rentang Pengukuran dan Alarm Dewasa : 0 ~ 120 rpm Nea / Ped : 0 ~ 150 rpm
- Resolution : 1 rpm
- Akurasi : ± 2 rpm
- Alarm Apean : 10 ~ 40 S

NIBP

- Metode : Oscillometric
- Mode : Manual, auto
- Interval pengukuran dalam mode AUTO : 1, 2, 3, 4, 5, 10, 15, 30, 60, 90, 120, 180, 240, 480 (min)
- Jenis alarm : SYS, DIA, MEAN
- rentang denyut nadi : 40 ~ 240 bpm
- Rentang pengukuran dan alarm : Mode Dewasa: SYS : 40 ~ 270 mmHg DIA : 10 ~ 215 mmHg MEAN : 20 ~ 235 mmHg Mode Anak: SYS : 40 ~ 200 mmHg DIA : 10 ~ 150 mmHg MEAN : 20 ~ 165 mmHg Neonatal Mode: SYS : 40 ~ 135 mmHg DIA : 10 ~ 100 mmHg MEAN : 20 ~ 110 mmHg
- Tekanan : 1mmHg
- Kesalahan Rata-rata Maksimum : $+ 5$ mmHg
- Simpangan baku maksimum : $+ 8$ mmHg
- Bentuk Gelombang : 2 ch
- Mode Dewasa : 297 ± 3 mmHg
- Mode Anak : 240 ± 3 mmHg
- Mode Neonatal : 147 ± 3 mmHg

SP02

- Rentang Pengukuran : 0 ~ 100 %
- Rentang Alarm : 0 ~ 100 %
- Resolusi : 1 %
- Akurasi : 70% ~ 100% ± 2 % 0% ~ 69% tidak ditentukan
- Penundaan Alarm : 10 sec
- Rentang Pengukuran dan Alarm : 0 ~ 254bpm
- Resolusi : 1bpm
- Akurasi : ± 2 bpm

TEMPERATURE

- Saluran : 1
- Rentang pengukuran dan alarm : 0 ~ 50 °C
- Resolusi : 0.1 °C
- Akurasi : ± 0.1 °C
- Interval aktualisasi : about 1 Sec
- Konstanta Waktu Rata-rata : < 10 Sec



+62 813-3638-5959



@beritojayamedika



beritojayamedika@gmail.com