CERTIFICATE OF CALIBRATION

AEROCENTER EPPS ATLANTA

Aero Center Epps Atlanta
1 Aviation Way
Atlanta, Georgia, 30341, United States
7704589851

Record Information

Certificate No: 5568-1
GageList ID: 5568
Instrument ID: A-1002
Serial No: T-56384
Asset No:
Company: Advanced AOG

Gage Information

Gage Status - As Found: Active
Model: TTL-300-ATG-1
Manufacturer: Tool Testing Lab
Date Acquired: 01/29/2026
Procedure No: 04.01

Description: Pressure Gauge 0-300 psi
Reference Standard: No
Shared Tool:
Customer Code:

Calibration Information

Interval: Scheduled (1 years, 0 months, 0 days)
Temperature: 66 F
Calibration Type: INITIAL

Humidity: 20%

Gage Assignment

Tool Location:

Assignee:

Calibration performed under the following environmental conditions:

Test No	Attribute	Nominal | Minimum | Maximum	As Found	As Left	Result
1	PRESSURE	0 | -5 | 5 Pounds per Square Inch (psi)	0	0	Pass
2	PRESSURE	30 | 25 | 35 Pounds per Square Inch (psi)	29	29	Pass
3	PRESSURE	60 | 55 | 65 Pounds per Square Inch (psi)	60	60	Pass
4	PRESSURE	90 | 85 | 95 Pounds per Square Inch (psi)	90	90	Pass
5	PRESSURE	120 | 115 | 125 Pounds per Square Inch (psi)	120	120	Pass
6	PRESSURE	150 | 145 | 155 Pounds per Square Inch (psi)	152	152	Pass
7	PRESSURE	180 | 175 | 185 Pounds per Square Inch (psi)	182	182	Pass
8	PRESSURE	210 | 205 | 215 Pounds per Square Inch (psi)	212	212	Pass
9	PRESSURE	240 | 235 | 245 Pounds per Square Inch (psi)	242	242	Pass
10	PRESSURE	270 | 265 | 275 Pounds per Square Inch (psi)	272	272	Pass
11	PRESSURE	300 | 295 | 305 Pounds per Square Inch (psi)	303	303	Pass

This Gage is within Specifications: (As Found) Pass - (As Left (overall pass/fail)) Pass

Gage is Fit for Intended Purpose: (As Left) Yes

CERTIFICATE OF CALIBRATION

Calibration Actions

Adjustments Required: No
Repairs Required: No

Action Required: No

Statement: The instrument on this certification has been calibrated against standards traceable to the National Institute of Standards and Technology (NIST) or other recognized national metrology institutes, derived from ratio type measurements, or compared to nationally or internationally recognized consensus standards. A test uncertainty ratio (T.U.R) of 4:1 [K=2, approx. 95% Confidence Level] was maintained unless otherwise stated. Lab Operations follow the guidelines of ANSI/NCSL Z540-1-1994, ISO/IEC/7025 NIST Handbook 150:2006, DAP-TM08 Rev 4.0 & NIST Tech Note 1297, 1994 All results contained within this certification relate only to item(s) calibrated. Any number of factors may cause the calibration item to drift out of calibration before the instrument's calibration interval has expired. This certificate shall not be reproduced except in full, without written consent of Aero Center Epps Atlanta Calibration Department

Review and Approval

Calibrated By: A MEREDITH

Calibration Date: 02/03/2026

Next Calibration Due: 02/03/2027

Gage Status - As Left: Active

Finalized By: A MEREDITH

Date Finalized: 02/03/2026

Reference Standard Information

Instrument ID	Description	Manufacturer	Model	Last Calibration	Next Calibration Due
N-PRESSURE GAUGE 4 (0-600 psi)	Pressure Gauge 0-600 psi	ASHCROFT INC.	45 1082PS 04L XC4	06/18/2025	06/18/2026
N-TEMP/HUM RECORDER	Temp/Humidity Recorder	Extech	RH520	07/17/2025	07/17/2026