CERTIFICATE OF CALIBRATION

AERO CENTER
EPPS ATLANTA

Aero Center Epps Atlanta
1 Aviation Way
Atlanta, Georgia, 30341, United States
7704589851

Record Information

Certificate No: 5567-1
GageList ID: 5567
Instrument ID: A-1004
Serial No: 0125114320
Asset No:
Company: Advanced AOG

Gage Information

Gage Status - As Found: Active
Model: 2502MRPH
Manufacturer: CDI Torque Products
Date Acquired: 01/29/2026
Procedure No: 02.01
Description: Torque Wrench 30-250 inlb
Reference Standard: No
Shared Tool:
Customer Code:

Calibration Information

Interval: Scheduled (1 years, 0 months, 0 days)
Temperature: 66 F
Humidity: 20%
Calibration Type: INITIAL

Gage Assignment

Tool Location:
Assignee:

Calibration performed under the following environmental conditions:

Test No	Attribute	Nominal | Minimum | Maximum	As Found	As Left	Result
1	Clockwise	74 | 71.04 | 76.96 In Lb	74.95	72.58	Pass
2	Clockwise	118 | 113.28 | 122.72 In Lb	121.65	118.51	Pass
3	Clockwise	162 | 152.52 | 168.48 In Lb	168.44	164.89	Pass
4	Clockwise	206 | 197.76 | 214.24 In Lb	216.13	212.48	Pass
5	Clockwise	250 | 240 | 260 In Lb	261.59	258.40	Pass

This Gage is within Specifications: (As Found) Fail - (As Left (overall pass/fail)) Pass
Gage is Fit for Intended Purpose: (As Left) Yes
Calibration Actions
Adjustments Required: Yes
Repairs Required: No
Action Required: No

Statement: The instrument on this certification has been calibrated against standards traceable to the National Institute of Standards and Technology (NIST) or other recognized national metrology institutes, derived from ratio type measurements, or compared to nationally or internationally recognized consensus standards. A test uncertainty ratio (T.U.R) of 4:1 [K=2, approx. 95% Confidence Level] was maintained unless otherwise stated. Lab Operations follow the guidelines of ANSI/NCSL Z540-1-1994, ISO/IEC/7025 NIST Handbook 150:2006, DAP-TM08 Rev 4.0 & NIST Tech Note 1297, 1994 All results contained within this certification relate only to item(s) calibrated. Any number of factors may cause the calibration item to drift out of calibration before the instrument's calibration interval has expired. This certificate shall not be reproduced except in full, without written consent of Aero Center Epps Atlanta Calibration Department

Review and Approval

Calibrated By: A MEREDITH
Calibration Date: 02/03/2026
Next Calibration Due: 02/03/2027
Gage Status - As Left: Active
Additional Comments: Clockwise ONLY

CERTIFICATE OF CALIBRATION

Finalized By: A MEREDITH

Date Finalized: 02/03/2026

Reference Standard Information

Instrument ID	Description	Manufacturer	Model	Last Calibration	Next Calibration Due
N-CDI-TORQUE SYSTEM	Torque Tester	CDI Torque	2000-400-02	03/06/2025	03/06/2026
N-TEMP/HUM RECORDER	Temp/Humidity Recorder	Extech	RH520	07/17/2025	07/17/2026

© 2026 GageList | All rights reserved | GageList.com | 281-816-7130