CERTIFICATE OF CALIBRATION

AEROCENTER EPPS ATLANTA

Aero Center Epps Atlanta
1 Aviation Way
Atlanta, Georgia, 30341, United States
7704589851

Record Information

Certificate No: 3904-4
GageList ID: 3904
Instrument ID: A-1007
Serial No: 23520
Asset No:
Company: Advanced AOG

Gage Information

Gage Status - As Found: Active
Model: UNKNOWN
Manufacturer: Dill
Date Acquired: 09/12/2022
Procedure No: 04.01
Description: Pressure Gauge 0-3000 psi
Reference Standard: No
Shared Tool:
Customer Code: AA50

Calibration Information

Interval: Scheduled (1 years, 0 months, 0 days)
Temperature: 66 F
Calibration Type: SCHEDULED
Humidity: 20%

Gage Assignment

Tool Location:
Assignee:

Calibration performed under the following environmental conditions:

Test No	Attribute	Nominal | Minimum | Maximum	As Found	As Left	Result
1	PRESSURE	0 | -100 | 100 Pounds per Square Inch (psi)	0	0	Pass
2	PRESSURE	300 | 200 | 400 Pounds per Square Inch (psi)	320	320	Pass
3	PRESSURE	600 | 500 | 700 Pounds per Square Inch (psi)	610	610	Pass
4	PRESSURE	900 | 800 | 1000 Pounds per Square Inch (psi)	890	890	Pass
5	PRESSURE	1200 | 1100 | 1300 Pounds per Square Inch (psi)	1180	1180	Pass
6	PRESSURE	1500 | 1400 | 1600 Pounds per Square Inch (psi)	1480	1480	Pass
7	PRESSURE	1800 | 1700 | 1900 Pounds per Square Inch (psi)	1780	1780	Pass
8	PRESSURE	2100 | 2000 | 2200 Pounds per Square Inch (psi)	2080	2080	Pass
9	PRESSURE	2400 | 2300 | 2500 Pounds per Square Inch (psi)	2400	2400	Pass
10	PRESSURE	2700 | 2600 | 2800 Pounds per Square Inch (psi)	2710	2710	Pass
11	PRESSURE	3000 | 2900 | 3100 Pounds per Square Inch (psi)	3000	3000	Pass

This Gage is within Specifications: (As Found) Pass - (As Left (overall pass/fail)) Pass
Gage is Fit for Intended Purpose: (As Left) Yes

© 2026 GageList | All rights reserved | GageList.com | 281-816-7130

CERTIFICATE OF CALIBRATION

Calibration Actions

Adjustments Required: No
Repairs Required: No

Action Required: No

Statement: The instrument on this certification has been calibrated against standards traceable to the National Institute of Standards and Technology (NIST) or other recognized national metrology institutes, derived from ratio type measurements, or compared to nationally or internationally recognized consensus standards. A test uncertainty ratio (T.U.R) of 4:1 [K=2, approx. 95% Confidence Level] was maintained unless otherwise stated. Lab Operations follow the guidelines of ANSI/NCSL Z540-1-1994, ISO/IEC/7025 NIST Handbook 150:2006, DAP-TM08 Rev 4.0 & NIST Tech Note 1297, 1994 All results contained within this certification relate only to item(s) calibrated. Any number of factors may cause the calibration item to drift out of calibration before the instrument's calibration interval has expired. This certificate shall not be reproduced except in full, without written consent of Aero Center Epps Atlanta Calibration Department

Review and Approval

Calibrated By: A MEREDITH

Calibration Date: 02/03/2026

Next Calibration Due: 02/03/2027

Gage Status - As Left: Active

Finalized By: A MEREDITH

Date Finalized: 02/03/2026

Reference Standard Information

Instrument ID	Description	Manufacturer	Model	Last Calibration	Next Calibration Due
N-PRESSURE GAUGE 5 (0-5000 psi)	Pressure Gauge 0 - 5000 psi	ASHCROFT INC.	45 1082PS 04L XC4	02/10/2025	02/10/2026
N-PRESSURE GAUGE 5 (0-5000 psi)	Pressure Gauge 0 - 5000 psi	ASHCROFT INC.	45 1082PS 04L XC4	02/10/2025	02/10/2026