

Design Portfolio

Agoston Walter

April 27, 2020

Measurect

November 2019 – April 2020

University Final Project for Biomedical Engineering – Group Project

Descriptive Paragraph:


Male spinal cord injury (SCI) patients have erectile dysfunction due to their sustained injuries, which worsens their quality of life. Current therapies such as Viagra do not solve the problem of erectile dysfunction in male SCI patients, spurring researchers at Duke and across the United States to look for other cures. A major setback for the researchers is that there is no universally accepted device to measure the stiffness of a penis during erection, the erection metric that is critical to successful coitus. Dr. Casey Steadman, a researcher at Duke, reached out to my Biomedical engineering design group with this design problem: to create a device that can accurately measure the stiffness of the penis over a period of two hours. With this device, Dr. Steadman can run her study on new therapies for the male SCI patients and have a robust, accurate tool to gauge the effectiveness of her treatments. There currently exists a variety of methods to measure erectile function, but all of them have significant issues such as a lack of evidence linking the data the sensor is measuring to penile stiffness, or interference with the actual erection. Our group performed extensive background research into this design problem, and we found that using ultrasound to measure the stiffness of the tissue is the optimal solution. The professor for the design class, Dr. Mark Palmeri, specializes in ultrasound and lent us an ultrasound probe which we were adapting to use for the project, which is what the first two pages of sketches is referencing. Due to the coronavirus epidemic in the United States, classes at Duke University were cancelled two weeks ago, and we were unable to proceed with the project as planned. Due to the situation, our group can no longer design a solution around the physical ultrasound probe, so we pivoted to designing our ideal device, including the ultrasound probe in our design. We plan on simulating our designed probes with simulation software to verify that they would work, and deliver our designs along with instructions in a report to our client, Dr. Steadman, so that they may create a physical version of the device once the epidemic has ceased. All the sketches below are mine. On this project, my teammates worked on the software and hardware side, while I chose to design the physical product. No user testing was performed, we did not have clearance to test our device on patients.

Level of the project: Fourth year (7th and 8th semester)

Name of the supervisors of the project: Dr. Mark Palmeri and Dr. Casey Steadman

Sketches

Designs around the original ultrasound probe (catheter form factor)



Cap solution:

- How do we seal bottom?
- with tape?
- Catheter will be hanging out.



Better solution: cuff



Valve can be used to fill the cuff w/ gel and to insert probe

Ultrasound gel pouch

Adapted Valve that allows:

- ① Vacuum Air out
- ② Fill w/ ultrasound gel
- ③ Insert probe

- lock probe into place w/ mechanical lock

④ Bottom has skin - friendly adhesive that allows one to press to skin and then peel off.

lock into position w/ adhesive to motion

Need to recreate more into valves

- without valves on the one-way ball in the middle

direction of flow

stopper bar

Ring pressure

ultrasound gel

Test probe into valve

adhesive on bottom

on/off valve for creating vacuum + pumping in ultrasound gel

latex material for ultrasound pouch

tricuspid-like valve to insert catheter through

adhesive on bottom to attach to penis

Cut off

① Latex Glove Finger




③ Insert Probe



② Fill up latex glove finger w/ Ultrasound Gel




④ Tape down setup to penis by taping top of finger and bottom; sealing ultrasound gel inside



Device Skin Coupling Designs and Testing Results


Concepts

Self sealing
Silicone Valve




IV valve

iv luer valve




Tegaderm




Custom US


Gel Pouch



Single walled latex
secured to skin
w/ mastix




Double walled latex
secured to skin using
either only one side,
all four sides, or
superglued shut



Failure modes

Failure mode: needle ~~punctures~~ ^{punctures} the skin



- Incorrect pushing out of air
- Tegaderm does not stay attached to skin for > 2 hrs
- Tegaderm leaks US gel


- Valve leaks when probe inserted through it
- mastix spirit gum doesn't hold down latex to skin well

- Same Problem as one side w/ tape
- mastix spirit gum should solve this problem (replace tape)


- tape doesn't hold due to ultrasound leaking out (w/ four sides it works better)
- superglue solution worked well under testing

Pivot: No ultrasound probe catheter form factor constraint; free thinking


3) Exploration.




Belt Design Sketch and Investigation:



Ultrasound Condom Design Sketch and Investigation:




Adhesive Patch Design Sketch and Investigation:



Finalizing designs, Elastic Band and Cap chosen:

Elastic Band

Cap




CAD of Final Deliverable:

Cap, designed for attachment to the glans penis



Full design, bottom-back view




Impedance-matching layer removed, bottom view

9 x 7 Ultrasound transducer array in center




Full design, front view


Elastic Band, designed to slide on and rest at the base of the penis



Full design, bottom-back view



Outside elastic layer removed, isometric view



Ultrasound Transducer Array Component, Front View

Linear 1 x 7 Ultrasound Transducer Array in center

ExiTrak

August 2019 – December 2020

University Final Project for Biomedical Engineering – Group Project


Descriptive Paragraph:

The towns of Welch and Williamson, West Virginia, have been economically devastated by the coal industry's disappearance from West Virginia in the later part of the 20th century. This economic crisis, compounded with a high incidence of COPD, heart disease, and diabetes in the region has created a health catastrophe: citizens cannot afford curative measures for their dangerous ailments. To address the growing need for healthcare, community health workers in these towns have prescribed exercise program for patients to perform in their homes using exercise bands. These exercise programs can dramatically improve patient outcomes, and the community health workers want to have a way to track patients' exercise, both to ensure the patient is following their prescribed exercises, and to gauge the program's success. Our design group inherited the project from a team who worked on it the prior semester and failed to create a successful product: their failures were due to faulty software and a poor mechanical design, the device had no securing mechanism to the resistance band and after a few stretches was unable to record any future stretches. We improved their mechanical design, fixing the problem of securing the band with a clip and improving the user ergonomics with a handle, and wrote an energy efficient software package using Python for the device's operation. My contribution to the effort was writing all the Python code which operated the device, brainstorming mechanical designs, and assembling the device. Our finished product could record up to 10,000 stretches per day for up to a month before the batteries had to be replaced. User research was also performed with the finished device, our group tested the public's acceptance of the device and the device's ease of use by asking ten Duke students to use the device after having read instructions, and the device passed all testing: users gave the product an average of 4.0/5.0 stars, and after reading instructions, successful stretches were performed by students in under a minute in all cases. Note: I completed some of the mechanical design work, pictures with no attribution are my work, I attribute the work of authors underneath the pictures.



Level of Project: Fourth year (7th semester)

Name of supervisor of project: Dr. Robert Malkin and Dr. Allan Shang



What we were given:





Improved Mechanical Design Iteration Sketches with labelled titles:



Device on the Handle



Additional slits Through device for band




Author of 'Additional slits through device for band': Madeline Manning

CAD Models for 3D prints:





First Iteration of Handle Design



First iteration of double slit design

Final CAD Design decided on after handle performed best on accuracy testing:



Author of rendered handle design in CAD: Qarch Hawk

Assembled ExiTrak Final Design



Pictured: User performing bicep curls and butterfly stretches



EquinOx

January 2020 – April 2020

University Electrical Engineering Final Project – Individual Project

Descriptive Paragraph:

Horses can go from healthy to fatally ill in the matter of an hour from a condition called colic, a blockage of the intestines. In a study done by the US Department of Agriculture in 2001, it was found that there is an average of 4.2 colic events / 100 horses in a year, with an 11% fatality rate for each colic event. This is a deadly fact of life for horses, and horse owners are willing to go to extreme lengths to protect their investments: in total, \$115 million was paid by horse owners for colic surgery in 1998. There is a real need for health monitoring technology for horses that will notify horse owners if the horse is having a colic event so that they can contact the veterinarian and intervene in time to save their horse's life. For this reason, my mentors, Dr. Brooke and Dr. Jokerst, and I are collaborating with the North Carolina State College of Veterinary Medicine to make this product a reality. Our plan is to build a reflective pulse oximeter that is wearable by horses. This device will monitor the horse's blood oxygenation levels and will notify the owner over a Wi-Fi connection if the horse's blood oxygenation levels dip below a defined threshold, a symptom of colic. With this product, I hope to save more horses from excruciating deaths and to give the horses' caretakers the peace of mind that wherever they are, they will always know their horse's health. Dr. Brooke, Dr. Jokerst and I designed the hardware layout of the device together, and I designed the mechanical attachment of the device to the horse individually. After the coronavirus epidemic cancelled classes, I had to restart the project from home because I no longer had access to the campus resources. As a result, the photos shown are of a rough mechanical prototype, I am still working on completing the final version that will be used by the North Carolina State College of Veterinary Medicine. All the work shown below is mine. No user research was done on horses because testing on horses was not included as part of the grant.

Level of Project: Fourth Year (8th semester)


Name of supervisor of project: Dr. Martin Brooke and Dr. Nan Jokerst

Identifying and Organizing the system components

Components:

- Microcontroller
 - Adafruit Feather
- battery tin
 - LiPo
- LED
 - 2x of each λ (4 each total)
- Proto board w/ circuitry
 - Voltage regulator
 - LED driver
 - ~~transistor~~ \rightarrow need to choose one
- Strap w/ blanket (optional)

Figure 1-A



1st iteration of planned attachment of device to horse saddle blanket components sewn under fabric




Figure 1-E

Sketch of 1st iteration for planned holder for the LEDs and Photodetector: they must be flush with skin for accurate measurements




Figure 1-0

Sketch of device's appearance on horse; explanation of attachment of saddle blanket through strap





Figure 1-1

LED and Photodetector case layouts (photodetector is in central hole, 4 LEDs in surrounding holes): each has different depths of penetration for measuring the blood oxygenation levels of different types of blood vessels




1st iteration: Laser Cut Rubber (smelly!), decided to switch to odorless Teflon for final version, but classes were cancelled before this could be implemented



Rough Physical Prototype created at home

Wood used for LED and photodetector case; LEDs and photodetectors placed with leads out



Black, light-absorbing saddle blanket with strap bottom-up, LED and photodetector case attached on bottom of blanket as seen in sketches



Prototype attached to pillow to simulate being attached to a horse, side view



Prototype attached to pillow to simulate being attached to a horse, top view



Prototype attached to pillow to simulate being attached to a horse, top view with electronics case open



Prototype attached to pillow to simulate being attached to a horse, isometric view



MedAware

February 2020

Individual Project – Reimagining Group Hackathon Project as an Industrial Design Exercise

Descriptive Paragraph:

Implicit bias kills in the United States healthcare system. There have been studies published that have established that implicit bias from doctors towards their patients can prove to have fatal consequences by showing that certain demographics have patterns of worse health outcomes than others when treated for the same ailments. My team won John Hopkins University's Medhacks 2019 by pitching MedAware, a suite that physicians and hospital management can use to track and quantify doctors' interactions with their patients, in hopes of educating health care providers of their implicit biases and improving their quality of care.


Please visit the following link to learn more about the original project:

<https://devpost.com/software/medaware>. The software and hardware components developed by myself and my team are ready to be implemented by hospital systems, but the design of the camera system can be rethought to be more user friendly and less obtrusive in the doctor's office. As an exercise of my industrial design skills, I decided to see if I can think of a better form factor that will fix the aforementioned problems with the original design.

Level of Project: Fourth Year (7th semester)

Name of Supervisor of Project: N/A

Complete Product Explanation:



Original Design: clunky mechanical design and scary for patients in the clinic to see a camera


May give the feeling that Big Brother is watching...



Exploration, brainstorming kind of form factors can we solve this issue with

Exploration:

One-way mirror
ceiling tile



Constraint: needs to be facing
down towards patient/doctor
(needs to be 90° to ground)

Smart Camera




What if room is too big
for camera lens?




Raspberry Pi





battery (optional)



Multiple lights that
cover whole room/camera in
each one.




Very conspicuous
(or have a complete ceiling
made from one-way glass)



Camera system that moves along
central axis of other x/y direction


Continued exploration, transition to idea of hiding in plain sight



Can be built into
fan base too.

Method should be dependent
on clinic.

Can be conspicuous but cute!
-kid friendly



looks on ceiling,
Camera inside, facing down
through a one-way section,



Camera
inside
play h
tag
sitting
near
to
doctor

Goal identified after analysis of exploration: either hide the camera or make it friendly!

Multiple options for different audiences:

~~Appropriate for kids~~
~~pieces of~~

Hide in plain sight:

- animals for kids (elephant, giraffe sitting, giraffe on ceiling)


~~Adults would be adults horizontal~~

Just hide

- don't need to optimize for audience
- place in light, ceiling tile
- place in fruit basket, radio

Goal: turn something scary either into something
not ~~be~~ seen or ~~not~~ into something friendly

Friendly Form Factor:





Continued Friendly Form Factors

Idea Sketches:

Tiger gecko for ceiling:


Side View



Top View

module embedded
inside dark stripes: not easily seen


Koala Sloth for ceiling





module ~~embedded~~
not visible on
the back:
inside school bag
worn by sloth/koala

Hiding the Camera: One-Way Mirror Ceiling Tile

Ceiling Tile Interface



Either battery
powered or wall
powered
(Context dependent)



Camera and
Raspberry Pi screwed
together


tile w/ cut
hole for
camera

Raspberry
Pi (camera)


One-way
mirror

Hiding the Camera: Light Interface

Light Interface Diagram



Side View



View from Bottom Up



replaced by



Fit camera +
Rpi into
compact lightbulb
interface that
is powered
by lightbulb
interface on
existing light fixture.

