



PT. TOPAZ BORNEO UTAMA

Jl. Ahmad Yani, KM. 36,5 No. 4
Banjarbaru, Kalimantan Selatan
Phone : (0511) 4781975

No. 004 / SRT / Januari / 2022..)

SERVICE REPORT

Number: 1	Unit Type: Norma Icon3	Serial No: ND30H17024	Version: 0.4.922.0	Date: Selasa 25 / 01 / 2022
				Travel Time (B-F): 01 h m
				Time Work Started: 12 h 00 m
				Time Work Finished: 13 h 40 m

Customer: UPTD Labkes
Kota Banjarmasin

Address/Ph Number:
Jl. Pramuka, Sungai Iwut, Kec. Banjarmasin Timur

Job Title: ☐ Install ☒ Repair ☐ Update
☐ Training ☒ Maintenance ☐ Refurbishment

Report Fault: Perbaikan hasil fastblank PTT tinggi

Action

- Cek reagent diluent, mengganti reagent diluent karena ada semut
- Drain full, Prime Reagent, Cleaning: open, RBC clean, Closed Rinsing
- Cek fastblank, Jalankan Kalibrasi & QC
- Cek sparepart yang perlu diganti

Remarks: Fastblank sudah o.k

- Kalibrasi & QC masuk target


Sparepart yang perlu diganti: Glass syringe with piston

Status After Service : ☒ Operational ☐ Non Operational ☐ To be Monitored

SPAREPART USE

No	Code Catalogue	Description	Qty	Invoice	Price
Total Price					

Report by Engineer


Rahmat Ramadhan

Acknowled and Approved by Customer:



Checked by Nama Perusahaan