

SERVICE LOGS EXTRACT

Source: BioServe shared portal and MedCare facility records

Period: February 1 to March 27, 2026

February Preventive Maintenance

- Tunis CT unit: completed February 8, signed by site manager
- Sousse ultrasound unit: completed February 14, signed by biomedical lead
- Ariana MRI unit: no signed preventive maintenance sheet located
- Remote telemetry note for Ariana MRI: February 26, status marked incomplete

March Events

- March 17, 07:52: Ariana MRI warning C-417 generated
- March 19, 18:12: Second C-417 warning generated
- March 21, 08:28: MedCare ticket opened
- March 21, 09:10: BioServe ticket accepted
- March 21, 14:30: BioServe technician onsite
- March 21, 17:40: Limited operation restored
- March 24, 16:05: Calibration stability test failed once, then passed

Log Integrity Concerns

The portal export does not contain a technician signature for the February Ariana entry. Two March warning events were automatically closed by remote monitoring without MedCare confirmation.

Potentially Helpful Evidence

The system event timestamps are machine-generated. They may be stronger evidence than later email narratives if authenticity is preserved.