

INTERNAL INCIDENT REPORT

Report ID: IR-MEDCARE-2026-03-21

Prepared by: MedCare Operations and Compliance

Date: March 22, 2026

Incident Summary

On March 21, 2026, the Ariana clinic MRI unit was unavailable from 08:15 to 17:40. Twenty-three patient appointments were delayed or moved. The incident was classified as critical because the affected unit supports urgent neurological diagnostic referrals.

Timeline

- 08:15: Radiology coordinator reported boot failure and calibration error C-417
- 08:28: Helpdesk ticket opened with BioServe
- 09:10: BioServe acknowledged ticket and requested log export
- 11:55: MedCare escalated because no technician was onsite
- 14:30: BioServe technician arrived
- 16:50: Temporary workaround applied
- 17:40: Unit returned to limited operation

Initial Findings

The BioServe response exceeded the 6 business hour onsite response standard. Preventive maintenance records for February 2026 do not show completion for the Ariana MRI unit. The unit had generated calibration warnings on March 17 and March 19 that were not escalated by BioServe.

Business Impact

- 23 patient appointments disrupted
- 9 outsourced scans booked at partner clinic
- Estimated direct outsourcing cost: 8,920 TND
- Potential regulator concern if repeat downtime affects continuity of care