IRENE

Product Quality Report

Record Number: ZL-230116-7895-25 Document Number: ZT/JL-8.2-03-ZL03-001 Ref. ACUTEC™ Headless **Product Product** Lot 2200114713 ZT.T52074040 Φ 4.0×40mm Number Specification Number Name Compression Screw Inspection National Medical Device Registration Standard 20163462434 Cannulated Screws Standard Inspection No. Inspection Item Standard Requirement Inspection Result Clause It shall meet the requirements of Clause 2.1 of Technical Requirements in National Medical Device **Fulfilled** 2.1 1 Material Registration Standard 20163462434. Mechanical Titanium Alloy Hardness of Cannulated screws and washers:≥260HV10 **Fulfilled** 2 2.2 **Properties** Surface defects: The surface of cannulated screws and washers must not have discontinuous defects; The surface roughness shall meet Clause 2.3.2 of Technical Requirements in National Medical Device **Fulfilled** 3 Surface Quality 2.3 Registration Standard 20163462434. Appearance: The surface of cannulated screws and washers should be smooth and free of defects such as sharp edges, burrs, and attachments. It shall meet the requirements of Clause 2.4 of Technical Requirements in National Medical Device 2.4 Fulfilled 4 **Dimensions** Registration Standard 20163462434. It shall meet the requirements of Clause 2.5 of Technical Requirements in National Medical Device **Fulfilled** 5 **Delivery Status** 2.5 Registration Standard 20163462434. Final Inspection Result □Disqualified Authorizer/Date Inspector/Date Verifier/Date 2022-08-16 2022-08-16 2022-08-16