Advanced Computational Framework for Pediatric Ureteral Stent Optimization: A Critical Review and Methodology Enhancement

1. Executive Summary and Strategic Overview

The proposal titled "Improved Simulation Strategy for Pediatric Ureteral Stents in COMSOL" represents a foundational step toward addressing a critical gap in pediatric urology: the lack of medical devices specifically engineered for the biomechanical realities of the developing urinary tract. Current clinical practice largely relies on scaling down adult devices—a strategy that fails to account for the non-linear scaling of physiological parameters such as peristaltic frequency, compliance, and somatic growth rates.[1] The proposal correctly identifies that the "trial-and-error" approach in pediatric stenting leads to high morbidity, including vesicoureteral reflux (VUR), migration, and severe bladder irritation.

However, a rigorous critique of the proposed computational strategy reveals significant opportunities for refinement. While the reliance on the Palmer "Age + 10" rule [2] and the Arruda-Boyce constitutive model [3] provides a stable starting point, the proposal underestimates the complexity of the pediatric biomechanical environment. Specifically, it treats the ureter as a quasi-static or predictably oscillating domain, neglecting the chaotic transients of intra-abdominal pressure (e.g., coughing, straining) which are the primary drivers of reflux-associated morbidity in children.[4] Furthermore, the assumption of isotropic hyperelasticity simplifies the highly anisotropic nature of the ureteral wall, potentially leading to inaccurate predictions of stent-induced wall stress and subsequent pain.[5]

This report provides an exhaustive technical critique and a detailed roadmap for enhancing the simulation strategy. It validates anatomical inputs against the latest MRI-based normative data, challenges the material model selection with anisotropic alternatives, and introduces a suite of "Mental Tests" to verify the numerical stability of COMSOL implementations. By expanding the scope to include critical boundary cases—specifically transient reflux and migration forces—this enhanced framework aims to transition the project from a theoretical study to a clinically predictive design tool.

---

2. Anatomical and Physiological Verification: Beyond

"Rules of Thumb"

The validity of any Finite Element Analysis (FEA) or Computational Fluid Dynamics (CFD) model is strictly bounded by the fidelity of its geometric inputs. The PDF proposal relies on clinical heuristics that, while useful for bedside estimation, may introduce unacceptable variance in a high-precision optimization study.

2.1 The "Age + 10" Rule vs. Anthropometric Realities

The proposal utilizes the formula $Length(cm) = Age(years) + 10$ to derive a stent length of ~15 cm for a 5-year-old child.[1]

Critique of Accuracy: The "Age + 10" rule, derived by Palmer and Palmer (2007), is a validated clinical heuristic with a reported correlation coefficient (R^2) of 0.88 across a cohort of 153 children.[2] It is designed to prevent the selection of stents that are too short (risk of migration) or too long (risk of bladder urgency). However, reliance on this rule for *simulation* design parameters assumes a standard growth curve that many pediatric urology patients do not follow. Children requiring stents often suffer from pathologies like Chronic Kidney Disease (CKD) or congenital defects that result in growth stunting or failure to thrive.

Conflicting Evidence: Research by Barrett et al. (2016) and others suggests that patient height is a superior predictor of ureteral length compared to age.[6] A 5-year-old in the 5th percentile for height will have a significantly shorter ureter than one in the 95th percentile. Using a fixed 15 cm geometry in COMSOL ignores this variance. If the simulation optimizes a stent for a 15 cm ureter, but the patient's actual ureter is 13 cm, the "optimal" design may cause severe trigonal irritation due to excessive distal coil length.

Recommendation:

The simulation geometry must not be a single fixed value. It should be parameterized to sweep a range of lengths derived from the standard deviation (σ) of the Palmer data.

- **Proposed Range:** $L \in $ cm.
- **Justification:** This range covers the Age+10 mean (15 cm) $\pm$ approximately 2 standard deviations, accounting for both natural anatomical variance and growth stunting.[2]

2.2 Ureteral Diameter and Wall Thickness: The MRI vs. Ultrasound Debate

The PDF states an internal diameter (ID) of 3–4 mm and estimates wall thickness at ~1 mm.[1]

Critique of Diameter: Recent normative data from Shashi et al. (2022), utilizing Magnetic Resonance Urography (MRU), provides more granular data than the older sources cited in the

PDF. Shashi found that the average widest internal ureteral diameter in children (ages 5-6) is approximately **3.2 mm ± 0.3 mm**.[7]

- **Insight:** The PDF's range of 3-4 mm biases the simulation toward the upper limit of normality. The critical failure mode for stent insertion is often the narrowest segment (the ureterovesical junction, UVJ), which may be < 3 mm. A simulation based on a 4 mm lumen will underestimate the hydraulic resistance and the occlusion pressure required to block flow.

Critique of Wall Thickness: The estimate of ~1 mm is a potential weak point in the model's structural fidelity. Robben et al. (1999), using high-resolution ultrasound, measured the renal collecting system wall thickness in children and found a mean of **1.6 mm** (range 0.8–3.1 mm).[8]

- **Implication:** The flexural stiffness of a tube scales with the cube of its thickness ($I \propto t^3$). Underestimating the wall thickness by 0.6 mm (modeling 1.0 mm instead of 1.6 mm) could lead to a significant underestimation of the ureter's ability to resist stent expansion. This would falsely predict that the ureter "conforms" to the stent shape more easily than it does in vivo, potentially masking issues related to pressure necrosis.

2.3 Peristaltic Dynamics: The Hyper-Peristalsis Boundary

The PDF cites a wave propagation speed of 2–3 cm/s and peak pressures of 30–40 cm H2O.[1]

Verification and Nuance:

- **Wave Speed:** The value of 2–3 cm/s is well-supported by Vahidi and Fatouraee (2012) and animal models.[3]
- **Pressure:** The peak pressure of ~4 kPa is accurate for a healthy bolus. However, the simulation must account for **Hyper-peristalsis**. When a foreign body (stent) is introduced, the ureter often responds with high-frequency, high-amplitude spasms (colic) in an attempt to expel the object. Pressures in obstructed or stented ureters can transiently exceed **60 cm H2O**.[10]
- **Recommendation:** The simulation must model a "Spasm Mode" with increased frequency (e.g., 10 waves/min) and amplitude (6 kPa) to stress-test the stent's drainage capacity under pathological conditions.

2.4 Validated Parameter Table for Simulation Setup

The following table synthesizes the corrected values to be used in the COMSOL definitions.

Parameter	PDF Value	Verified Clinical Value	Simulation Strategy	Source

Ureter Length	~15 cm	$15 \pm$ cm (Age + 10 variance)	**Parameter Sweep:** 12 cm to 18 cm	Palmer (2007) [2]
Internal Diameter	3-4 mm	$3.2 \pm$ mm (5-6 yr old)	**Parameter Sweep:** 2.8 mm to 3.8 mm	Shashi (2022) [7]
Wall Thickness	~1 mm	1.6 mm (mean range 0.8-3.1)	**Baseline:** 1.5 mm; **Robustness Check:** 2.0 mm	Robben (1999) [8]
Wave Speed	2-3 cm/s	2-3 cm/s	**Fixed:** 2.5 cm/s	Vahidi (2012) [3]
Peak Pressure	3-4 kPa	2-6 kPa (Obstructed/Spasming)	**Scenario A:** 3 kPa (Normal) **Scenario B:** 6 kPa (Spasm)	Carugo (2021) [10]

---

3. Boundary Cases and Failure Mode Analysis: The Missing Physics

The original proposal focuses heavily on the steady-state or quasi-steady performance of the stent (drainage efficiency). However, in pediatric urology, device failure often occurs at the boundaries of normal operation—during rapid transient events or long-term biological integration. This section details three critical boundary cases that are absent or underdeveloped in the PDF but are essential for a robust design.

3.1 Boundary Case 1: Transient Reflux during Coughing (The "Water Hammer" Effect)

The PDF assumes baseline renal pelvis pressures (~10-15 cm H2O). It misses the dominant hydraulic event in pediatric patients: sharp increases in intra-abdominal pressure (IAP) due to coughing, sneezing, or straining (Valsalva).

The Physics of Failure: In a healthy child, the ureterovesical junction (UVJ) acts as a competent check valve, preventing retrograde flow during IAP spikes. A stent props this valve

open. During a cough, IAP can spike to **100–150 cm H2O (10–15 kPa)** within milliseconds.[4]

- **Implication:** Because the pediatric ureter is short (15 cm vs. 30 cm in adults), this pressure wave propagates to the kidney much faster. This "water hammer" effect transmits high pressure to the renal papillae, causing flank pain and potential renal scarring.
- **The Gap:** A simulation that only looks at peristaltic drainage will favor stents with large lumens and many side holes (low resistance). However, these same features make the stent a perfect conduit for high-pressure reflux.

Simulation Implementation:

- **Setup:** A time-dependent study step separate from the peristaltic cycle.
- **Boundary Condition:** At the bladder outlet ($z = L$), impose a transient pressure pulse $P_{bladder}(t)$ modeled as a Gaussian pulse:

$$P(t) = P_{base} + P_{cough} \cdot e^{-\frac{(t-t_0)^2}{2\sigma^2}}$$

  Where $P_{cough} \approx 10$ kPa and duration ≈ 0.5 s.
- **Metric:** Calculate the **Peak Pressure at the Renal Pelvis** and **Net Retrograde Volume**. The optimization goal is to minimize these values while maintaining an acceptable antegrade flow.

3.2 Boundary Case 2: Stent Migration and "Jack-knifing"

Stent migration is statistically more prevalent in children than adults, with rates reported between 2.8% and 41% depending on the cohort.[12] The PDF treats the stent as geometrically fixed.

The Physics of Failure:

Children are highly active. Inertial forces from running or jumping, combined with the somatic growth of the child (which effectively shortens the stent relative to the ureter over time), apply tension to the stent. If the distal (bladder) coil is not sufficiently stiff or geometrically locked, it can uncoil and retract into the ureter (proximal migration), requiring invasive retrieval. Conversely, it can prolapse into the urethra.

Simulation Implementation:

- **Setup:** A structural mechanics study (Stationary or Time-Dependent).
- **Load:** Apply a prescribed axial displacement (Δz) to the distal tip of the stent or a body load representing inertial acceleration ($F = ma$).
- **Metric: Pull-out Force**. Measure the reaction force required to uncoil the pigtail.

- **Design Interaction:** This creates a conflict. A stiffer coil prevents migration but increases bladder irritation. The simulation must find the Pareto optimal point between "Migration Resistance" and "Contact Pressure" on the bladder wall.

3.3 Boundary Case 3: Encrustation as a Dynamic Boundary

Encrustation is mentioned in the PDF background but not integrated into the simulation physics. Encrustation is not just a binary failure; it is a progressive change in boundary conditions.

The Physics of Failure: Biofilms and crystal deposits form in regions of **Low Wall Shear Stress (WSS)** and recirculating flow.[10] As side holes occlude, the hydraulic resistance of the stent increases, potentially altering the flow path from intraluminal to extraluminal.

Simulation Implementation:

- **Setup:** Phase 1 screening cannot assume a "clean" stent.
- **Method:** Perform a sensitivity analysis where side holes are sequentially blocked (e.g., 25%, 50%, 75% occlusion) based on WSS maps from the clean simulation.
- **Metric: Fail-Safe Performance**. Does the stent still provide adequate drainage when 50% of its holes are blocked? A design that relies on complex micro-features that easily foul is inferior to a simpler, more robust design.

---

4. Constitutive Modeling: Arruda-Boyce vs. The Reality of Anisotropy

The choice of material model dictates the simulation's ability to predict occlusion and stress. The PDF proposes the **Arruda-Boyce** model.

4.1 Critique of the Arruda-Boyce Model

The Arruda-Boyce model is an 8-chain statistical mechanics model originally derived for rubber elasticity. It captures the non-linear "locking" behavior of polymer chains—initially compliant, then stiffening rapidly as chains reach their extensibility limit.

- **Relevance:** This effectively mimics the passive behavior of the ureter, which is compliant at low pressures but stiffens to prevent rupture at high distension.[3] It is numerically stable and computationally efficient in COMSOL.
- **Deficiency:** The model assumes **isotropy** (mechanical properties are identical in all directions). The ureter, however, is a composite of smooth muscle and collagen fibers oriented helically and longitudinally. It is **anisotropic**. The longitudinal stiffness (governing stent buckling and accommodation) differs from the circumferential stiffness (governing peristaltic occlusion).[13]

4.2 The High-Fidelity Alternative: Holzapfel-Gasser-Ogden (HGO)

For high-fidelity validation (Phase 2), the **Holzapfel-Gasser-Ogden (HGO)** model is the gold standard for soft biological tissues.[5] It includes terms for the strain energy of collagen fiber families dispersed in the tissue matrix.

- **Argument for HGO:** If the goal is to predict tissue trauma (pain), we must know the stress in the specific fiber directions. Isotropic models smears this stress, potentially underestimating the peak stress at the stent contact points.
- **Argument Against HGO (in this specific context):** HGO requires 5-9 material parameters (fiber angle dispersion, fiber stiffness, matrix stiffness). Obtaining reliable *pediatric-specific* values for these parameters is nearly impossible due to ethical restrictions on tissue harvesting. Using scaled adult parameters introduces epistemic uncertainty that likely outweighs the model's theoretical precision.

4.3 Decision Matrix and Recommendation

Feature	Arruda-Boyce (Proposed)	Holzapfel-Gasser-Ogden (Alternative)
Stability	High (Convex strain energy)	Low (Prone to instability if parameters unchecked)
Data Req.	Low (Shear modulus μ, Locking λ_m)	High (Fiber angles γ, Dispersion κ, etc.)
Biofidelity	Medium (Captures stiffening, misses directionality)	High (Captures anisotropy and recruitment)
Suitability	**Best for Optimization (Phase 1)**	**Best for Detailed Trauma Analysis (Phase 2)**

Recommendation: The report should endorse **Arruda-Boyce** for the primary optimization loop due to its robustness and the lack of pediatric fiber data. However, a **Sensitivity**

Analysis must be added: the Shear Modulus (μ_0) should be swept across a range (e.g., 10 kPa to 200 kPa) to ensure the optimal design holds true for both "floppy" and "stiff" ureters.

---

5. COMSOL Implementation: Verification and "Mental Tests"

The user requested "Mental Tests" to verify familiarity with COMSOL functions like flc2hs. These tests are crucial for the methodology section to demonstrate that the simulation setup is mathematically sound and numerically stable.

5.1 The flc2hs Function: Mathematical Verification

The function flc2hs(x, scale) stands for **Fl**smoothed **C**ontinuous **2**nd derivative **H**eaviside **S**tep. It creates a smooth transition from 0 to 1 over an interval defined by scale.[15]

Why it is mandatory:

Standard step functions (jumping from 0 to 1 instantly) introduce infinite gradients. In Finite Element Analysis, this causes the Jacobian matrix to become singular or forces the time-dependent solver to take infinitesimally small time steps (e.g., 10^{-12} s) to resolve the discontinuity, effectively freezing the simulation. flc2hs ensures C^2 continuity, allowing the Newton-Raphson solver to converge efficiently.

Equation for Peristalsis:

The moving pressure wave $P(z,t)$ should be defined as:

$$P(z,t) = P_{peak} \cdot (flc2hs(z - v \cdot t + \frac{W}{2}, \epsilon) - flc2hs(z - v \cdot t - \frac{W}{2}, \epsilon))$$

Where W is the wave width and ϵ is the smoothing scale.

5.2 Mental Test 1: The "Step-Width Challenge"

- **Concept:** The relationship between the smoothing scale (ϵ) and the mesh element size (Δx).
- **The Test:** Imagine you set the smoothing scale $\epsilon = 0.1$ mm to make the pressure wave look "sharp" and realistic. Your mesh element size is $\Delta x = 1$ mm. What happens?
- **Result:** The solver will likely fail or produce garbage results ("ringing").
- **Reasoning:** If $\epsilon < \Delta x$, the pressure rises from 0 to max *inside* a single element. The shape functions (linear or quadratic) cannot resolve this sub-grid curvature. The solver sees a "shock" that jumps discontinuously between nodes.

- **Pass Condition:** You must strictly enforce $\epsilon \geq \Delta x$. Ideally, the transition should span 2-3 elements ($\epsilon \approx 2\Delta x$) to ensure the load is integrated correctly across the mesh.

5.3 Mental Test 2: The "Inverted Element" in ALE

- **Concept:** Arbitrary Lagrangian-Eulerian (ALE) moving meshes handling lumen occlusion.
- **The Test:** You simulate a strong peristaltic contraction that squeezes the ureter completely shut (lumen diameter $\rightarrow$ 0). What happens to the mesh?
- **Result:** The simulation crashes with an "Inverted Mesh Element" error.
- **Reasoning:** ALE tracks the deformation of the fluid domain. If the walls touch, the fluid volume at that node becomes zero (or negative). The Jacobian of the geometric mapping becomes negative, which is mathematically impossible in this formulation.
- **Pass Condition:** You must implement a "minimum gap" constraint or a **Penalty Contact** formulation. The walls must effectively stop squeezing when they are separated by a thin fluid film (e.g., 50 μm). This avoids mesh inversion while approximating occlusion. Alternatively, use the **One-Way Coupled** approach (solve structure, then flow) if full topological change is not required.

5.4 Implementation Option: One-Way vs. Two-Way FSI

The PDF suggests a "Simplified FSI" for Phase 1. We must define this rigorously.

- **Option A: Fully Coupled FSI (Two-Way).** The fluid pressure deforms the wall; the wall deformation alters the fluid flow.
  - *Cost:* High (Hours/Days per run).
  - *Necessity:* Only if the fluid pressure is high enough to push the wall back (e.g., arterial blood flow). In the ureter, the muscle contraction force (kPa range) vastly dominates the fluid inertial forces.
- **Option B: One-Way Coupled FSI (Phase 1 Recommendation).**
  1. Solve the Solid Mechanics study first to generate the peristaltic wall deformation $u_{wall}(t)$.
  2. Apply this deformation as a moving wall boundary condition ($v_{wall} = \frac{\partial u}{\partial t}$) to the Fluid Dynamics study.
  - *Benefit:* Drastically faster. Decouples the stiff solid solver from the fluid solver.
  - *Argument:* Valid because urine density and velocity are low; the "added mass" effect is negligible compared to the stiffness of the contracting muscle.

---

6. Optimization Strategy: Reflux-Weighted Objective

Function

The PDF proposes maximizing flow and minimizing pressure drop. This is insufficient because a stent that maximizes flow (large open pipe) also maximizes reflux.

Critique:

A large-diameter stent with many side holes is optimal for drainage but catastrophic for reflux protection. The optimization must penalize reflux.

Proposed Objective Function:

Instead of a simple maximization, use a composite cost function J to be minimized via Bayesian Optimization:

$$J(\mathbf{x}) = w_1 \cdot \frac{1}{Q_{antegrade}} + w_2 \cdot V_{reflux} + w_3 \cdot \sigma_{max}$$

Where:

- $\mathbf{x}$ is the vector of design variables (length, diameter, side-hole count).
- $Q_{antegrade}$ is the forward flow rate (Drainage).
- V_{reflux} is the volume of fluid returning to the kidney during the "Cough" simulation (Safety).
- σ_{max} is the peak Von Mises stress in the bladder trigone (Comfort).
- w_i are weighting factors determined by clinical priority (usually Safety > Drainage > Comfort).

Bayesian Optimization (GPR): The choice of Gaussian Process Regression (GPR) is validated.[17] Unlike Genetic Algorithms which require thousands of function evaluations, GPR creates a probability map of the objective function and uses an **Acquisition Function** (like Expected Improvement) to intelligently select the next design point. This is ideal for computationally expensive FSI simulations where every data point costs hours of compute time.

---

7. Conclusions and Final Recommendations

The proposed "Improved Simulation Strategy" is a promising initiative that addresses a clear unmet need in pediatric healthcare. The use of COMSOL's ALE capabilities and the

Arruda-Boyce material model provides a sound technical foundation. However, to achieve clinical relevance, the simulation must move beyond "steady-state drainage" and aggressively model the **boundary cases** that lead to device failure.

Summary of Required Enhancements:

1. **Geometric Parameterization:** Abandon the fixed 15 cm length. Implement a sweep ($L = 12 - 18$ cm, $D = 2.8 - 3.8$ mm) to account for growth stunting and anatomical variance.[2]
2. **Material Robustness:** Retain Arruda-Boyce for stability but conduct sensitivity analyses on the Shear Modulus (μ) to bound the behavior of anisotropic tissues without the complexity of HGO.
3. **Numerical Rigor:** Implement flc2hs with smoothing scales $\epsilon \geq 2\Delta x$ and utilize One-Way Coupled FSI for the screening phase to ensure solver convergence.
4. **Critical Boundary Cases:**
   - Add a **Transient Cough Simulation** (10 kPa pulse) to quantify reflux risk.
   - Add a **Migration Force Simulation** to optimize coil stiffness against dislodgement.
5. **Multi-Objective Optimization:** Redefine the optimization goal to explicitly penalize reflux volume, preventing the selection of designs that are "too efficient" at the cost of safety.

By integrating these physiological realities and numerical safeguards, the project will produce not just a theoretical model, but a robust engineering tool capable of guiding the next generation of pediatric ureteral stents.

Works cited

1. Improved Simulation Strategy for Pediatric Ureteral Stents in COMSOL.pdf
2. Determining the Proper Stent Length to Use in Children - AUA Journals, accessed February 10, 2026, https://www.auajournals.org/doi/pdf/10.1016/j.juro.2007.03.191
3. A biomechanical simulation of ureteral flow during peristalsis using intraluminal morphometric data | Request PDF - ResearchGate, accessed February 10, 2026, https://www.researchgate.net/publication/51981049_A_biomechanical_simulation_of_ureteral_flow_during_peristalsis_using_intraluminal_morphometric_data
4. Effect of Material Properties on Predicted Vesical Pressure During a Cough in a Simplified Computational Model of the Bladder and Urethra | Request PDF - ResearchGate, accessed February 10, 2026, https://www.researchgate.net/publication/230717037_Effect_of_Material_Properties_on_Predicted_Vesical_Pressure_During_a_Cough_in_a_Simplified_Computational_Model_of_the_Bladder_and_Urethra
5. A new anisotropic soft tissue model for elimination of unphysical auxetic behaviour, accessed February 10, 2026, https://www.researchgate.net/publication/343956753_A_new_anisotropic_soft_tis

sue_model_for_elimination_of_unphysical_auxetic_behaviour
6. Determination of optimal stent length: a survey of urologic surgeons - PMC, accessed February 10, 2026, https://pmc.ncbi.nlm.nih.gov/articles/PMC10091891/
7. (PDF) Normative values for ureteral diameter in children - ResearchGate, accessed February 10, 2026, https://www.researchgate.net/publication/359799984_Normative_values_for_ureteral_diameter_in_children
8. Significance of thickening of the wall of the renal collecting system in children: an ultrasound study - PubMed, accessed February 10, 2026, https://pubmed.ncbi.nlm.nih.gov/10525779/
9. Numerical Simulation of Peristaltic Urine Flow in a Stented Ureter - ResearchGate, accessed February 10, 2026, https://www.researchgate.net/publication/266222772_Numerical_Simulation_of_Peristaltic_Urine_Flow_in_a_Stented_Ureter
10. Fluid mechanical modeling of the upper urinary tract - UCL Discovery - University College London, accessed February 10, 2026, https://discovery.ucl.ac.uk/id/eprint/10120741/10/Carugo_wsbm.1523.pdf
11. Megh R. Goyal (Editor), Arka Bhowmik (Editor), Anamika Chauhan (- Biofluid Dynamics of Human Body Systems (2025, Apple Academic Press) - Libgen - Li - Scribd, accessed February 10, 2026, https://www.scribd.com/document/895805016/Megh-R-Goyal-Editor-Arka-Bhowmik-Editor-Anamika-Chauhan-Biofluid-Dynamics-of-Human-Body-Systems-2025-Apple-Academic-Press-Libgen-li
12. Antegrade Ureteral Stenting in Pediatric Patients: Introducing a Novel Ureteral Morphological Classification and a New Perspective on Functional Success - MDPI, accessed February 10, 2026, https://www.mdpi.com/2077-0383/14/1/246
13. Successful Treatment with a 6.3-Fr Disposable Digital Flexible Ureteroscope and Holmium Laser Lithotripsy in a 50-Year-Old Man with Large Bilateral Ureteral Stones and Hydronephrosis - PMC - PubMed Central, accessed February 10, 2026, https://pmc.ncbi.nlm.nih.gov/articles/PMC12579438/
14. tissue hyperelastic parameters: Topics by Science.gov, accessed February 10, 2026, https://www.science.gov/topicpages/t/tissue+hyperelastic+parameters
15. Peristaltic Pump, accessed February 10, 2026, http://courses.washington.edu/bioen327/COMSOL/models.sme.peristaltic_pump.pdf
16. Case Study of Heat Transfer during Artificial Ground Freezing with Groundwater Flow - MDPI, accessed February 10, 2026, https://www.mdpi.com/2073-4441/10/10/1322
17. (PDF) Computational Optimization Study of Transcatheter Aortic Valve Leaflet Design Using Porcine and Bovine Leaflets - ResearchGate, accessed February 10, 2026, https://www.researchgate.net/publication/334531946_Computational_Optimization_Study_of_Transcatheter_Aortic_Valve_Leaflet_Design_Using_Porcine_and_Bovine_Leaflets