

	Gautam Solar Private Limited		
	Specifications for Calibration	Doc.No.: GSPL/QA/S/11	
		Issue Date:- 01-12-2024	
		Rev.No./Rev.Date:- 00	

Golden Module Serial No.	GS04770TG0452500014	Dimension of Module	2278mm*1134mm	
Parameters			Lower Calibration Limit(W)	Upper Calibration Limit(W)
Pmax (W): 590.07		VALIDATION (± 1%)	584.1	595.9
Voc (V) : 51.57		CALIBRATION (±0.5%)	51.05	52.08
Isc (A) : 14.28		OUT OF RANGE (± 1.5%)	14.1372	14.4228
Module Manufacturing Date: 14/02/2025		Module Calibration Date: 20/03/2025 Calibration Expiry Date: 19/03/2026		

Prepared By	Reviewed By	Approved By
Ashutosh Kumar	Akhilesh Singh	Aman Kushwaha