




FORTEST

Testing Technology. Together.

Pinout RS232 cable (T-Series and M-Series ForTest)



COM1 side (cable male)



PC/PLC side (cable female)

In order to communicate with the instrument, the serial cable must be done with the following pinout:

	COM1 (male)	PC/PLC (female)
PIN NR:	6	2
PIN NR:	7	3
PIN NR:	5	5

The cable have to be connected to the COM1 port on the rear of the instrument.
To test the communication please follow [this](#) procedure.

On the COM2 port, the communication will not work with this pinout but it's necessary to use a RS232 standard cable.

