



Filename: Tapestation Lib Report TCP_2627_023_1.cRNA


Default image (Contrast 19%), Image is Scaled to Highest Sample Peak

Sample Info

Well	RIN ^e	28S/18S (Area)	Conc. [ng/μl]	Sample Description	Alert	Observations
A1	-	-	70.4	Electronic Ladder		Ladder
B1	6.6	1.0	189	SAM8875 P41BC		
C1	1.5	-	40.8	SAM8876 P42BC		The upper ribosomal fragment has degraded
D1	6.8	1.6	826	SAM8877 P40BC	⚠	Sample concentration outside functional range for RIN ^e and the assay
E1	5.7	0.9	31.9	SAM8878 P39BC		The upper ribosomal fragment has degraded
F1	4.6	2.0	2820	SAM8882 P35BC	⚠	Sample concentration outside functional range for RIN ^e and the assay; RIN ^e edited
G1	3.7	0.4	223	SAM8885 P31BC		The upper ribosomal fragment has degraded
H1	6.0	1.1	28.5	SAM8886 P32BC		
A2	6.8	0.6	374	SAM8887 P34BC		
B2	6.2	0.7	12.1	SAM8888 A	⚠	Sample concentration outside functional range for RIN ^e and the assay; The upper ribosomal fragment has degraded
C2	8.6	1.6	326	SAM8890 P24BC		
D2	8.5	1.4	324	SAM8891 P26BC		

A1: Electronic Ladder




Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/μl]	Sample Description	Alert	Observations
A1	-	-	70.4	Electronic Ladder		Ladder

Peak Table

Size [nt]	Calibrated Conc. [ng/μl]	Assigned Conc. [ng/μl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
200	4.92	-	72.4	7.80	
500	13.2	-	77.6	20.88	
1000	11.8	-	34.6	18.62	
2000	11.4	-	16.8	18.11	
4000	12.9	-	9.47	20.38	
6000	8.98	-	4.40	14.22	

B1: SAM8875 P41BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/ul]	Sample Description	Alert	Observations
B1	6.6	1.0	189	SAM8875 P41BC		

Peak Table

Size [nt]	Calibrated Conc. [ng/ul]	Assigned Conc. [ng/ul]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1662	28.6	-	50.6	49.15	
4660	29.6	-	18.7	50.85	

C1: SAM8876 P42BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/μl]	Sample Description	Alert	Observations
C1	1.5	-	40.8	SAM8876 P42BC		The upper ribosomal fragment has degraded

Peak Table

Size [nt]	Calibrated Conc. [ng/μl]	Assigned Conc. [ng/μl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
970	1.52	-	4.62	100.00	


D1: SAM8877 P40BC**Sample Table**

Well	RINe	28S/18S (Area)	Conc. [ng/μl]	Sample Description	Alert	Observations
D1	6.8	1.6	826	SAM8877 P40BC	⚠	Sample concentration outside functional range for RINe and the assay

Peak Table

Size [nt]	Calibrated Conc. [ng/μl]	Assigned Conc. [ng/μl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1777	114	-	188	38.67	
4780	180	-	111	61.33	

E1: SAM8878 P39BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/μl]	Sample Description	Alert	Observations
E1	5.7	0.9	31.9	SAM8878 P39BC		The upper ribosomal fragment has degraded

Peak Table

Size [nt]	Calibrated Conc. [ng/μl]	Assigned Conc. [ng/μl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1712	2.42	-	4.15	51.40	
6049	2.28	-	1.11	48.60	

F1: SAM8882 P35BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/ul]	Sample Description	Alert	Observations
F1	4.6	2.0	2820	SAM8882 P35BC		Sample concentration outside functional range for RINe and the assay; RINe edited

Peak Table

Size [nt]	Calibrated Conc. [ng/ul]	Assigned Conc. [ng/ul]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
627	274	-	1290	11.59	
987	150	-	446	6.33	
1724	456	-	778	19.29	
4517	890	-	579	37.66	
5232	342	-	192	14.46	
6254	252	-	119	10.67	


G1: SAM8885 P31BC**Sample Table**

Well	RINe	28S/18S (Area)	Conc. [ng/μl]	Sample Description	Alert	Observations
G1	3.7	0.4	223	SAM8885 P31BC		The upper ribosomal fragment has degraded

Peak Table

Size [nt]	Calibrated Conc. [ng/μl]	Assigned Conc. [ng/μl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1769	24.8	-	41.2	69.56	
5028	10.8	-	6.34	30.44	

H1: SAM8886 P32BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/ul]	Sample Description	Alert	Observations
H1	6.0	1.1	28.5	SAM8886 P32BC		

Peak Table

Size [nt]	Calibrated Conc. [ng/ul]	Assigned Conc. [ng/ul]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1675	1.18	-	2.08	47.95	
5171	1.28	-	0.730	52.05	

A2: SAM8887 P34BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/ul]	Sample Description	Alert	Observations
A2	6.8	0.6	374	SAM8887 P34BC		

Peak Table

Size [nt]	Calibrated Conc. [ng/ul]	Assigned Conc. [ng/ul]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1738	37.8	-	64.0	61.57	
4612	23.6	-	15.1	38.43	


B2: SAM8888 A**Sample Table**

Well	RINe	28S/18S (Area)	Conc. [ng/μl]	Sample Description	Alert	Observations
B2	6.2	0.7	12.1	SAM8888 A	⚠	Sample concentration outside functional range for RINe and the assay; The upper ribosomal fragment has degraded

Peak Table

Size [nt]	Calibrated Conc. [ng/μl]	Assigned Conc. [ng/μl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1689	1.20	-	2.09	57.76	
5090	0.876	-	0.506	42.24	

C2: SAM8890 P24BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/ul]	Sample Description	Alert	Observations
C2	8.6	1.6	326	SAM8890 P24BC		

Peak Table

Size [nt]	Calibrated Conc. [ng/ul]	Assigned Conc. [ng/ul]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1610	68.0	-	124	38.27	
3944	110	-	81.8	61.73	

D2: SAM8891 P26BC



Sample Table

Well	RINe	28S/18S (Area)	Conc. [ng/ul]	Sample Description	Alert	Observations
D2	8.5	1.4	324	SAM8891 P26BC		

Peak Table

Size [nt]	Calibrated Conc. [ng/ul]	Assigned Conc. [ng/ul]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment
25	36.0	36.0	4240	-	
1691	69.9	-	122	41.37	
4264	99.0	-	68.3	58.63	


A1	B1	C1	D1	E1	F1	G1
RIN ^e 1.6	RIN ^e -	RIN ^e -	RIN ^e -	RIN ^e -	RIN ^e 6.6	RIN ^e -

Default image (Contrast 19%), Image is Scaled to Highest Peak

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
A1	-	-	3750	Electronic Ladder		Ladder
B1	1.6	-	537	SAM8879 P38BC		Caution! Expired ScreenTape device; Markers outside standard running position; The lower ribosomal fragment is missing; RNA concentration outside recommended range for RINe; RINe edited
C1	-	-	137	SAM8880 P37BC		Caution! Expired ScreenTape device; Markers outside standard running position; Sample concentration outside functional range for RINe and the assay; RINe edited
D1	-	-	194	SAM8881 P36BC		Caution! Expired ScreenTape device; Markers outside standard running position; Sample concentration outside functional range for RINe and the assay; RINe edited
E1	-	-	125	SAM8883 P29BC		Caution! Expired ScreenTape device; Markers outside standard running position; Sample concentration outside functional range for RINe and the assay; RINe edited

F1	6.6	-	916	SAM8884 P30BC		Caution! Expired ScreenTape device; Markers outside standard running position; The lower ribosomal fragment is missing; RNA concentration outside recommended range for RINe; RINe edited
G1	-	-	237	SAM8889 P28BC		Caution! Expired ScreenTape device; Markers outside standard running position; Sample concentration outside functional range for RINe and the assay; RINe edited

A1: Electronic Ladder




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
A1	-	-	3750	Electronic Ladder		Ladder

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
200	189	-	2770	5.98		
500	638	-	3760	20.25		
1000	557	-	1640	17.66		
2000	580	-	854	18.42		
4000	634	-	466	20.11		
6000	554	-	272	17.58		

B1: SAM8879 P38BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
B1	1.6	-	537	SAM8879 P38BC		Caution! Expired ScreenTape device; Markers outside standard running position; The lower ribosomal fragment is missing; RNA concentration outside recommended range for RINe; RINe edited

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		edited Lower Marker

C1: SAM8880 P37BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
C1	-	-	137	SAM8880 P37BC		Caution! Expired ScreenTape device; Markers outside standard running position; Sample concentration outside functional range for RINe and the assay; RINe edited

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		edited Lower Marker
1007	8.13	-	23.7	55.98		18S edited
3217	6.39	-	5.84	44.02		28S edited

D1: SAM8881 P36BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
D1	-	-	194	SAM8881 P36BC		Caution! Expired ScreenTape device; Markers outside standard running position; Sample concentration outside functional range for RINe and the assay; RINe edited

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		edited Lower Marker
1066	16.6	-	45.8	100.00		18S edited

E1: SAM8883 P29BC




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
E1	-	-	125	SAM8883 P29BC		Caution! Expired ScreenTape device; Markers outside standard running position; Sample concentration outside functional range for RINe and the assay; RINe edited

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		edited Lower Marker
55	18.3	-	980	25.05		
90	7.16	-	233	9.80		
1018	19.8	-	57.3	27.16		18S edited
2903	27.8	-	28.1	38.00		28S edited


F1: SAM8884 P30BC**Sample Table**

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
F1	6.6	-	916	SAM8884 P30BC	⚠	Caution! Expired ScreenTape device; Markers outside standard running position; The lower ribosomal fragment is missing; RNA concentration outside recommended range for RINe; RINe edited

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		edited Lower Marker
57	138	-	7090	16.53		
96	63.8	-	1960	7.64		
185	83.2	-	1320	9.96		
250	29.3	-	346	3.51		
276	31.0	-	330	3.70		
366	54.9	-	442	6.57		
571	41.4	-	213	4.95		
863	58.2	-	198	6.96		
1021	134	-	386	16.01		
1782	35.7	-	59.0	4.28		
2670	166	-	183	19.88		

G1: SAM8889 P28BC



Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
G1	-	-	237	SAM8889 P28BC		Caution! Expired ScreenTape device; Markers outside standard running position; Sample concentration outside functional range for RINe and the assay; RINe edited

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		edited Lower Marker
1003	46.0	-	135	57.59		18S edited
2648	33.9	-	37.7	42.41		28S edited