## Beam Scheme v0.52

Results		
Protocol: BasePar	rams ~	
Parameters	X Profile	Y Profile
Field stats		
CAX dose	50392.00	50580.00
Max dose	52420.00	52770.00
Max pos (cm)	6.20	5.46
Min dose	50223.00	50358.00
Sum	24563902.00	24487380.00
Sum squares	126775273142	241 <b>26</b> 2495664554.00
Num points	476.00	475.00
Lata maratata di manana		
Interpolated params		0.00
90% Left (cm)	-9.88	-9.82
90% Right (cm)		9.82
80% Left (cm)	-9.95	-9.89
80% Right (cm)	9.88	9.89
50% Left (cm)	-10.05	-10.02
50% Right (cm)	9.98	9.99
20% Left (cm)	-10.25	-10.36
20% Right (cm)	10.19	10.37
10% Left (cm)	-10.68	-10.67
10% Right (cm)	10.61	10.69
Inflection point		
80% Left (cm)	-9.88	-9.79
80% Right (cm)	9.80	9.78
Inf pt left (cm)	-10.06	-10.04
Inf pt right (cm)	9.99	10.02
50% Left (cm)	-10.07	-10.07
50% Right (cm)	10.01	10.04
20% Left (cm)	-10.24	-10.30
20% Right (cm)	10.18	10.26
Slope left (%/cm)	84.86	61.32
Slope right (%/cm)	-80.69	-64.65
Dose point values		
Left 20% FW	51206.13	51216.34
Right 20% FW	51261.21	51337.14
Left 50% FW	51970.83	51969.36
Right 50% FW	51974.25	52109.71
Left 60% FW	52279.19	51869.08
Right 60% FW	52259.06	52171.85
Left 80% FW	51808.45	52209.81
Right 80% FW	51878.09	51847.01
Summatur:		
Symmetry Area left	514007 07	511900 12
Area left	514927.37	511809.12
Area right	511686.18	511660.79
Ratio Difference	1.01 338.00	1.01 691.00