

id	日期	类别	客户姓名	客户ID	内镜型号	内镜ID	预处理开始	预处理结束	预处理员	洗消步骤	洗消员
814	2019-10-11	自洗消									
815	2019-10-11	自洗消									
816	2019-10-12	内镜洗消			GIF-H290Z	2934702				测漏-08:15:28 初洗开始-08:15:29 初洗结束-08:52:00	袁芳
817	2019-10-12	内镜洗消			GIF-H290Z	2934702				测漏-08:52:15 初洗开始-08:52:16 机洗开始-08:52:19 机洗结束-09:16:54	袁芳
818	2019-10-12	内镜洗消			GIF-H290Z	2934696				测漏-08:18:42 初洗开始-08:18:43 机洗开始-08:18:46 机洗结束-09:01:51	袁芳
819	2019-10-12	内镜洗消			GIF-H290Z	2934698				测漏-08:22:32 初洗开始-08:22:33 机洗开始-08:22:36 机洗结束-09:04:21	袁芳
820	2019-10-12	内镜洗消			GIF-H290Z	2934705				测漏-08:26:53 初洗开始-08:26:54 机洗开始-08:26:58 机洗结束-09:15:01	袁芳
821	2019-10-12	内镜洗消			GIF-H290Z	2934702				测漏-08:52:15 初洗开始-08:52:16 机洗开始-08:52:19 机洗结束-09:16:54	袁芳
822	2019-10-12	内镜洗消			CF-HQ290ZI	2922160				测漏-09:04:15 初洗开始-09:04:16 机洗开始-09:04:18 机洗结束-09:45:44	袁芳
823	2019-10-12	内镜洗消			CF-HQ290ZI	2922153				测漏-09:09:46 初洗开始-09:09:47 机洗开始-09:09:50 机洗结束-09:47:24	袁芳
824	2019-10-12	内镜洗消			CF-HQ290ZI	2922161				测漏-09:22:40 初洗开始-09:22:41 机洗开始-09:22:51 机洗结束-10:06:09	袁芳
825	2019-10-12	内镜洗消			CF-HQ290ZI	2922164				测漏-09:21:14 初洗开始-09:21:15 机洗开始-09:21:22 机洗结束-10:07:09	袁芳
826	2019-10-15	内镜洗消	方华	1908140001	GIF-H290Z	2934698	09:36:24	09:36:58	阴雪燕	测漏-07:38:08 初洗开始-07:38:09 机洗开始-07:38:11 机洗结束-08:10:29	张磊
827	2019-10-15	内镜洗消	方华	1908140001	CF-HQ290ZI	2922160	09:50:50	09:51:31	阴雪燕	测漏-07:39:02 初洗开始-07:39:03 机洗开始-07:39:11 机洗结束-08:10:10	杨荣江
828	2019-10-15	内镜洗消	孟凡玲	1909230001	GIF-H290Z	2934696	11:31:10	11:31:48	阴雪燕	测漏-07:41:28 初洗开始-07:41:28 机洗开始-07:41:56 机洗结束-08:16:00	张磊
829	2019-10-15	内镜洗消	孟凡玲	1909230001	CF-HQ290ZI	2922164	11:47:39	11:48:34	阴雪燕	测漏-07:42:47 初洗开始-07:42:48 机洗开始-07:42:53 机洗结束-08:15:36	杨荣江
830	2019-10-15	内镜洗消	徐鹏	1909240001	GIF-H290Z	2934705	10:56:25	10:56:59	阴雪燕	测漏-08:11:35 初洗开始-08:11:36 机洗开始-08:11:39 机洗结束-08:39:55	张磊
831	2019-10-15	内镜洗消	李凌波	1908140004	CF-HQ290ZI	2922161	10:23:01	10:23:42	阴雪燕	测漏-08:17:25 初洗开始-08:17:26 机洗开始-08:17:28 机洗结束-08:40:17	张磊
832	2019-10-15	内镜洗消	徐鹏	1909240001	CF-HQ290ZI	2922153	11:04:02	11:04:51	阴雪燕	测漏-08:18:07 初洗开始-08:18:08 机洗开始-08:18:12 机洗结束-08:43:33	张磊
833	2019-10-15	内镜洗消	李凌波	1908140004	GIF-H290Z	2934702	10:13:09	10:13:41	阴雪燕	测漏-08:20:02 初洗开始-08:20:03 机洗开始-08:20:06 机洗结束-08:45:32	张磊
834	2019-10-15	内镜洗消			GIF-H290Z	2934698				测漏-09:39:24 初洗开始-09:43:36 机洗开始-09:52:24 机洗结束-10:37:00	张磊
835	2019-10-15	内镜洗消			CF-HQ290ZI	2922160				测漏-09:55:39 初洗开始-10:00:29 机洗开始-10:08:27 机洗结束-10:48:44	张磊
836	2019-10-15	内镜洗消			GIF-H290Z	2934702				测漏-10:16:11 初洗开始-10:19:22 机洗开始-10:26:28 机洗结束-10:26:28	张磊
837	2019-10-15	内镜洗消			CF-HQ290ZI	2922161				测漏-10:30:22 初洗开始-10:34:27 机洗开始-10:42:00 机洗结束-11:21:07	张磊
838	2019-10-15	内镜洗消			GIF-H290Z	2934705				测漏-11:02:14 初洗开始-11:05:32 机洗开始-11:14:07 机洗结束-11:53:37	张磊
839	2019-10-15	内镜洗消			CF-HQ290ZI	2922153				测漏-11:16:06 初洗开始-11:20:28 机洗开始-11:28:33 机洗结束-12:07:10	张磊
840	2019-10-15	内镜洗消			GIF-H290Z	2934696				测漏-11:34:25 初洗开始-11:38:26 机洗开始-11:48:00 机洗结束-12:26:29	张磊
841	2019-10-15	内镜洗消			CF-HQ290ZI	2922164				测漏-11:50:31 初洗开始-11:54:06 机洗开始-12:01:52 机洗结束-12:47:15	张磊

id	814	id	815	id	816	id	817	id	818	id	819
fileId	1570759221518	fileId	1570759244233	fileId	1570840308321	fileId	1570841421227	fileId	1570842012360	fileId	1570842127959
Category	自洗消	Category	自洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消
CycleCompletionDate	2019-10-11	CycleCompletionDate	2019-10-11	CycleCompletionDate	2019-10-12	CycleCompletionDate	2019-10-12	CycleCompletionDate	2019-10-12	CycleCompletionDate	2019-10-12
TimeBegin	07:50:53	TimeBegin	07:52:16	TimeBegin	08:18:00	TimeBegin	08:49:48	TimeBegin	08:21:09	TimeBegin	08:27:18
TimeEnd	09:53:48	TimeEnd	09:56:04	TimeEnd	08:29:04	TimeEnd	08:49:58	TimeEnd	08:56:34	TimeEnd	09:02:43
CYCLE	DISINFECTED	CYCLE	DISINFECTED	CYCLE	FAIL	CYCLE	FAIL	CYCLE	PASS	CYCLE	PASS
MachineSerialNumber	92773003, Left	MachineSerialNumber	92773002, Left	MachineSerialNumber	92773003, Left	MachineSerialNumber	92773003, Left	MachineSerialNumber	92773003, Right	MachineSerialNumber	92773002, Left
Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital
Endoscope		Endoscope		Endoscope	104	Endoscope	104	Endoscope	101	Endoscope	102
EndoscopeType		EndoscopeType		EndoscopeType	GIF-H290Z	EndoscopeType	GIF-H290Z	EndoscopeType	GIF-H290Z	EndoscopeType	GIF-H290Z
InternalID		InternalID		InternalID	2934702	InternalID	2934702	InternalID	2934696	InternalID	2934698
SerialNumber		SerialNumber		SerialNumber	2934702	SerialNumber	2934702	SerialNumber	2934696	SerialNumber	2934698
Hookup		Hookup		Hookup	2-8-611 1LPT	Hookup	2-8-611 1LPT	Hookup	2-8-611 1K9H	Hookup	2-8-611 1LPK
ParameterSet	S-WLDisinfect-2hr	ParameterSet	S-WLDisinfect-2hr	ParameterSet	1-39-603 A	ParameterSet	1-35-603 C	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A
CycleNumber	198	CycleNumber	140	CycleNumber	199	CycleNumber	201	CycleNumber	160	CycleNumber	141
OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE
OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE
MRCValidation		MRCValidation		MRCValidation		MRCValidation		MRCValidation	Passed	MRCValidation	Passed
07:50:53 1 模块		07:52:16 1 模块		08:18:00 1 测漏		08:49:48 1 测漏		08:21:09 1 测漏		08:27:18 1 测漏	
07:50:53 1 测漏		07:52:16 1 测漏		08:18:28 2 冲洗		08:49:56 1 (-360) 被操作者停止		08:21:36 2 冲洗		08:27:47 2 冲洗	
07:51:48 1 连接性检测		07:53:12 1 连接性检测		08:22:37 3 清洗剂清洗		08:49:57 14 吹干		08:25:32 3 清洗剂清洗		08:31:38 3 清洗剂清洗	
07:52:18 1 双槽排水		07:53:42 1 双槽排水		08:26:04 3 (-300) SCU F-075/S035 (Minimum spray arm speed)		08:49:57 14 (-104) 内镜轻微漏气,循环继续平稳进行		08:30:10 3 清洁剂接触时间 = 00:02:06		08:36:13 3 清洁剂接触时间 = 00:02:06	
07:53:34 2 A/B液注入定量器		07:54:55 2 A/B液制备		08:26:04 3 清洁剂接触时间 = 00:02:06		08:49:58 循环结束		08:30:10 3 槽内温度 [°C] = 24.00		08:36:13 3 槽内温度 [°C] = 24.10	
07:53:34 2 A/B液制备		07:54:55 2 A/B液注入定量器		08:26:04 3 槽内温度 [°C] = 24.50				08:30:10 4 冲洗		08:36:13 4 冲洗	
07:54:02 2 A/B液配比混匀		07:55:22 2 A/B液配比混匀		08:26:04 14 吹干				08:30:10 5 漂洗		08:36:13 5 漂洗	
07:54:02 2 注入并检测腔道 (全部)		07:55:22 2 注入并检测腔道 (全部)		08:29:04 循环结束				08:36:20 5 清洗接触时间 = 00:02:06		08:42:23 5 清洗接触时间 = 00:02:06	
07:54:38 2 清洗槽注入		07:56:03 2 清洗槽注入						08:36:20 5 槽内温度 [°C] = 26.10		08:42:23 5 槽内温度 [°C] = 26.20	
07:56:25 2 A/B液注入定量器		07:57:59 2 A/B液注入定量器						08:36:20 6 清洗剂清洗		08:42:23 6 清洗剂清洗	
07:57:22 2 等待化学制剂入水		07:58:55 2 等待化学制剂入水						08:36:22 7 冲洗		08:42:24 7 冲洗	
07:57:22 2 等待断开供水		07:58:55 2 等待断开供水						08:36:22 8 漂洗		08:42:25 8 漂洗	
07:57:22 2 通过腔道液体混匀		07:58:55 2 通过腔道液体混匀						08:36:23 9 消毒		08:42:26 9 消毒	
07:59:10 2 消毒供水管路和水过滤器		08:00:43 2 消毒供水管路和水过滤器						08:45:38 9 消毒接触时间 = 00:05:15		08:51:47 9 消毒接触时间 = 00:05:15	
08:31:59 2 消毒清洗槽		08:33:39 2 消毒清洗槽						08:45:38 9 槽内温度 [°C] = 26.00		08:51:47 9 槽内温度 [°C] = 27.00	
09:32:55 2 双槽排水		09:34:44 2 双槽排水						08:45:38 10 冲洗		08:51:47 10 冲洗	
09:34:43 2 等待连接供水		09:36:40 2 等待连接供水						08:47:34 11 漂洗		08:53:36 11 漂洗	
09:34:43 2 漂洗供水管路		09:36:40 2 漂洗供水管路						08:47:35 12 漂洗		08:53:37 12 漂洗	
09:38:12 2 双槽排水		09:40:21 2 双槽排水						08:47:35 13 未洗		08:53:38 13 未洗	
09:39:03 2 冲洗腔道		09:41:12 2 冲洗腔道						08:52:02 13 清洗接触时间 = 00:02:22		08:58:11 13 清洗接触时间 = 00:02:22	
09:41:03 2 双槽排水		09:43:13 2 双槽排水						08:52:02 13 槽内温度 [°C] = 25.30		08:58:11 13 槽内温度 [°C] = 27.20	
09:41:56 2 漂洗清洗槽		09:44:07 2 漂洗清洗槽						08:52:03 14 吹干		08:58:11 14 吹干	
09:50:23 2 双槽排水		09:52:46 2 双槽排水						08:56:33 14 酒精吹干		09:02:42 14 酒精吹干	
09:52:53 3 双槽排水		09:55:09 3 双槽排水						08:56:34 循环结束		09:02:43 循环结束	
09:52:53 3 吹干		09:55:09 3 吹干									
09:53:38 3 吹干		09:55:54 3 吹干									
09:53:48 循环结束		09:56:04 循环结束									

id	820	id	821	id	822	id	823	id	824	id	825
fileId	1570842508735	fileId	1570842988773	fileId	1570844685897	fileId	1570844793131	fileId	1570845785569	fileId	1570845818778
Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消
CycleCompletionDate	2019-10-12	CycleCompletionDate	2019-10-12	CycleCompletionDate	2019-10-12	CycleCompletionDate	2019-10-12	CycleCompletionDate	2019-10-12	CycleCompletionDate	2019-10-12
TimeBegin	08:31:32	TimeBegin	08:52:36	TimeBegin	09:08:14	TimeBegin	09:11:18	TimeBegin	09:27:22	TimeBegin	09:26:29
TimeEnd	09:07:18	TimeEnd	09:16:06	TimeEnd	09:43:44	TimeEnd	09:46:37	TimeEnd	10:02:57	TimeEnd	10:03:07
CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS
MachineSerialNumber	92773002, Right	MachineSerialNumber	92773003, Left	MachineSerialNumber	92773003, Right	MachineSerialNumber	92773002, Left	MachineSerialNumber	92773002, Right	MachineSerialNumber	92773003, Left
Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital
Endoscope	103	Endoscope	104	Endoscope	105	Endoscope	108	Endoscope	106	Endoscope	107
EndoscopeType	GIF-H290Z	EndoscopeType	GIF-H290Z	EndoscopeType	CF-HQ290ZI	EndoscopeType	CF-HQ290ZI	EndoscopeType	CF-HQ290ZI	EndoscopeType	CF-HQ290ZI
InternalID	2934705	InternalID	2934702	InternalID	2922160	InternalID	2922153	InternalID	2922161	InternalID	2922164
SerialNumber	2934705	SerialNumber	2934702	SerialNumber	2922160	SerialNumber	2922153	SerialNumber	2922161	SerialNumber	2922164
Hookup	2-8-611 1LPI	Hookup	2-8-611 1LPT	Hookup	2-8-611 1K9H	Hookup	2-8-611 1LPK	Hookup	2-8-611 1LPI	Hookup	2-8-611 1LPT
ParameterSet	1-39-603 A	ParameterSet	1-35-603 C	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A
CycleNumber	110	CycleNumber	202	CycleNumber	161	CycleNumber	142	CycleNumber	111	CycleNumber	203
OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE
OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE
MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed
08:31:32	1 测漏	08:52:36	1 测漏	09:08:14	1 测漏	09:11:18	1 测漏	09:27:22	1 测漏	09:26:29	1 测漏
08:32:00	2 冲洗	08:52:53	2 冲洗	09:08:43	2 冲洗	09:11:48	2 冲洗	09:27:53	2 冲洗	09:26:58	2 冲洗
08:35:57	3 清洗剂清洗	08:57:00	3 清洗剂清洗	09:12:31	3 清洗剂清洗	09:15:42	3 清洗剂清洗	09:31:39	3 清洗剂清洗	09:31:12	3 清洗剂清洗
08:40:37	3 清洁剂接触时间 = 00:02:06	08:57:01	4 冲洗	09:17:25	3 清洁剂接触时间 = 00:02:06	09:20:26	3 清洁剂接触时间 = 00:02:06	09:36:17	3 清洁剂接触时间 = 00:02:06	09:36:06	3 清洁剂接触时间 = 00:02:06
08:40:37	3 槽内温度 [°C] = 26.10	08:57:02	5 漂洗	09:17:25	3 槽内温度 [°C] = 25.20	09:20:26	3 槽内温度 [°C] = 27.10	09:36:17	3 槽内温度 [°C] = 27.10	09:36:06	3 槽内温度 [°C] = 25.60
08:40:37	4 冲洗	08:57:03	6 清洗剂清洗	09:17:25	4 冲洗	09:20:26	4 冲洗	09:36:17	4 冲洗	09:36:06	4 冲洗
08:40:37	5 漂洗	08:57:04	7 冲洗	09:17:26	5 漂洗	09:20:27	5 漂洗	09:36:18	5 漂洗	09:36:06	5 漂洗
08:46:42	5 清洗接触时间 = 00:02:06	08:57:04	8 漂洗	09:23:24	5 清洗接触时间 = 00:02:06	09:26:28	5 清洗接触时间 = 00:02:06	09:42:27	5 清洗接触时间 = 00:02:17	09:42:16	5 清洗接触时间 = 00:02:06
08:46:42	5 槽内温度 [°C] = 26.70	08:57:05	9 消毒	09:23:24	5 槽内温度 [°C] = 25.10	09:26:28	5 槽内温度 [°C] = 27.20	09:42:27	5 槽内温度 [°C] = 27.30	09:42:16	5 槽内温度 [°C] = 25.60
08:46:42	6 清洗剂清洗	09:06:42	9 消毒接触时间 = 00:05:15	09:23:24	6 清洗剂清洗	09:26:28	6 清洗剂清洗	09:42:27	6 清洗剂清洗	09:42:16	6 清洗剂清洗
08:46:44	7 冲洗	09:06:42	9 槽内温度 [°C] = 25.30	09:23:25	7 冲洗	09:26:29	7 冲洗	09:42:28	7 冲洗	09:42:17	7 冲洗
08:46:44	8 漂洗	09:06:42	10 冲洗	09:23:26	8 漂洗	09:26:30	8 漂洗	09:42:29	8 漂洗	09:42:17	8 漂洗
08:46:45	9 消毒	09:08:37	11 漂洗	09:23:27	9 消毒	09:26:31	9 消毒	09:42:29	9 消毒	09:42:18	9 消毒
08:56:11	9 消毒接触时间 = 00:05:15	09:08:38	12 漂洗	09:32:53	9 消毒接触时间 = 00:05:15	09:35:48	9 消毒接触时间 = 00:05:15	09:51:56	9 消毒接触时间 = 00:05:15	09:51:57	9 消毒接触时间 = 00:05:15
08:56:11	9 槽内温度 [°C] = 27.20	09:08:38	13 未洗	09:32:53	9 槽内温度 [°C] = 25.60	09:35:48	9 槽内温度 [°C] = 27.30	09:51:56	9 槽内温度 [°C] = 27.30	09:51:57	9 槽内温度 [°C] = 25.50
08:56:11	10 冲洗	09:13:18	13 清洗接触时间 = 00:02:20	09:32:53	10 冲洗	09:35:48	10 冲洗	09:51:56	10 冲洗	09:51:57	10 冲洗
08:58:04	11 漂洗	09:13:18	13 槽内温度 [°C] = 25.50	09:34:46	11 漂洗	09:37:31	11 漂洗	09:53:46	11 漂洗	09:53:55	11 漂洗
08:58:05	12 漂洗	09:13:18	14 吹干	09:34:47	12 漂洗	09:37:32	12 漂洗	09:53:47	12 漂洗	09:53:55	12 漂洗
08:58:06	13 未洗	09:16:06	循环结束	09:34:47	13 未洗	09:37:33	13 未洗	09:53:47	13 未洗	09:53:56	13 未洗
09:02:42	13 清洗接触时间 = 00:02:22	09:39:14	13 清洗接触时间 = 00:02:22	09:39:14	13 清洗接触时间 = 00:02:22	09:42:05	13 清洗接触时间 = 00:02:20	09:58:25	13 清洗接触时间 = 00:02:22	09:58:34	13 清洗接触时间 = 00:02:22
09:02:42	13 槽内温度 [°C] = 27.30	09:39:14	13 槽内温度 [°C] = 25.50	09:39:14	13 槽内温度 [°C] = 25.50	09:42:05	13 槽内温度 [°C] = 27.30	09:58:25	13 槽内温度 [°C] = 27.20	09:58:34	13 槽内温度 [°C] = 25.70
09:02:43	14 吹干	09:39:14	14 吹干	09:39:14	14 吹干	09:42:05	14 吹干	09:58:25	14 吹干	09:58:34	14 吹干
09:07:17	14 酒精吹干	09:43:43	14 酒精吹干	09:43:43	14 酒精吹干	09:46:36	14 酒精吹干	10:02:56	14 酒精吹干	10:03:06	14 酒精吹干
09:07:18	循环结束	09:43:44	循环结束	09:43:44	循环结束	09:46:37	循环结束	10:02:57	循环结束	10:03:07	循环结束

id	826	id	827	id	828	id	829	id	830	id	831
fileId	1571098165683	fileId	1571098171720	fileId	1571098325037	fileId	1571098332291	fileId	1571099954756	fileId	1571100004458
Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消
CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15
TimeBegin	07:42:08	TimeBegin	07:41:30	TimeBegin	07:48:51	TimeBegin	07:48:41	TimeBegin	08:16:06	TimeBegin	08:16:33
TimeEnd	08:08:35	TimeEnd	08:08:59	TimeEnd	08:11:57	TimeEnd	08:12:18	TimeEnd	08:38:47	TimeEnd	08:40:04
CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS
MachineSerialNumber	92773003, Left	MachineSerialNumber	92773003, Right	MachineSerialNumber	92773002, Left	MachineSerialNumber	92773002, Right	MachineSerialNumber	92773003, Right	MachineSerialNumber	92773003, Left
Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital
Endoscope	102	Endoscope	105	Endoscope	101	Endoscope	107	Endoscope	103	Endoscope	106
EndoscopeType	GIF-H290Z	EndoscopeType	CF-HQ290ZI	EndoscopeType	GIF-H290Z	EndoscopeType	CF-HQ290ZI	EndoscopeType	GIF-H290Z	EndoscopeType	CF-HQ290ZI
InternalID	2934698	InternalID	2922160	InternalID	2934696	InternalID	2922164	InternalID	2934705	InternalID	2922161
SerialNumber	2934698	SerialNumber	2922160	SerialNumber	2934696	SerialNumber	2922164	SerialNumber	2934705	SerialNumber	2922161
Hookup	2-8-611 1LPI	Hookup	2-8-611 1LPT	Hookup	2-8-611 1LPK	Hookup	2-8-611 1K9H	Hookup	2-8-611 1LPT	Hookup	2-8-611 1LPI
ParameterSet	1-35-603 C	ParameterSet	1-35-603 C	ParameterSet	1-35-603 C	ParameterSet	1-35-603 C	ParameterSet	1-35-603 C	ParameterSet	1-35-603 C
CycleNumber	204	CycleNumber	162	CycleNumber	143	CycleNumber	112	CycleNumber	163	CycleNumber	205
OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE
OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE
MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed
07:42:08 1	测漏	07:41:30 1	测漏	07:48:51 1	测漏	07:48:41 1	测漏	08:16:06 1	测漏	08:16:33 1	测漏
07:42:38 2	冲洗	07:42:01 2	冲洗	07:49:22 2	冲洗	07:49:14 2	冲洗	08:16:22 2	冲洗	08:17:05 2	冲洗
07:43:45 2	(~283) 入口水流过低	07:43:46 2	(~283) 入口水流过低	07:53:18 3	清洗剂清洗	07:53:16 3	清洗剂清洗	08:20:06 3	清洗剂清洗	08:20:58 3	清洗剂清洗
07:49:41 3	清洗剂清洗	07:49:50 3	清洗剂清洗	07:53:19 4	冲洗	07:53:17 4	冲洗	08:20:07 4	冲洗	08:21:00 4	冲洗
07:49:42 4	冲洗	07:49:51 4	冲洗	07:53:19 5	漂洗	07:53:18 5	漂洗	08:20:07 5	漂洗	08:21:00 5	漂洗
07:49:42 5	漂洗	07:49:51 5	漂洗	07:53:20 6	清洗剂清洗	07:53:19 6	清洗剂清洗	08:20:08 6	清洗剂清洗	08:21:01 6	清洗剂清洗
07:49:43 6	清洗剂清洗	07:49:52 6	清洗剂清洗	07:53:21 7	冲洗	07:53:20 7	冲洗	08:20:09 7	冲洗	08:21:02 7	冲洗
07:49:44 7	冲洗	07:49:53 7	冲洗	07:53:21 8	漂洗	07:53:20 8	漂洗	08:20:10 8	漂洗	08:21:03 8	漂洗
07:49:45 8	漂洗	07:49:54 8	漂洗	07:53:22 9	消毒	07:53:21 9	消毒	08:20:11 9	消毒	08:21:04 9	消毒
07:49:46 9	消毒	07:49:55 9	消毒	08:02:53 9	消毒接触时间 = 00:05:15	08:03:05 9	消毒接触时间 = 00:05:15	08:29:24 9	消毒接触时间 = 00:05:15	08:30:51 9	消毒接触时间 = 00:05:15
07:59:25 9	消毒接触时间 = 00:05:15	07:59:43 9	消毒接触时间 = 00:05:15	08:02:53 9	槽内温度 [°C] = 25.00	08:03:05 9	槽内温度 [°C] = 25.20	08:29:24 9	槽内温度 [°C] = 25.20	08:30:51 9	槽内温度 [°C] = 25.30
07:59:25 9	槽内温度 [°C] = 24.00	07:59:43 9	槽内温度 [°C] = 24.00	08:02:53 10	冲洗	08:03:05 10	冲洗	08:29:24 10	冲洗	08:30:51 10	冲洗
07:59:25 10	冲洗	07:59:44 10	冲洗	08:04:35 11	漂洗	08:04:49 11	漂洗	08:31:19 11	漂洗	08:32:36 11	漂洗
08:01:06 11	漂洗	08:01:30 11	漂洗	08:04:36 12	漂洗	08:04:50 12	漂洗	08:31:20 12	漂洗	08:32:37 12	漂洗
08:01:07 12	漂洗	08:01:31 12	漂洗	08:04:37 13	未洗	08:04:51 13	未洗	08:31:21 13	未洗	08:32:38 13	未洗
08:01:08 13	未洗	08:01:31 13	未洗	08:09:10 13	清洗接触时间 = 00:02:20	08:09:30 13	清洗接触时间 = 00:02:20	08:35:59 13	清洗接触时间 = 00:02:22	08:37:17 13	清洗接触时间 = 00:02:22
08:05:47 13	清洗接触时间 = 00:02:22	08:06:10 13	清洗接触时间 = 00:02:22	08:09:10 13	槽内温度 [°C] = 26.30	08:09:30 13	槽内温度 [°C] = 26.30	08:35:59 13	槽内温度 [°C] = 25.30	08:37:17 13	槽内温度 [°C] = 25.60
08:05:47 13	槽内温度 [°C] = 24.70	08:06:10 13	槽内温度 [°C] = 24.60	08:09:10 14	吹干	08:09:30 14	吹干	08:35:59 14	吹干	08:37:17 14	吹干
08:05:47 14	吹干	08:06:10 14	吹干	08:11:57	循环结束	08:12:18	循环结束	08:38:47	循环结束	08:40:04	循环结束
08:08:35	循环结束	08:08:59	循环结束								

id	832	id	833	id	834	id	835	id	836	id	837
fileId	1571100182688	fileId	1571100298986	fileId	1571106946435	fileId	1571107673764	fileId	1571108854569	fileId	1571109592007
Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消
CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15
TimeBegin	08:21:12	TimeBegin	08:23:09	TimeBegin	09:55:09	TimeBegin	10:11:07	TimeBegin	10:31:37	TimeBegin	10:44:24
TimeEnd	08:44:15	TimeEnd	08:46:19	TimeEnd	10:31:36	TimeEnd	10:46:45	TimeEnd	11:07:12	TimeEnd	11:20:32
CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS
MachineSerialNumber	92773002, Left	MachineSerialNumber	92773002, Right	MachineSerialNumber	92773003, Left	MachineSerialNumber	92773003, Right	MachineSerialNumber	92773002, Left	MachineSerialNumber	92773002, Right
Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital
Endoscope	108	Endoscope	104	Endoscope	102	Endoscope	105	Endoscope	104	Endoscope	106
EndoscopeType	CF-HQ290ZI	EndoscopeType	GIF-H290Z	EndoscopeType	GIF-H290Z	EndoscopeType	CF-HQ290ZI	EndoscopeType	GIF-H290Z	EndoscopeType	CF-HQ290ZI
InternalID	2922153	InternalID	2934702	InternalID	2934698	InternalID	2922160	InternalID	2934702	InternalID	2922161
SerialNumber	2922153	SerialNumber	2934702	SerialNumber	2934698	SerialNumber	2922160	SerialNumber	2934702	SerialNumber	2922161
Hookup	2-8-611 1LPK	Hookup	2-8-611 1K9H	Hookup	2-8-611 1LPI	Hookup	2-8-611 1LPT	Hookup	2-8-611 1LPK	Hookup	2-8-611 1K9H
ParameterSet	1-35-603 C	ParameterSet	1-35-603 C	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A
CycleNumber	144	CycleNumber	113	CycleNumber	206	CycleNumber	164	CycleNumber	145	CycleNumber	114
OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE
OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE
MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed
08:21:12 1	测漏	08:23:09 1	测漏	09:55:14 1	测漏	10:11:07 1	测漏	10:31:42 1	测漏	10:44:24 1	测漏
08:21:46 2	冲洗	08:23:40 2	冲洗	09:55:42 2	冲洗	10:11:37 2	冲洗	10:32:11 2	冲洗	10:44:54 2	冲洗
08:25:40 3	清洗剂清洗	08:27:38 3	清洗剂清洗	09:59:48 3	清洗剂清洗	10:15:18 3	清洗剂清洗	10:36:10 3	清洗剂清洗	10:49:03 3	清洗剂清洗
08:25:41 4	冲洗	08:27:39 4	冲洗	10:04:42 3	清洁剂接触时间 = 00:02:06	10:20:16 3	清洁剂接触时间 = 00:02:06	10:40:50 3	清洁剂接触时间 = 00:02:06	10:53:46 3	清洁剂接触时间 = 00:02:06
08:25:42 5	漂洗	08:27:40 5	漂洗	10:04:42 3	槽内温度 [°C] = 24.60	10:20:16 3	槽内温度 [°C] = 24.30	10:40:50 3	槽内温度 [°C] = 25.70	10:53:46 3	槽内温度 [°C] = 26.80
08:25:43 6	清洗剂清洗	08:27:41 6	清洗剂清洗	10:04:42 4	冲洗	10:20:16 4	冲洗	10:40:50 4	冲洗	10:53:46 4	冲洗
08:25:44 7	冲洗	08:27:42 7	冲洗	10:04:42 5	漂洗	10:20:17 5	漂洗	10:40:51 5	漂洗	10:53:47 5	漂洗
08:25:44 8	漂洗	08:27:42 8	漂洗	10:11:08 5	清洗接触时间 = 00:02:06	10:26:21 5	清洗接触时间 = 00:02:06	10:46:54 5	清洗接触时间 = 00:02:06	11:00:09 5	清洗接触时间 = 00:02:06
08:25:45 9	消毒	08:27:43 9	消毒	10:11:08 5	槽内温度 [°C] = 24.70	10:26:21 5	槽内温度 [°C] = 24.50	10:46:54 5	槽内温度 [°C] = 26.70	11:00:09 5	槽内温度 [°C] = 27.30
08:34:58 9	消毒接触时间 = 00:05:15	08:37:14 9	消毒接触时间 = 00:05:15	10:11:08 6	清洗剂清洗	10:26:21 6	清洗剂清洗	10:46:54 6	清洗剂清洗	11:00:09 6	清洗剂清洗
08:34:58 9	槽内温度 [°C] = 26.60	08:37:14 9	槽内温度 [°C] = 26.70	10:11:10 7	冲洗	10:26:22 7	冲洗	10:46:55 7	冲洗	11:00:10 7	冲洗
08:34:58 10	冲洗	08:37:14 10	冲洗	10:11:10 8	漂洗	10:26:22 8	漂洗	10:46:55 8	漂洗	11:00:10 8	漂洗
08:36:49 11	漂洗	08:38:56 11	漂洗	10:11:11 9	消毒	10:26:23 9	消毒	10:46:56 9	消毒	11:00:11 9	消毒
08:36:50 12	漂洗	08:38:57 12	漂洗	10:20:34 9	消毒接触时间 = 00:05:15	10:35:43 9	消毒接触时间 = 00:05:15	10:56:14 9	消毒接触时间 = 00:05:15	11:09:28 9	消毒接触时间 = 00:05:15
08:36:51 13	未洗	08:38:58 13	未洗	10:20:34 9	槽内温度 [°C] = 25.30	10:35:43 9	槽内温度 [°C] = 25.50	10:56:14 9	槽内温度 [°C] = 27.20	11:09:28 9	槽内温度 [°C] = 27.50
08:41:27 13	清洗接触时间 = 00:02:22	08:43:33 13	清洗接触时间 = 00:02:22	10:20:34 10	冲洗	10:35:43 10	冲洗	10:56:14 10	冲洗	11:09:28 10	冲洗
08:41:27 13	槽内温度 [°C] = 27.00	08:43:33 13	槽内温度 [°C] = 27.10	10:22:17 11	漂洗	10:37:38 11	漂洗	10:58:06 11	漂洗	11:11:22 11	漂洗
08:41:27 14	吹干	08:43:33 14	吹干	10:22:18 12	漂洗	10:37:39 12	漂洗	10:58:06 12	漂洗	11:11:23 12	漂洗
08:44:15	循环结束	08:46:19	循环结束	10:22:19 13	未洗	10:37:40 13	未洗	10:58:07 13	未洗	11:11:24 13	未洗
				10:27:05 13	清洗接触时间 = 00:02:22	10:42:13 13	清洗接触时间 = 00:02:22	11:02:41 13	清洗接触时间 = 00:02:22	11:16:00 13	清洗接触时间 = 00:02:22
				10:27:05 13	槽内温度 [°C] = 25.80	10:42:13 13	槽内温度 [°C] = 25.10	11:02:41 13	槽内温度 [°C] = 27.70	11:16:00 13	槽内温度 [°C] = 27.30
				10:27:05 14	吹干	10:42:13 14	吹干	11:02:41 14	吹干	11:16:00 14	吹干
				10:31:35 14	酒精吹干	10:46:44 14	酒精吹干	11:07:11 14	酒精吹干	11:20:31 14	酒精吹干
				10:31:36	循环结束	10:46:45	循环结束	11:07:12	循环结束	11:20:32	循环结束

id	838	id	839	id	840	id	841
fileId	1571111552605	fileId	1571112383789	fileId	1571113543944	fileId	1571114770714
Category	内镜洗消	Category	内镜洗消	Category	内镜洗消	Category	内镜洗消
CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15	CycleCompletionDate	2019-10-15
TimeBegin	11:15:32	TimeBegin	11:30:14	TimeBegin	11:50:53	TimeBegin	12:03:29
TimeEnd	11:52:22	TimeEnd	12:05:32	TimeEnd	12:26:32	TimeEnd	12:39:24
CYCLE	PASS	CYCLE	PASS	CYCLE	PASS	CYCLE	PASS
MachineSerialNumber	92773003, Left	MachineSerialNumber	92773003, Right	MachineSerialNumber	92773002, Left	MachineSerialNumber	92773003, Left
Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital	Location	HUA SAN Hospital
Endoscope	103	Endoscope	108	Endoscope	101	Endoscope	107
EndoscopeType	GIF-H290Z	EndoscopeType	CF-HQ290ZI	EndoscopeType	GIF-H290Z	EndoscopeType	CF-HQ290ZI
InternalID	2934705	InternalID	2922153	InternalID	2934696	InternalID	2922164
SerialNumber	2934705	SerialNumber	2922153	SerialNumber	2934696	SerialNumber	2922164
Hookup	2-8-611 1LPI	Hookup	2-8-611 1LPT	Hookup	2-8-611 1LPK	Hookup	2-8-611 1LPI
ParameterSet	1-39-603 A	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A	ParameterSet	1-39-603 A
CycleNumber	207	CycleNumber	165	CycleNumber	146	CycleNumber	208
OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE	OperatorLoading	MVFSE
OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE	OperatorUnloading	MVFSE
MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed	MRCValidation	Passed
11:15:37	1 测漏	11:30:14	1 测漏	11:50:53	1 测漏	12:03:29	1 测漏
11:16:05	2 冲洗	11:30:42	2 冲洗	11:51:23	2 冲洗	12:03:58	2 冲洗
11:20:11	3 清洗剂清洗	11:34:31	3 清洗剂清洗	11:55:18	3 清洗剂清洗	12:08:17	3 清洗剂清洗
11:25:07	3 清洁剂接触时间 = 00:02:06	11:39:14	3 清洁剂接触时间 = 00:02:06	12:00:08	3 清洁剂接触时间 = 00:02:06	12:12:55	3 清洁剂接触时间 = 00:02:06
11:25:07	3 槽内温度 [°C] = 25.10	11:39:14	3 槽内温度 [°C] = 24.80	12:00:08	3 槽内温度 [°C] = 26.70	12:12:55	3 槽内温度 [°C] = 25.30
11:25:07	4 冲洗	11:39:14	4 冲洗	12:00:08	4 冲洗	12:12:55	4 冲洗
11:25:07	5 漂洗	11:39:15	5 漂洗	12:00:09	5 漂洗	12:12:55	5 漂洗
11:31:30	5 清洗接触时间 = 00:02:06	11:45:10	5 清洗接触时间 = 00:02:06	12:06:11	5 清洗接触时间 = 00:02:06	12:18:56	5 清洗接触时间 = 00:02:06
11:31:30	5 槽内温度 [°C] = 25.70	11:45:10	5 槽内温度 [°C] = 24.70	12:06:11	5 槽内温度 [°C] = 26.80	12:18:56	5 槽内温度 [°C] = 25.60
11:31:30	6 清洗剂清洗	11:45:10	6 清洗剂清洗	12:06:11	6 清洗剂清洗	12:18:56	6 清洗剂清洗
11:31:31	7 冲洗	11:45:11	7 冲洗	12:06:12	7 冲洗	12:18:57	7 冲洗
11:31:31	8 漂洗	11:45:11	8 漂洗	12:06:13	8 漂洗	12:18:58	8 漂洗
11:31:32	9 消毒	11:45:12	9 消毒	12:06:14	9 消毒	12:18:59	9 消毒
11:41:15	9 消毒接触时间 = 00:05:15	11:54:28	9 消毒接触时间 = 00:05:15	12:15:34	9 消毒接触时间 = 00:05:15	12:28:24	9 消毒接触时间 = 00:05:15
11:41:15	9 槽内温度 [°C] = 25.20	11:54:28	9 槽内温度 [°C] = 25.10	12:15:34	9 槽内温度 [°C] = 27.10	12:28:24	9 槽内温度 [°C] = 25.70
11:41:15	10 冲洗	11:54:28	10 冲洗	12:15:34	10 冲洗	12:28:24	10 冲洗
11:43:10	11 漂洗	11:56:23	11 漂洗	12:17:23	11 漂洗	12:30:16	11 漂洗
11:43:11	12 漂洗	11:56:24	12 漂洗	12:17:24	12 漂洗	12:30:17	12 漂洗
11:43:12	13 未洗	11:56:25	13 未洗	12:17:25	13 未洗	12:30:18	13 未洗
11:47:51	13 清洗接触时间 = 00:02:22	12:01:00	13 清洗接触时间 = 00:02:22	12:22:01	13 清洗接触时间 = 00:02:22	12:34:54	13 清洗接触时间 = 00:02:22
11:47:51	13 槽内温度 [°C] = 25.10	12:01:00	13 槽内温度 [°C] = 25.10	12:22:01	13 槽内温度 [°C] = 27.20	12:34:54	13 槽内温度 [°C] = 25.20
11:47:51	14 吹干	12:01:00	14 吹干	12:22:02	14 吹干	12:34:54	14 吹干
11:52:21	14 酒精吹干	12:05:31	14 酒精吹干	12:26:31	14 酒精吹干	12:39:23	14 酒精吹干
11:52:22	循环结束	12:05:32	循环结束	12:26:32	循环结束	12:39:24	循环结束