

Case Report

00009182938 Replica Testing Association 1

Case #: 1001242	Date of Report: Sept. 23, 2024
Patient Name: Ailyah Christopher	Patient DOB: Sept. 1, 2024
Medical Record No:	Gestational Age at Birth: None None
Located in: Akutan Alaska	Current Post Conceptional Age: None weeks None days
Serial Number: 3000043314	Model: SmartMonitor 2 w/ Modem
Case Opened: 03-Sep-2024	No. of Data Uploads: 1 upload(s)
Case Closed: Open	Last Event Upload: 03-Sep-2024

Device Report

Date From: 24-Jan-2024 Date To: 24-Jan-2024

Device Settings					
Monitor Version Information		Alarm Settings		Channel Status	
Monitor Type	Smart Monitor 2	Apnea Alarm Delay	20 secs	QRS	On
Firmware Revision	2.01	Brady Alarm	80 bpm	HR	On
Software Revision	1.1	Brady Onset Delay	5 secs	Impedance	On
Boot Block Revision	1.04	Tachy Alarm	230 bpm	RespRate	Off
System Settings		Tachy Onset Delay	5 secs	Pleth	On
Date Format	Month/Day/Year	Low Breath Rate Alarm	Off	SpO2	On
Rate Display On LCD	On	Record Settings		Pulse Rate	On
Memory Full Alert	50%	Recording Mode	Continuous	Auxiliary1	Off
Avg. Heart Rate	4 Beat Average	Record Pre/Post Time	30/15 secs	Auxiliary2	Off
Memory Full Audible	Off	Apnea Record Delay	Off secs	Auxiliary3	Off
Ext Eqp Trigger	Off	Brady Record	Off	Auxiliary4	Off
Ext. Phy Trigger	Off	Hardware Installed			
Hospital Mode	On	Modem			
Oximeter	Yes	PC Card			

DATE	START TIME	DURATION	EVENT	START TIME	DURATION	EVENT
------	------------	----------	-------	------------	----------	-------