LABELING MANAGEMENT

Function:	Prepared by:	Reviewed by:	Approved by:
Department:	Quality Assurance	Systems and cybersecurity	Quality Assurance
Name:	Ivan Perez	David Pozo	Fernando Pozo
Signature:			
Date:			

Valid until:	05/07 /2028

DOCUMENT HISTORY			
VERSION	CAUSES OF MODIFICATION	EFFECTIVE DATE	REPLACES (CODE, REVISION)
01	Initial edition	22/02/2024	N/A
02	Change in Spika departments, change in label format and location	02/06/2025	PNT-019.01
03	Changing the label image to a general outline	05/07 /2025	PNT-019.02

Content

1. Objective 4
2. Scope 4
3. Responsibilities and Departments affected 4
4. Definitions and Acronyms 5
5. Procedure 5
5.1. Generalities 5
5.2. Recommendations 6
5.3. Tag Management 7
5.4. UDI 7
6. Distribution of Procedure 8
7. Annexes 8
8. Related documentation 8
9. Formation 8

1. Aim

The objective of this procedure is to define the operating procedure for label design at Spika Tech, hereinafter Spika, understanding it as a fundamental task with regard to product identification.

2. Scope

This procedure applies to all labeling generated by Spika for the medical devices it distributes.

3. Responsibilities and Departments Affected

The Technical Manager is responsible for supervising labeling management.

The Research and Development Department is responsible for uploading the labeling to a location visible to the user and keeping it updated.

The affected departments are:

Department	Apply	
	Yeah	No
Address	☒	☐
Environment	☒	☐
Quality Assurance	☒	☐
Systems and Cybersecurity	☒	☐
Research and Development	☒	☐
Marketing and Sales	☒	☐
Management	☒	☐

4. Definitions and Acronyms

- Label : The written, printed or graphic information appearing on the product itself, on the packaging of each unit or on the packaging of several products.
- Intended purpose : the intended use of a product according to the data provided by the manufacturer on the label, in the instructions for use or in the promotional or sales material or statements, and as indicated by the manufacturer in the clinical evaluation.
- Instructions for use : information provided by the manufacturer to inform the user about the intended purpose of a product, its correct use and the precautions to be taken
- Unique Product Identifier or UDI Device Identifier : a series of numeric or alphanumeric characters that is created according to internationally accepted product identification and coding standards and enables the unambiguous identification of products on the market.
- CE marking of conformity" or "CE marking" means a marking by which a manufacturer indicates that a product conforms to the applicable requirements set out in this Regulation and other applicable Union harmonisation legislation providing for its affixing
- ICAD : The ICAD system refers to Automatic Identification and Data Capture (or AIDC) . Identification for Data Capture) and can be a linear barcode or a DataMatrix , etc.

5. Procedure

5.1. Generalities

In the case of medical devices, European Council Regulation 2017/745 must be followed, where the labeling must include:

- Name or trade name of the product
- Information necessary to identify the product, its contents and its intended purpose
- The name, registered trade name or registered trademark of the manufacturer and its registered office

- If the manufacturer's registered office is located outside the Union, the name of the authorized representative and the address of the registered office of the authorized representative.
- The batch number or serial number of the product, preceded by the words "BATCH NUMBER" or "SERIAL NUMBER" or an equivalent symbol, as applicable.
- The support for the unique identification referred to in Article 27, paragraph 4, and Annex VII, Part C.
- Unambiguous indication of the end-date for safe use or implementation of the product, expressed at least in terms of year and month, where applicable.
- Where no indication of the date until which the product can be safely used is given, the date of manufacture is given. This date of manufacture may be included as part of the batch or serial number, provided that the date is clearly identifiable.
- Indication of any specific storage or handling conditions that may apply.

When a lot number, serial number, software identification, or expiration date appears on the label, they will be included in the UDI-PI. When the label also includes the manufacturing date, it need not be included in the UDI-PI. When only the manufacturing date appears on the label, this will be used as the UDI-PI.

It is recommended to use the symbols permitted in each case, following the guidelines of UNE EN ISO 15223-1 Medical Devices. Symbols to be used on labels, labeling, and information to be provided.

5.2. Recommendations

In the labeling, instructions for use, marketing, commissioning and advertising of products, the use of texts, names, trademarks, photographs and images or other signs that may mislead the user or patient as to the intended purpose, safety and operation of the product by any of the following means is prohibited:

a) attributing to the product functions and properties that it does not possess;
b) create a false impression about treatment or diagnosis, functions or properties that the product does not possess;
c) Failure to inform the user or patient about the potential risks involved in using the product for its intended purpose;

d) imply uses of the product other than those indicated as part of the intended purpose for which the conformity assessment was carried out.

5.3. Tag management

Spika will include product information in the form of a label on the home screen of the VR - cardio viewing app, downloading it and making it visible to the user, as shown in the example below. Additionally, the label code and version will be encoded in the lower right margin.

The label will be previously designed and approved in physical /paper/electronic format by at least:

- For software, the designer of the production department and the Technical Manager, thus complying with applicable regulations, which is duly checked by the technical manager of the organization.

5.4. UDI placement criteria

In the case of software, the UDI will be provided on a screen that is easily accessible to the user in an easily readable text format, such as in the “about…” product information file or on the home screen.

6. Distribution of the Procedure

Staff members in the departments listed in Section 3 (Responsibilities and Affected Departments) must be provided with a copy of this procedure. In addition, controlled copy No. 1 will be submitted to the Archives for paper archiving.

Controlled copies to be issued:

Controlled Copy No.	Departments
1	Archive

7. Annexes

NO. / REV.	QUALIFICATION
N/A	N/A

8. Related documentation

Code	QUALIFICATION
N/A	UNE EN ISO – 15223-1 Medical Devices. Symbols to be used on labels, labeling, and information to be provided.
N/A	European Council Regulation 2017/745

9. Training

This version requires that the affected departments indicated in section 3 receive the training indicated below:

Mark with an X	Type of training

☒	**Theoretical** (reading and understanding of the procedure)
☐	**Theoretical – Practical** (If you select this option, contact the Responsible Technician)