APPROVAL, REJECTION AND RELEASE OF PRODUCTS

Function:	Prepared by:	Reviewed by:	Approved by:
Department:	Quality Assurance	Systems and cybersecurity	Quality Assurance
Name:	Iván Pérez	David Pozo	Fernando Pozo
Signature:			
Date:			

Valid until:	02/06/2028

DOCUMENT HISTORY			
VERSION	CAUSES OF CHANGE	DATE OF ENTRY INTO FORCE	REPLACES (CODE, REVISION)
01	Initial edition	22/02/2024	N/A
02	Updating of Spika departments, definition of software reproducibility procedures, downloading and installation process, verification of labelling and correct functioning in the annexes.	15/07/2024	PNT-016.01
03	Amendment of Annex 1 to include the software version and update the applicable regulations. Amendment of Annexes 4 and 5 to include the signature of the responsible technician.	20/08/2024	PNT-016.02
04	Change in Spika's departments	02/06/2025	PNT-016.03

Content

1. Target 4
2. Outreach 4
3. Responsibilities and Departments concerned 4
4. Definitions and Acronyms 5
5. Procedure 5
5.1. Intermediate product 5
5.2. Finished products 6
5.3. Rejected products 6
5.4. Product sampling 7
6. Distribution of the Procedure 7
7. Annexes 7
8. Related documentation 7
9. Training 8
Model release certificate 9
Software reproducibility 11
Download and installation of the software 12
Software labelling 13
Review and approval by the Technical Manager 14

1. Target

The objective of this procedure is to establish guidelines on the systematic approach to ensure compliance of all products with the approved specifications.

The aim of this procedure is to prevent products that do not meet the above-mentioned conditions from being used incorrectly in the production process or from being placed on the market.

2. Outreach

This procedure applies to all medical devices that are released at Spika Tech, hereinafter Spika, regardless of the state of the production process.

3. Responsibilities and Departments concerned

It is the responsibility of the Responsible Technician to ensure that materials and products meet the approved specifications. He/she has the power to approve or reject them as appropriate.

The departments concerned are listed in the table below:

Department	Applies	
	Yes	No
Address	☒	☐
Environment	☒	☐
Quality Assurance	☒	☐
Systems and Cybersecurity	☒	☐
Research and Development	☒	☐
Marketing and Sales	☒	☐

Management	☒	☐

4. Definitions and Acronyms

- Intermediate Product: Partially processed material that still has to pass through other stages of production before becoming a final product.
- Finished product: it is understood to be that which has gone through all the production stages, including its packaging in the final container.
- Specifications: These describe in detail the requirements to be met by the products or materials used or obtained during manufacture. They serve as the basis for quality assessment.

5. Procedure

5.1. Intermediate product

The intermediate products will be checked against their specifications. In the same way, they will be placed in the route https://github.com/SpikaTechSL/VRCardio-VHET360-Visualization (route only accessible to members of the company as it is a private repository) until their release by the Responsible Technician.

If the product corresponds in each and every one of its parameters, it will be declared APPROVED. If, on the other hand, there is any discrepancy between the specifications and the results obtained, a non-conformity report will be opened to determine whether the product meets the specifications or not:

- The discrepancy is a minor defect that does not affect the quality of the product itself, the product may be approved.
- The discrepancy is a major defect that may affect quality, will be ruled as REJECTED/OBSOLETE, will be identified as such and will be located in the path https://github.com/SpikaTechSL/VRCardio-VHET360-Visualization (path only accessible to members of the company as it is a private repository).

In the time after the data acquisition devices are received by the manufacturer and all information is reviewed by Spika personnel, the product remains in quarantine and is identified with an orange label.

The intermediate software product shall be approved according to the model certificate in Annex 1. The VR Cardio product consists of software, which is released by the relevant Spika staff and approved by the Technical Manager, following the certificate in Annex 1.

5.2. Finished products

Once the product has been finalised, the compilation must be checked to see if it has been done according to the product specifications, the variable data corresponds and the intermediate product is duly approved, in which case the product will be approved and released for sale.

If discrepancies exist, a Non-Conformity report should be initiated for:

- To rule if the discrepancy is minor without affecting the quality of the product and its final result, under the authorisation of the Responsible Technician, is that the product is APT for release and supply to the market.
- Ruling if the discrepancy is major and adversely affects safety or the product must be reprocessed before sale, it may be REJECTED or APPROVED after reprocessing. In both cases, under the supervision and approval of the T. Responsible.

In order for the product to be approved and released to the market, it is necessary to:

- Intermediate versions are approved as indicated in Annex 1.
- The final version produced complies with the software reproducibility procedure (Annex 2) that ensures that the released software can be delivered at the point of use without deformation or unauthorised changes and with correct installation.
- The software can be downloaded and installed according to Annex 3.
- The labelling of the software complies with regulatory requirements (Annex 4).
- The technical manager has reviewed all of the above procedures to ensure that the software is working properly and can be released to the market (Annex 5).

5.3. Rejected products

Any product that obtains a rejected opinion must:

- Obsolete/rejected software is located at: https://github.com/SpikaTechSL/VRCardio-VHET360-Visualization/releases (path only accessible to members of the company as it is a private repository).

5.4. Product sampling

SPIKA does not show the product because the device is not yet in series production.

6. Distribution of the Procedure

Staff belonging to the departments listed in Section 3 (Responsibilities and Departments concerned) should be provided with a copy of this procedure. In addition, controlled copy no. 1 will be given to the Archives for archiving on paper.

Controlled copies to be issued:

No. Controlled copy	Departments
1	Archive

7. Annexes

NUMBER / REV.	TITLE
1/01	Certified model
2/01	Software reproducibility
3/01	Download and installation of the software
4/01	Software labelling
5/01	Review and approval by the Technical Manager

8. Related documentation

Code	TITLE

N/A	Regulation (EU) 2017/745 of the European Parliament and of the Council on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 178/2002
N/A	Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC
MQ-01	Quality Manual
PNT-015	Traceability and identification of products and processes
PNT-004	Non-Conformity Management

9. Training

This version requires the following training for the departments concerned as indicated in paragraph 3:

Mark with an X	Type of training
☒	Theoretical (reading and understanding of the procedure)
☐	Theoretical - Practical (If this option is selected, please contact the Technician in charge).

Annex 1_VERSION_04

Model release certificate

SPIKA TECH

RELEASE CERTIFICATE

Product:

Software version released:

Date of manufacture:

Release date:

Determination	Specification / Acceptance Criteria	Result
Review of technical documentation	Complete and updated documentation	
Functional verification of the software	All critical functions tested and verified	
Software validation	Complies with the validation plan and defined criteria.	
Version control and change management	Documented change history	
Risk review (updated risk analysis)	Acceptable risks after mitigation	

I hereby certify that all stages of release of this product have been carried out in compliance with the requirements of Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices.

Any non-conformities, if any, have been fully evaluated and approved by authorised personnel.

It is concluded that the indicated lot is approved for release.

OUTCOME:

☐ CONFORME ☐ NON-COMPLIANCE

S.D.: Responsible Technician

Indicate name and surname

Software reproducibility

Task	Expected Outcome	Compliance (YES/NO)
Compiling the software in a controlled environment	Error-free compiled software	
Integrity testing in different environments	Software running correctly in all environments	
Documentation of changes	Detailed record of all authorised and documented changes	
Installation tests in a controlled environment	Correct installation of the software without documented errors	

Meets all criteria:

Final Result: _____(Pass / Fail)

Signature:

Annex 3_VERSION_04

Download and installation of the software

Task	Expected Outcome	Compliance (YES/NO)
Distribution of software on secure media	Software accessible only by authorised personnel	
Provision of download instructions	Clear and detailed instructions provided	
Provision of installation instructions	Detailed installation manual with screenshots and descriptions	
Post-installation functional tests	Software working correctly after installation	
Technical support information provided	Technical support contact details available	

Meets all criteria:

Final Result: _____(Pass / Fail)

Signature:

Annex 4_VERSION_04

Software labelling

Task	Expected Outcome	Compliance (YES/NO)
Product identification	Clearly labelled name, version and release date	
Manufacturer's information	Clearly labelled name, address and contact details of manufacturer	
System requirements	Minimum and recommended specifications clearly indicated	
Instructions for use	Description of use and links to manuals provided	
Warnings and precautions	Clearly indicated warnings and safety measures	
Legal information	Licence, copyright and terms of use clearly stated	

Meets all criteria:

Final Result: _____(Pass / Fail)

Signature:

S.D.: Responsible Technician

Indicate name and surname

Annex 5_VERSION_04

Review and approval by the Technical Manager

Task	Expected Outcome	Compliance (YES/NO)
Review of documentation	Complete and correctly filed documentation	
Validation of the final product	Final product meeting all specifications and requirements	
Authorisation for release	Declaration of conformity issued and product authorised for release	
Record of approval	Record of approval with signature, date and relevant comments	

Meets all criteria:

Final Result: _____(Pass / Fail)

Signature:

S.D.: Responsible Technician

Indicate name and surname