Annex 2_VERSION_05

CAPA Report

LAYER REPORT	CODE
	AC-001/25

Person opening the action:	Iván Pérez	Department	Quality Assurance
Detection date	15/07/2025	Origin	Notified Body Audit
Description of the finding	The SOP PNT-030 for Strategy for regulatory compliance does not clearly describes the identification of relevant legal requirements, the identification of the sources as well as the responsibilities for the tasks. In addition the follow up of CS and/or harmonized standards is not covered		
Root cause analysis	Procedure PNT-030 does not explicitly and in a structured way address the mechanisms to: - Identify applicable legal requirements (MDR, national legislation, etc.). - Determine the official sources of information (EU Official Journal, European Commission website, notified bodies, industry associations). - Assign clear responsibilities for monitoring these requirements. - Monitor changes in Common Specifications (CS) or relevant harmonised standards. - This omission is due to the fact that the procedure was initially drafted early in the development of the system, and was not revised after the full entry into force of the MDR.		
Immediate correction	Record the non-conformity in the quality system and provide traceability of the follow-up.		
Corrective/preventive action	Revision of procedure PNT-030, including: - Specific section for the identification of applicable legal and regulatory requirements, including MDR, applicable directives, national legislation, and MDCG guidelines.		

	- Inclusion of a list of official sources of updated information (websites, bulletins, standards databases, etc.). - Clear assignment of responsible parties for monitoring, reviewing and updating these requirements. - Description of the monitoring process for harmonised standards and Common Specifications (CS), including frequency, monitoring tools (e.g. CEN notifications, OJEU) and criteria for assessing their applicability. - Integration with the risk assessment and change management procedure to ensure that the system is updated when requirements change.		
Verification of the impact of the action on the product		Yes (if you check this box, explain below)	No
	Is there an impact on regulatory requirements?	☒	☐
	Is there an impact on safety and operation?	☐	☒
	Impact assessment:	The non-conformity has no direct impact on the product and product safety, but it does compromise the ability of the system to adapt to regulatory changes with assurance. Corrective actions will establish a robust regulatory surveillance system to ensure continued compliance with the MDR and other applicable regulations.	
Responsible for the action	Iván Pérez	Department	Quality Assurance
Deadline	31/10/2025		
To be completed by the person who opens	Signature and date	Firmado por PEREZ CIRUELOS IVAN - ***4861** el día 16/07/2025 con un certificado emitido por AC FNMT Usuarios	

To be completed by the quality manager:	Signature and date Date:	Firmado por PEREZ CIRUELOS IVAN - ***4861** el día 16/07/2025 con un certificado emitido por AC FNMT Usuarios

Evaluation of the effectiveness of the change – To be completed when the change is made			
Has the change been implemented successfully?	☐	Yes	
	☐	No	Observations

Effectiveness of change – To be completed when the change is made

Closing carried out by:	Date