CLINICAL EVALUATION

Function:	Prepared by:	Reviewed by:	Approved by:
Department:	Quality assurance	Systems and cybersecurity	Quality assurance
Name:	Ivan Perez	David Pozo	Fernando Pozo
Signature:			
Date:			

Valid until:	05/07 /2028

HISTORY OF THE DOCUMENT			
VERSION	CAUSES OF CHANGE	EFFECTIVE DATE	REPLACES (CODE, REVISION)
01	Initial edition	05/07 /2025	N/A

Content

1. Objective 4
2. Diffusion 4
3. Responsibilities and Departments involved 4
4. Definitions and acronyms 4
5. Procedure 5
5.1. Clinical evaluator 5
5.2. Clinical evaluation 5
5.3. Clinical evaluation plan 6
5.4. Clinical evaluation report 7
6. Record control 7
7. Distribution of Procedure 8
8. Annexes 8
9. Related documentation 8
10. Workout 9

1. Aim

The purpose of this procedure is to provide a system for performing the clinical evaluation of medical devices produced at Spika Tech, hereinafter SPIKA.

2. Beat

The procedure applies to medical devices produced by SPIKA management.

3. Responsibilities and Departments involved

- It is the responsibility of the TD/PRRC and M to select the Clinical Evaluator.
- It is the responsibility of the Clinical Evaluator to perform the clinical evaluation and prepare the Clinical Evaluation Report.
- It is the responsibility of the TD/PRRC and M to approve and sign the Clinical Evaluation Report.
- It is the responsibility of M and the Clinical Evaluator to sign the Declaration of Interests.

4. Definitions and acronyms

- CEP : Clinical Evaluation Plan
- CER : Clinical Evaluation Report
- M : Management
- MD : Medical Device
- TD : Technical Director
- PRRC : Person Responsible for Regulatory Compliance
- M : Management

5. Procedure

5.1. Clinical evaluator

The M will designate one or more individuals to conduct the clinical assessment, known as the Clinical Evaluator. The Clinical Evaluator need not be a SPIKA employee .

The Clinical Evaluator must have a higher education degree in the respective field and five years of documented professional experience; or ten years of documented professional experience if a degree is not a prerequisite for the specific role.

The clinical evaluator, with respect to the particular device under evaluation, should have knowledge of the device technology and its application, the diagnosis and management of the conditions the device is intended to diagnose or manage, and knowledge of medical alternatives, treatment standards, and technology (e.g., specialized clinical experience in the relevant medical specialty).

As a general principle, clinical evaluators should possess knowledge of at least the following:

- Research methodology (including clinical research design and biostatistics).
- Information management (e.g. scientific training or library qualification).
- Experience with relevant databases, such as Embase and Medline.
- Requirements regulatory .
- Medical writing (e.g., postgraduate experience in a relevant science or medicine; training and experience in medical writing, systematic review, and clinical data evaluation).

The clinical evaluator should be familiar with the Medical Device Regulation (MDR) guidelines, MEDDEV 2.7.1./rev 4 and other relevant regulatory requirements.

The Clinical Evaluator must submit the CV to the SPIKA Ministry.

The Clinical Evaluator must complete a Declaration of Interest. This Declaration should include relevant financial interests related to the medical device for which the clinical evaluation is being conducted, in addition to their current work as an evaluator. The Declaration of Interest must be dated and signed by both the evaluator and the SPIKA M.

5.2. Clinical evaluation

The clinical evaluator performs the clinical evaluation in order to obtain and evaluate all available clinical data on the medical device or equivalent devices to determine whether the

device or system is safe and effective, and to demonstrate the device's compliance with the general safety and performance requirements (Annex I of MDR 2017/745).

The clinical evaluator conducts the clinical evaluation based on a thorough analysis of available pre- and post-marketing clinical data relevant to the intended use of the device in question, including clinical performance and safety data. During the clinical evaluation, the clinical evaluator must include in the analysis both the specific medical device in question and any data relating to products declared equivalent by the manufacturer.

The clinical assessor must perform the clinical assessment during the medical device conformity assessment process, prior to its marketing and market launch. To perform the clinical assessment, the Clinical Assessor must prepare the Clinical Assessment Plan (CEP). Once the clinical assessment is completed, the Clinical Assessor must prepare the Clinical Assessment Report (CER). Once the medical device is placed on the market, the clinical assessor must perform the clinical assessment periodically, as new clinical information regarding the safety and performance of the device during use becomes available. The information resulting from the clinical assessment is included in the ongoing risk analysis and may lead to changes in the use of the device.

The clinical evaluator must perform the clinical evaluation at least once a year for implantable medical devices and Class III medical devices, and every two years for other medical devices.

5.3. Clinical evaluation plan

The Clinical Evaluator is responsible for preparing the Clinical Evaluation Plan.

The Clinical Evaluation Plan must include information on the safety or performance of the medical device that is generated from the use of the device, which is obtained from the following:

- Clinical investigation(s) of the device in question.
- Clinical investigation(s) or other studies reported in the scientific literature, of a device for which equivalence with the device in question can be demonstrated.
- Reports published in peer-reviewed scientific literature on other clinical experience with the device in question or a device for which equivalence to the device in question can be demonstrated.
- Clinically relevant information from post-marketing surveillance , in particular from post-marketing clinical follow-up .

5.4. Clinical evaluation report

The Clinical Evaluator is responsible for conducting literature research and preparing the Clinical Evaluation Report and Literature Research Protocol.

The Clinical Evaluation Report must detail:

- The technology on which the medical device is based, the intended use of the device, and any claims made about the clinical performance or safety of the device.
- The nature and extent of the clinical data that have been evaluated.
- How the referenced information (recognized standards and/or clinical data) demonstrates the clinical performance and safety of the device in question.
- That the device(s) operate as intended by the manufacturer.
- That the device(s) do not pose any undue security risk.
- That the risks associated with the use of the device(s) are acceptable when compared to the benefits to the patient.

The Clinical Evaluation Report will be signed and dated by the evaluator(s) and accompanied by the manufacturer's justification for the choice of evaluator.

If any customer property is lost, damaged, or deemed unusable, we will notify you by email . electronics are kept up to date.

6. Record control

ID	Location	Minimum retention time	Responsible	Elimination
Declaration of interests	B2.2. Intended use, MDR SharePoint	10 years after the last device came onto the market	TD/PRRC	TD/PRRC
Clinical evaluation plan	K1 Clinical Assessment Plan, MDR SharePoint			

Clinical evaluation report	K2 Clinical Assessment Report, MDR SharePoint			
Bibliographic research protocol	K4. Bibliographic search, MDR SharePoint			

7. Distribution of the Procedure

Staff in the departments listed in Section 3 (Responsibilities and Affected Departments) shall be provided with a copy of this procedure. Controlled Copy 1 shall also be submitted to the Archives for paper filing.

Controlled copies to be issued:

No. Controlled copy	Departments
1	Archive

8. Appurtenances

NUMBER / REV.	EXTERNAL (BUT)	QUALIFICATION
N/A	N/A	N/A

9. Related documentation

Code	QUALIFICATION
N/A	N/A

10. Training

This version requires the following training for the departments involved as indicated in section 3:

Mark with an X	Type of training
☒	Theoretical (reading and understanding of the procedure)
☐	Theoretical - Practical