PREPARATION OF TECHNICAL DOCUMENTATION: MEDICAL DEVICES

Function:	Prepared by:	Reviewed by:	Approved by:
Department:	Quality Assurance	Systems and cybersecurity	Quality Assurance
Name:	Ivan Perez	David Pozo	Fernando Pozo
Signature:			
Date:			

Valid until:	02/06/2025

DOCUMENT HISTORY			
VERSION	CAUSES OF MODIFICATION	EFFECTIVE DATE	REPLACES (CODE, REVISION)
01	Initial edition	22/02/2024	N/A
02	Change in Spika's departments	02/06/2025	PNT-024.01

Content

1. Objective 4
2. Scope 4
3. Responsibilities and Departments affected 4
4. Definitions and Acronyms 5
5. Procedure 5
5.1. Description and generalities 5
5.2. Contents of the technical file 5
5.3. Review and approval of technical file 7
6. Distribution of Procedure 7
7. Annexes 7
8. Related documentation 7
9. Formation 7
Technical File Medical Device 9

1. Aim

This procedure describes the process followed by Spika Tech, hereinafter SPIKA, for the preparation of technical documentation for medical devices, ensuring compliance with applicable regulations in force.

2. Scope

This procedure applies to the documentation generated (technical file) on a medical device generated at SPIKA facilities.

3. Responsibilities and Departments Affected

The Technical Manager will be responsible for reviewing and approving the technical file, ensuring that all manufacturing and quality control of the medical device has been carried out in compliance with all applicable requirements (regulatory, internal, and customer requirements). The technical file will be prepared by competent personnel with knowledge of the product, its characteristics, and the production and quality control process.

The affected departments are those indicated in the following table:

Department	Apply	
	Yeah	No
Address	☒	☐
Environment	☒	☐
Quality Assurance	☒	☐
Systems and Cybersecurity	☒	☐
Research and Development	☒	☐
Marketing and Sales	☒	☐
Management	☒	☐

4. Definitions and Acronyms

- Technical file : Set of technical documentation that allows the assessment of conformity with the applicable regulatory requirements according to current regulations.

5. Procedure

5.1. Description and generalities

For each product or family of Medical Devices designed and manufactured by SPIKA, a Technical File must be generated, which must be periodically reviewed and maintained with post-marketing monitoring data, so that compliance with the applicable requirements indicated in Regulation 746/2017 and/or Regulation 745/2017, and section 4.2.3 of the ISO 13485:2016 standard is evidenced.

The Technical File will be recorded in the template included in Annex 1 of this procedure.

5.2. Contents of the technical file

The following information will be included in the technical file :

CONTENTS OF THE TECHNICAL FILE
Product Description and Specifications

• Possible variants of the same will be included, if applicable.
• Denomination, or trade name
• General description, which may include operating principles and instructions for use.
• Intended use or intended purpose, which may include information about:
  ▪ what it detects or measures
  ▪ its monitoring, control or diagnostic function
  ▪ whether it is quantitative or qualitative
  ▪ user (if it is for professional use, *point of care*, other)
• *in vitro* diagnostic medical device .

CONTENTS OF THE TECHNICAL FILE
• Product risk class, in accordance with Annex VIII of the Regulation • Indication of the materials, ingredients or reagents that compose it • Indication of other accessories or materials that, although not considered components of the product, must be used with it to achieve its intended use/intended purpose.
Reference to previous and similar generations of the product (if they exist)
• Information about the previous generation of the product • Compilation of bibliographic references, or information from manufacturers of similar products on the market
Information provided by the manufacturer
• Instructions for use • Labeled • Product safety data sheets • Specs
Design and Development
• Description of the development and design process and the controls applied
Product stability
• Expiration or retest dates • Permitted transportation excursions (if permitted) • Storage conditions • Transport conditions (if special)
Other information (if applicable)
• Technology transfer process • Risk Management • Validations (methods/processes)

Table I. Contents of the Technical File of medical device.

5.3. Review and approval of technical file

The Technical File will be reviewed and updated when sufficient data from the product's monitoring or post-marketing phase is available, or when relevant data is available, to ensure its suitability for the product's use.

6. Distribution of the Procedure

Staff members in the departments listed in Section 3 (Responsibilities and Affected Departments) must be provided with a copy of this procedure. In addition, controlled copy No. 1 will be submitted to the Archives for paper archiving.

Controlled copies to be issued:

Controlled Copy No.	Departments
1	Archive

7. Annexes

NO. / REV.	QUALIFICATION
1/01	Technical file template

8. Related documentation

Code	QUALIFICATION
N/A	EN ISO 13485 (4.2.3 Medical device archiving)
N/A	Regulation (EU) 2017/745 on medical devices

9. Training

This version requires that the affected departments indicated in section 3 receive the training indicated below:

Mark with an X	Type of training
☒	**Theoretical** (reading and understanding of the procedure)
☐	**Theoretical – Practical** (If you select this option, contact the Responsible Technician)

Annex 1_VERSION_02

Technical File for Medical Devices

Index

1. Product Description and Specifications 10
2. References to previous and similar generations of the product 11
3. Information provided by the manufacturer 11
3.1. Instructions for Use 11
3.2. Labeling 11
3.3. Product safety data sheets 11
3.4. Specifications 11
4. Design and Development 12
5. Product stability 12
6. Additional information 13

1. Product Description and Specifications

Product identification (name, trade name, code or reference):	
Variants (if applicable):	
General description (operating principles, instructions for use):	
Intended use or intended purpose:	
Monitoring or diagnostic function:	
Indicate whether qualitative or quantitative	
User (professional , point of care):	
Justification for its classification as a medical device:	
Risk class (Annex VIII Regulation):	
Other accessories or materials required:	

2. References to previous and similar generations of the product

Information about the previous generation of the product:	
Bibliographic references of similar products marketed:	
Information from manufacturers of similar products on the market:	

3. Information provided by the manufacturer

3.1. Instructions for Use

3.2. Labeled

3.3. Product safety data sheets

3.4. Specs

4. Design and Development

5. Product stability

Expiration date or retest :	
Permitted excursions for transportation:	
Storage conditions:	
Conditions of transport:	

6. Additional information

Technology transfer process:	
Risk management:	
Method validation:	
Process validation:	