


3DQLayers: Volumetric Layer Based Analysis for Quantitative Renal MRI

Alexander J Daniel¹ and Susan T Francis^{1,2}

¹ Sir Peter Mansfield Imaging Centre, University of Nottingham, Nottingham, United Kingdom ² NIHR Nottingham Biomedical Research Centre, Nottingham University Hospitals NHS Trust and the University of Nottingham, Nottingham, United Kingdom  Corresponding author

DOI: [10.xxxxxx/draft](https://doi.org/10.xxxxxx/draft)

Software

- [Review](#) 
- [Repository](#) 
- [Archive](#) 

Editor: [Open Journals](#) 

Reviewers:

- [@openjournals](#)

Submitted: 01 January 1970

Published: unpublished

License

Authors of papers retain copyright and release the work under a Creative Commons Attribution 4.0 International License ([CC BY 4.0](#)).

Summary

Informative measurements of the kidneys structure and function can be performed using quantitative Magnetic Resonance Imaging (MRI) where each voxel of the image is a measurement of the physical properties of the tissue being image. Traditionally, analysis of these images is performed by segmenting the kidney into its constituent tissue types and calculating the average of each measurement for each tissue type. The process of segmenting renal tissue types is time consuming and inaccurate. An alternative to tissue segmentation proposed by Pruijm *et al* involves dividing the kidney into layers based on each voxels distance between the outer and inner surface of the kidney, a method known as Twelve Layer Concentric Objects (TLCO) (Li *et al.*, 2020; Milani *et al.*, 2017; Piskunowicz *et al.*, 2015). Layer based analysis only requires segmenting the whole kidney rather than the tissues within and is therefore quicker and more repeatable. The TLCO method does however have some limitation, it can only be performed on a single slice image, requires the image to be acquired at a specific angle to the kidneys, requires manual delineation of the outside and inside surface of the kidney, divides the kidney into the same number of layers irrespective of the size of the kidney and the software itself is closed source. These limitations are addressed by 3DQLayers, a Python package to automatically define 3D, multi-slice, renal layers of known thickness.

Statement of need

Background

The kidneys are a pair of structurally and functionally complex organs in the lower abdomen that participate in the control of bodily fluids by regulating the balance of electrolytes, excreting waste products of metabolism and excess water from blood to urine (Lote, 2012). The each kidney is separated into two tissue types; cortical tissue located towards the outside of each organ, and medullary tissue arranged in small pyramids towards the centre of the organ (Hall, 2015), Figure 1. MRI is the ideal medical imaging modality to study the kidneys due to its non-ionising, non-invasive and quantitative nature (Francis *et al.*, 2023; Selby & Francis, 2024). Quantitative MRI is the process of taking measurements where the value of each voxel has numerical significance, in physical units, based on the tissues underlying properties rather than simply representing signal intensity in arbitrary units. Example quantitative measurements include how readily water can diffuse through the tissue and the rate at which blood perfuses into the tissue. To help interpret quantitative images, regions of interest (ROI) are defined and statistical measures taken of the voxels within each region.

Segmenting ROI for the renal cortex and medulla manually is time consuming, difficult and prone to intra- and inter-reader variation thus decreasing the repeatability of measurements. Pruijm *et al* proposed an alternative to tissue ROI based analysis in the Twelve Layer Concentric

Object (TLCO) method (Li et al., 2020; Milani et al., 2017; Piskunowicz et al., 2015) where users delineate the inner and outer boundaries of the kidney to generate twelve equidistant layers between the renal pelvis and the surface of the kidney. The outer layers are analogous to the cortex and inner layers are analogous the medulla.

TLCO requires the MR image to be a single slice cutting through the kidneys on their longest axis (coronal-oblique) however, this is not always desirable (Bane et al., 2020). Often researchers prefer to acquire multi-slice images to increase the number of voxels in the image and gain a better understanding of the heterogeneity of the kidney. Additionally flexibility in the orientation images are acquired at is highly desirable. These limitations of TLCO were the motivation for the development of 3DQLayers, a volumetric, quantitative-depth based analysis method for renal MRI data.

Methods

Example Usage

Figures




Figure 1: a. A schematic of the kidneys showing the renal cortex and medullary pyramids. b. An anatomical MR Image of the abdomen showing the kidneys with the renal cortex appearing as a light band towards the outside of the kidneys and medullary pyramids as darker patches towards the centre of the kidneys.

Acknowledgements

References

- Bane, O., Mendichovszky, I. A., Milani, B., Dekkers, I. A., Deux, J.-F., Eckerbom, P., Grenier, N., Hall, M. E., Inoue, T., Laustsen, C., Lerman, L. O., Liu, C., Morrell, G., Pedersen, M., Pruijm, M., Sadowski, E. A., Seeliger, E., Sharma, K., Thoeny, H., ... Prasad, P. V. (2020). Consensus-based technical recommendations for clinical translation of renal BOLD MRI. *Magnetic Resonance Materials in Physics, Biology and Medicine*, 33(1), 199–215. <https://doi.org/10.1007/s10334-019-00802-x>
- Francis, S. T., Selby, N. M., & Taal, M. W. (2023). Magnetic Resonance Imaging to Evaluate Kidney Structure, Function, and Pathology: Moving Towards Clinical Application. *American Journal of Kidney Diseases*. <https://doi.org/10.1053/j.ajkd.2023.02.007>

- 67 Hall, J. E. (2015). *Guyton and Hall Textbook of Medical Physiology*. Elsevier Health Sciences.
68 ISBN: 978-1-4557-7005-2
- 69 Li, L.-P., Milani, B., Pruijm, M., Kohn, O., Sprague, S., Hack, B., & Prasad, P. (2020). Renal
70 BOLD MRI in patients with chronic kidney disease: Comparison of the semi-automated
71 twelve layer concentric objects (TLCO) and manual ROI methods. *Magnetic Resonance*
72 *Materials in Physics, Biology and Medicine*, 33(1), 113–120. [https://doi.org/10.1007/](https://doi.org/10.1007/s10334-019-00808-5)
73 [s10334-019-00808-5](https://doi.org/10.1007/s10334-019-00808-5)
- 74 Lote, C. J. (2012). *Principles of Renal Physiology*. Springer Science & Business Media.
75 ISBN: 978-94-011-4086-7
- 76 Milani, B., Ansaloni, A., Sousa-Guimaraes, S., Vakilzadeh, N., Piskunowicz, M., Vogt, B.,
77 Stuber, M., Burnier, M., & Pruijm, M. (2017). Reduction of cortical oxygenation in
78 chronic kidney disease: Evidence obtained with a new analysis method of blood oxygenation
79 level-dependent magnetic resonance imaging. *Nephrology Dialysis Transplantation*, 32(12),
80 2097–2105. <https://doi.org/10.1093/ndt/gfw362>
- 81 Piskunowicz, M., Hofmann, L., Zuercher, E., Bassi, I., Milani, B., Stuber, M., Narkiewicz, K.,
82 Vogt, B., Burnier, M., & Pruijm, M. (2015). A new technique with high reproducibility
83 to estimate renal oxygenation using BOLD-MRI in chronic kidney disease. *Magnetic*
84 *Resonance Imaging*, 33(3), 253–261. <https://doi.org/10.1016/j.mri.2014.12.002>
- 85 Selby, N. M., & Francis, S. T. (2024). Assessment of Acute Kidney Injury using MRI. *Journal*
86 *of Magnetic Resonance Imaging*, n/a(n/a). <https://doi.org/10.1002/jmri.29281>

DRAFT