Alexia Crawford

Mathematics of Medical Imaging

March 9th, 2026

Advancements in Computed Tomography: A Comparative Analysis of Analytical, Iterative, and AI-Driven Reconstruction

1.0 Introduction

Since the start of Computed Tomography (CT) in the 1970s, doctors have faced a difficult tradeoff between needing clear images to make accurate diagnoses and using the lowest radiation possible to keep patients safe (Shulz et al.). One answer to solving this issue is a reconstruction algorithm. This algorithm is the math used to turn X-ray data into clear pictures that can be viewed on a screen. For a long time, the Filtered Back Projection (FBP) was most commonly used. It was fast and reliable. When the radiation dose is lowered, the FBP images become less clear. Because of this, the medical field has advanced its algorithms. Two of these more advanced algorithms are the Iterative Reconstruction (IR) and Deep Learning Reconstruction (DLR). IR uses a feedback loop to clear up images and DLR uses artificial intelligence to create high-quality images. While Filtered Back-Projection provided the mathematical foundation for CT imaging, the shift toward Iterative and Deep Learning reconstructions represents a move from "perfect-geometry" assumptions to "data-driven" statistical modeling.

2.0 Filtered Back Projection

The Filtered Back-Projection has been the "classic analytical algorithm for computed tomography reconstruction" because of its "high computational efficiency," (Tan et al.). This algorithm relies on two main steps. The first step is the filtering. Raw projection data is multiplied by the ramp filter, a high pass filter used to prepare the projection data for reconstruction. This prepares the frequency domain for image creation. Window functions can be added to this filter to limit noise and "enforce the values … outside a given frequency range to be zero," (Tan et al.). The second step is the back projection operation. In this step, the original image is reconstructed. This occurs by putting the processed data back across the image grid. Because CT sensors are discrete, this step needs a linear interpolation to estimate projections values. This can be done at un-sampled locations.

2.1 Challenges with Filtered Back Projection

While FBP is the standard tool for non-invasive diagnosis, like surgery, it still doesn't reach the requirement to "decrease radiation dose" (Tan et al.) by reducing tube current and exposure time. When doctors lower the radiation dosage, the FBP produces images that are grainy and noisy. This makes it hard for the doctors to see what is wrong. A major part of this problem is how the math handles the gaps in data. The standard linear interpolation used to fill these gaps in is "vulnerable to noise" which can cause information loss. This means important details may disappear when the final image is created.

3.0 iterative Reconstruction

Iterative reconstruction differs from the standard methos because it doesn't rely on one mathematical transform. These "algorithms can be considered as a feedback mechanism, with a simulator of the CT-physics (re-projection) in the feedback loop". The process starts with an initial estimate and "updates the reconstruction image, based on deviations between the measured and simulated scans," (Nuyts et al.). This cycle continues until the simulated projections match the data collected by the scanner. One of the main reasons IR is becoming more popular, and interest is increasing is that it has potential to significantly reduce the dosage. IR can "significantly reduce patient dose" and still maintain the quality of the diagnostic. IR also allows researchers to fix different problems that ruin image quality. This can't be done with many other formulas. IR does this by using computer models that mimic how X-ray particles travel and how sensors detect them. Because these algorithms aren't limited to a single mathematical pattern, they are able to reconstruct images from "arbitrary X-ray system geometries," (Nuyts et al.). This makes them more versatile than other standard methods such as the filtered back projection.

3.1 Limitations with Iterative Reconstruction

Although it has many benefits, IR needs a lot of computer power to work well, along with specific digital models. Getting accurate model-based correction with IR requires a lot of effort and expensive, high-powered hardware. The computer has to break the patient's body down to a "finite number of basis functions," the math leads to differences between the measured data and

the simulated data (Nuyts et al.). While Iterative Reconstruction is powerful, it still has issues and doesn't fix everything. For example, "the conventional FBP reconstruction (left) shows a ghosting artifact near the heart wall and blurred vessels... these artifacts are significantly reduced by KPIR," a form of iterative reconstruction.

4.0 Deep Learning Reconstruction

Deep Learning reconstruction (DLR) has become more popular over the past few years as a method to "further improve image quality and possibly reduce radiation dose" (Koetzier et al.). DLR uses artificial intelligence to "reconstruct high-quality images" by using convolutional neural networks (CNNs). To teach the computers how to do this, researchers use supervised learning where they show the AI low dose scans, which are grainy, and then show it a high dose scan, which are clear, of the same image so the computer can learn what a clear image should look like.

4.1 DLR Frameworks and Approaches

There are three DLR techniques that are categorized based on how they interact with traditional reconstruction steps. The first one is Indirect DLR. These still use FBP or IR at some point and include sinogram-based models that optimize data before reconstruction, image-based models that optimize the image after reconstruction, and hybrid models that combine both. The second technique is Direct DLR. These algorithms "reconstruct the sinogram directly into an image, without the use of FBP or IR" (Koetzir et al.). This method can help avoid mistakes that can

show up in other reconstruction methods as long as it is trained by good data. The third technique is Denoising. Some software operates as “deep learning-based denoising,” which works on images that have already been reconstructed. It then improves “heterogeneous liver parenchymal texture evaluation" and reduces the noise.

DLR has many advantages compared to the filtered back projection and iterative reconstruction. One advantage is that the image quality compared to FBP is improved. It also lowers “image noise and artifacts while preserving spatial resolution," (Koetzir et al.). Because of the improved noise reduction, DLR allows for dose reduction between 30 and 71% compared to other options (Koetzir et al.). Another advantage is that DLR can reconstruct images faster than some other techniques. One of the most important benefits is that DLR doesn’t have the “plastic” or “blotchy” texture that is seen in other methods. It has a “desired noise texture” that is similar to high dose FBP scans.

4.2 Deep Learning Reconstruction Advantages and Disadvantages

While DLR has many improvements over the FBP and IR, it also faces some challenges. For example, in low does settings, the AI can confuse background noise for something else and replace it with a “nonexistent object.” This is referred to as a hallucination (Koetzir et al.). On the other hand, if the AI wasn’t trained on examples of small blood vessels or rare growths, it might remove them or blue them when it shouldn’t. Koetzir et al. write the “performance of DLR algorithms relies on the quality of data used for model training."

5.0 Comparative Analysis

These three methods, Filtered back projection, iterative Reconstruction, and Deep Learning Reconstruction, are compared based on the noise quality, resolution, and efficiency.

5.1 Noise and Image Quality

While all three methods aim to reduce the image noise in low dose scans, they all use different techniques to do so and end up with different results. When studies were performed on the same patients using FBP and IR, IR was shown to improve the image quality and decrease the noise in the images compared to the traditional FBP (Fareed et al.). It was also shown by Son et al. that while hybrid IR can lower noise magnitude, DLR often provides the desired noise texture and keeps high quality results. Koetzir et al. explain that IR methods reduce the noise magnitude, they produce low quality results that are grainy that radiologists describe as “objectionable.”

5.2 Spatial Resolution

The ability to see important and small details, spatial resolution, is an important point of comparison. In the study by Fareed et al., they found that the “eduction in image noise does not affect spatial resolution. Newer studies show that “DLR algorithm may provide improved noise characteristics and better spatial resolution than the hybrid IR algorithm,” (Son et al.). Specifically, the DLR has been measured with a shorter edge rise distance which suggests sharper edges in the image.

5.3 Efficiency and Clinical Use

When choosing which method to use, the choice often comes down to how much time the math takes to run. According to Koetzir et al., FBP is still the standard reconstruction method for CT images because it is simple, fast, and reliable. While IR produces better images than the FBP, these algorithms are time consuming because they have a lot of iteration to do (Tan et al.). Lastly, the DLR uses AI to "reconstruct high-quality images from lower-dose CT faster than MBIR (Model-Based Iterative Reconstruction)," (Koetzier et al.). The choice between which method to use is a balance between the efficiency of FBP, the noise reduction of IR, and the ability of DLR to quickly reconstruct images with good resolution.

6.0 Conclusion

The evolution of CT image reconstruction represents a shift from using simple math formulas to more advanced data-driven AI models. For over 40 years, the filtered back projection was the standard way to reconstruct CT images. This is because it was fast and reliable. As medical professionals tried to decrease the radiation dosage to patients, the FBP started to fault. The images became grainy and noisy and made it hard to see what was wrong on these images. Iterative Reconstruction was able to help with this issue by using feedback loops to clean up the images. Even though IR was helpful, it was more time consuming and make the pictures look plasticky or blotchy.

Deep learning reconstruction helps solve these problems by using artificial intelligence to quickly reconstruct images with low noise and attention to detail. By replacing the math with AI steps, the models significantly reduce radiation dosage and still produce high quality images for doctors. Although the AI produces better results than older methods, it can still make mistakes. For example, blurring out important details.

This technology will continue to advance as it connects with new and improved hardware and software. Advanced scanners will provide the higher quality data that AI needs to get even better at identifying and removing artifacts. The computing power of AI is great at handling spectral data. Spectral data contains more hidden information about the body. As these technologies continue to work together, medical imaging moves toward a future where non-invasive and low radiation dosage diagnoses are faster and more accurate than ever before.

Works Cited

Fareed, Ahmed, et al. “Impact of Iterative Reconstruction vs. Filtered Back Projection on Image Quality in 320-Slice CT Coronary Angiography.” *Medicine*, vol. 96, no. 48, Dec. 2017, p. e8452, https://doi.org/10.1097/md.0000000000008452.

Johan Nuyts, et al. “Modelling the Physics in the Iterative Reconstruction for Transmission Computed Tomography.” *Physics in Medicine and Biology*, vol. 58, no. 12, 21 June 2013, pp. R63–R96, https://doi.org/10.1088/0031-9155/58/12/r63. Accessed 27 Apr. 2023.

Koetzier, Lennart R, et al. “Deep Learning Image Reconstruction for CT: Technical Principles and Clinical Prospects.” *Deep Learning Image Reconstruction for CT: Technical Principles and Clinical Prospects*, vol. 306, no. 3, 31 Jan. 2023, https://doi.org/10.1148/radiol.221257.

Schulz, Raymond A., et al. “How CT Happened: The Early Development of Medical Computed Tomography.” *Journal of Medical Imaging*, vol. 8, no. 05, 29 Oct. 2021, www.ncbi.nlm.nih.gov/pmc/articles/PMC8555965/, https://doi.org/10.1117/1.jmi.8.5.052110.

Son, Wookon, et al. “Comparison of a Deep Learning-Based Reconstruction Algorithm with Filtered Back Projection and Iterative Reconstruction Algorithms for Pediatric Abdominopelvic CT.” *Korean Journal of Radiology/Korean Journal of Radiology*, vol. 23, no. 7, 1 Jan. 2022, pp. 752–752, https://doi.org/10.3348/kjr.2021.0466. Accessed 3 May 2024.

Tan, Xi, et al. “Deep Filtered Back Projection for CT Reconstruction.” *IEEE Access*, 1 Jan. 2024, pp. 1–1, https://doi.org/10.1109/access.2024.3357355.