

# EEG and MEG Data Preprocessing

## Oral presentation

Alexi CANESSE and Esteban CHRISTIANN,  
ENS Paris-Saclay, France

Project for the time series course  
Part of the MVA program at ENS Paris-Saclay.

10<sup>th</sup> January 2024

# Table of content

1 Introduction

2 Methods

3 Application : Reconstruction

# Introduction

# Contributions

# Table of content

1 Introduction

2 Methods

3 Application : Reconstruction

# Robust detrending

Usual detrending methods are sensible to artefacts.

⇒ Robust detrending: estimate the trend ignoring outliers, and flag high errors as outliers for the next iteration




Figure: Application of the detrending algorithm.

# Inpainting

Inpainting is the process of reconstructing the signal at timesteps affected by glitches. Here, we assume that glitches locations are known.

This algorithm first estimates the linear relationship between channels and use it to reconstruct the signal.




Figure: Application of the inpainting algorithm.

# Outlier detection

To flag outliers, we use the previous algorithm and flag poorly reconstructed values as outliers for the next iteration.




Figure: Application of the outlier detection algorithm.



# Robust rereferencing




Figure: Application of the robust rereferencing algorithm.

# Step removal




Figure: Application of the step removal algorithm.

# Ringling removal

Ringling artefacts are caused by the antialiasing filter. First, we estimate the parameters of this filter to model its impulse response and cancel it, which removes ringling artefacts.




Figure: Left: A full signal. Right: zoom on an artefact

# Table of content

1 Introduction

2 Methods

3 Application : Reconstruction

# Data

# A complete pipeline for preprocessing

Remove steps → Remove ringing artefacts → Robust detrend → Detect outliers → Inpainting → Robust rereference.





Figure: Left: Raw MEG data. Right: Cleaned MEG data

# Results



**Figure:** Top: Raw MEG and its reconstruction. Bottom: Cleaned MEG with its reconstruction