Nesher – Nurse call system

Geriatric Medical Center · floors B02 – L13 · ACE BG / Advantecnia

Floor	Bed / chair points	Numbers (green)	Toilet points	Numbers (blue)	Rooms (est.)
B02	0	–	0	–	0
L00	0	–	2	1–2	0
L01	15	1–15	0	–	7
L02	30	16–45	11	3–13	13
L03	32	46–77	11	14–24	14
L04	32	78–109	11	25–35	14
L05	50	110–159	10	36–45	22
L06	50	160–209	10	46–55	22
L07	50	210–259	30	56–85	22
L08	48	260–307	37	86–122	20
L09	46	308–353	19	123–141	20
L10	46	354–399	19	142–160	20
L11	46	400–445	19	161–179	20
L12	45	446–490	17	180–196	20
L13	45	491–535	17	197–213	20
TOTAL	**535**	**1 – 535**	**213**	**1 – 213**	**234**

Points read from base drawings GCN-ELC-B02…L13-03 (sockets plans, rev 0, 29.07.24). Rooms: from room labels on L03, L04, L07, L08; other floors estimated – to be confirmed.

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 01

Legend

Green number
Bed / dialysis chair call point – ACE82012153 + ACE39060521

Blue number
Toilet / shower SOS pull-cord – ACE40320303

Sheets

- 01 Summary · 02 System architecture · 03 Typical room & toilet
- 04 – 17 Floors L00 – L13 (B02: no points)

Quantities (all floors)



535 × ACE82012153
Red bed call panel + bayonet socket



213 × ACE40320303
SOS toilet / shower pull-cord




234 × ACE82082137
Presence / cancel (green) · 1 per room

234 × EOP004
Corridor light LED R·G·B · 1 per room

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 02

System architecture · riser




Notes

- One SmartHub per floor (capacity per floor to be confirmed by ACE BG).
- Nurse iPad per nurse station on Wi-Fi (network by others).
- Year 1 software maintenance included; years 2–3 optional line.

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 03

Typical double room & toilet



Equipment per typical room



2 × ACE82012153 + 2 × ACE39060521
1 per bed · handset plugs into bayonet



1 × ACE40320303
toilet / shower pull-cord



1 × ACE82082137
presence / cancel next to the door




1 × EOP004
corridor light above the door



1 × ECV169
room controller, corridor ceiling

Mounting / wiring notes

- Every device cabled radially to the room PCB with Cat6.
- PCBs in daisy chain on the bus (RS485 + 12-48 VDC).
- Nurse call wiring in its own tray / conduit (spec 2.3.п).
- Heights: bed panel 1.15 · presence 1.35 · SOS 1.85 · light 2.42 m.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 04

Floor L00

0 bed / chair points (–)

2 toilet points (1–2)

Equipment on this sheet



2 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



1 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L00-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 05

Floor L01

15 bed / chair points (1–15)

0 toilet points (–)

Equipment on this sheet



15 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



7 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



7 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



7 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L01-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 06

Floor L02

30 bed / chair points (16–45)

11 toilet points (3–13)

Equipment on this sheet



30 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



11 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



13 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



13 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



19 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling




1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L02-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 07

Floor L03

32 bed / chair points (46–77)

11 toilet points (14–24)

Equipment on this sheet



32 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



11 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



14 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



14 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



20 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L03-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 08

Floor L04

32 bed / chair points (78–109)

11 toilet points (25–35)

Equipment on this sheet



32 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



11 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



14 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



14 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



20 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling




1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L04-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 09

Floor L05

50 bed / chair points (110–159)

10 toilet points (36–45)

Equipment on this sheet



50 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



10 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



22 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



22 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



27 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling




1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L05-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 10

Floor L06

50 bed / chair points (160–209)

10 toilet points (46–55)

Equipment on this sheet



50 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



10 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



22 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



22 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



27 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling




1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L06-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 11

Floor L07

50 bed / chair points (210–259)

30 toilet points (56–85)

Equipment on this sheet



50 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



30 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



22 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



22 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



37 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling




1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L07-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 12

Floor L08

48 bed / chair points (260–307)

37 toilet points (86–122)

Equipment on this sheet



48 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



37 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



20 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



20 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



39 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L08-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 13



Floor L09

46 bed / chair points (308–353)

19 toilet points (123–141)

Equipment on this sheet



46 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



19 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



20 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



20 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



30 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L09-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 14



Floor L10

46 bed / chair points (354–399)

19 toilet points (142–160)

Equipment on this sheet



46 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



19 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



20 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



20 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



30 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L10-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 15

Floor L11

46 bed / chair points (400–445)

19 toilet points (161–179)

Equipment on this sheet



46 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



19 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



20 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



20 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



30 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L11-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 16

Floor L12

45 bed / chair points (446–490)

17 toilet points (180–196)

Equipment on this sheet



45 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



17 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



20 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



20 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



29 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L12-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.



ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. November 2026
	Reference.	Nesher_esq_REV00 · Sheet 17

Floor L13

45 bed / chair points (491–535)

17 toilet points (197–213)

Equipment on this sheet



45 × ACE82012153
Red bed call panel + bayonet socket · 1 per bed



17 × ACE40320303
SOS toilet / shower pull-cord · 1 per toilet



20 × ACE82082137
Presence / cancel (green) · 1 per room (est.)



20 × EOP004
Corridor light LED R·G·B · 1 per room (est.)



29 × ECV169
24-pin PCB 12-48 V room controller · corridor ceiling



1 × EVV049
SmartHub connectia · floor rack



1 × EOP002
Nurse iPad 10.9" Wi-Fi · nurse station

Notes

- Positions of bed and toilet points as on base drawing GCN-ELC-L13-03.
- Presence and corridor light: next to / above each room door (see sheet 03).
- Green / blue numbers match the review list and the BOQ.