
ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. October 2026
	Reference.	9722600017_esq_REV03



Equipment (quantities per proposal 9722600017)



11 × ECV169
PCB 24 pins · #1-#8 rooms (A/B), #9-#10 WC B3, #11 WC B2



12 × ACE82012153
Bed call panel + handset (1 per bed)



6 × ACE82082137
Presence / cancel · 4 rooms + 2 WCs



7 × EOP004
Corridor light · 4 rooms + WC B3 ×2 + WC B2



3 × ACE40320303
SOS pull-cord · WC B2 ×1 · WC B3 ×2



1 × EVV049
SmartHub (Raspberry Pi) · rack



1 × EOP006
LVNS server · rack



1 × EOP002
Nurse iPad · wifi

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. October 2026
	Reference.	9722600017_esq_REV03

Equipment on this sheet



1 × EVV049
SmartHub connectia



1 × EOP006
LVNS server



1 × EVV038
Licences + MDM



2 x 2 x 0,8 mm

»»» To Room 4 · PCB #1

SmartHub connectia · EVV049 (Raspberry Pi)



LVNS server · EOP006

PoE switch · UPS · rack

(not included)

Rj45 - LAN PoE - UTP Cat6a - RJ45


»» Nurse iPad (wifi)

Bus daisy chain · 11 × PCB ECV169 (corridor ceiling)

Room 4	Room 3	Room 2	Room 1	WC B3	WC B2
#1-#2	#3-#4	#5-#6	#7-#8	#9-#10	#11

2 x 2 x 0,8 mm · RS485 + 12-48 VDC

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. October 2026
	Reference.	9722600017_esq_REV03



Equipment on this sheet



3 × ACE82012153
Bed call panel + bayonet · LL · h 1,15 m



3 × ACE39060521
Pear handset · 2 m cable to the bayonet socket



1 × ACE82082137
Presence / cancel · next to the door



1 × EOP004
3-colour LED · above the door




2 × ECV169
PCB 24 pins · corridor ceiling

Mounting / wiring notes

- Radial wiring: each device has its own UTP Cat6
- cable to its PCB in the corridor ceiling.
- PCB A: bed calls 1-3 + presence + corridor light.
- PCB B: handset disconnection beds 1-3.
- Bus 2x2x0,8 mm daisy chain A → B → next room.
- Light switches 1,05 m (by others).

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. October 2026
	Reference.	9722600017_esq_REV03



Equipment on this sheet



3 × ACE82012153
Bed call panel + bayonet · LL · h 1,15 m



3 × ACE39060521
Pear handset · 2 m cable to the bayonet socket



1 × ACE82082137
Presence / cancel · next to the door



1 × EOP004
3-colour LED · above the door




2 × ECV169
PCB 24 pins · corridor ceiling

Mounting / wiring notes

- Radial wiring: each device has its own UTP Cat6
- cable to its PCB in the corridor ceiling.
- PCB A: bed calls 1-3 + presence + corridor light.
- PCB B: handset disconnection beds 1-3.
- Bus 2x2x0,8 mm daisy chain A → B → next room.
- Light switches 1,05 m (by others).

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. October 2026
	Reference.	9722600017_esq_REV03



Equipment on this sheet



3 × ACE82012153
Bed call panel + bayonet · LL · h 1,15 m



3 × ACE39060521
Pear handset · 2 m cable to the bayonet socket



1 × ACE82082137
Presence / cancel · next to the door



1 × EOP004
3-colour LED · above the door




2 × ECV169
PCB 24 pins · corridor ceiling

Mounting / wiring notes

- Radial wiring: each device has its own UTP Cat6
- cable to its PCB in the corridor ceiling.
- PCB A: bed calls 1-3 + presence + corridor light.
- PCB B: handset disconnection beds 1-3.
- Bus 2x2x0,8 mm daisy chain A → B → next room.
- Light switches 1,05 m (by others).

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. October 2026
	Reference.	9722600017_esq_REV03



Equipment on this sheet



3 × ACE82012153
Bed call panel + bayonet · LL · h 1,15 m



3 × ACE39060521
Pear handset · 2 m cable to the bayonet socket



1 × ACE82082137
Presence / cancel · next to the door



1 × EOP004
3-colour LED · above the door




2 × ECV169
PCB 24 pins · corridor ceiling

Mounting / wiring notes

- Radial wiring: each device has its own UTP Cat6
- cable to its PCB in the corridor ceiling.
- PCB A: bed calls 1-3 + presence + corridor light.
- PCB B: handset disconnection beds 1-3.
- Bus 2x2x0,8 mm daisy chain A → B → next room.
- Light switches 1,05 m (by others).

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. October 2026
	Reference.	9722600017_esq_REV03



Equipment on this sheet



3 × ACE40320303
SOS pull-cord · TI/TD · h 1,85 m



2 × ACE82082137
Presence / cancel · corridor, next to entrance



3 × EOP004
3-colour LED · above the entrance



3 × ECV169
PCB 24 pins · corridor ceiling

Mounting / wiring notes

- WC B3: PCB B3-1 = shower cord + P-B3 + LP-B3;
- PCB B3-2 = toilet cord + LP2-B3.
- WC B2: PCB B2 = toilet cord + P-B2 + LP-B2.
- Staff WC: no nurse call.
- Pull-cord at 1,85 m, cord down to 0,20 m from floor.
- Bus ends at PCB B2 (#11).

ADVANTECNIA ACE BG · Orad	Made by.	ACE BG · Advantecnia
	Revised by.	Technical dept. October 2026
	Reference.	9722600017_esq_REV03

Equipment on this sheet



1 × EOP002
Nurse iPad · Wi-Fi



1 × EVV038
ionTab licences + MDM

Mounting / wiring notes

- Nurse desk = iPad.
- Wi-Fi network by others.



((·))

wifi

Nurse iPad 10,9" · EOP002 · nurse station desk

Receives and resolves all calls · sets / cancels presence in any room