Dose-volume histogram — Lung_Patient_6

Clinician plan vs. GA (180° excluded). A curve further left and lower means less dose to that structure.



Curves further left are better for organs; the PTV curve should stay high, then drop sharply.
Target curves should overlap; organ curves show the tradeoffs. Esophagus is on page 2.

DVH by structure — Lung_Patient_6

Same curves, one structure per panel, all plans, each annotated with that structure's protocol metrics.
Bold = clinically preferable: organs favor lower dose; PTV favors the value closest to goal within the limit. Grey = all plans are within 0.01 of each other. Amber misses a goal,



Clinical criteria — Lung_Patient_6

Organs: lower is better. PTV: closest to goal without exceeding the limit. Bold marks the best plan.

Criterion	Limit	Goal	Clinician	GA 180° excluded	
GTV max	69	66	61.89	**61.86**	Gy
PTV max	69	66	**63.46**	63.15	Gy
ESOPHAGUS max	66	–	**11.41**	11.98	Gy
ESOPHAGUS mean	34	21	1.45	**0.36**	Gy
ESOPHAGUS V60Gy	17	–	0.00	0.00	%
HEART max	66	–	**43.46**	48.66	Gy
HEART mean	27	20	3.42	**3.25**	Gy
HEART V30Gy	50	48	0.28	**0.17**	%
LUNG_L max	66	–	**11.45**	26.36	Gy
LUNG_R max	66	–	63.46	**63.15**	Gy
CORD max	50	48	**7.40**	8.53	Gy
SKIN max	60	–	48.11	**45.65**	Gy
LUNGS_NOT_GTV max	66	–	63.46	**63.15**	Gy
LUNGS_NOT_GTV mean	21	20	**6.68**	6.69	Gy
LUNGS_NOT_GTV V20Gy	37	–	12.06	**11.56**	%
PTV D95 (target coverage)			59.76	59.74	
Objective value (search signal)			44.28	44.35	
Solve time			21 s	20 s	

Beam angles (gantry°)	Clinician	0, 185, 210, 240, 270, 300, 330
	GA	15, 30, 90, 185, 240, 285, 345

Grey rows: all plans agree within 0.01 — heart, lung and LUNGS_NOT_GTV max sit at 66 Gy because those structures abut the target. Amber misses the goal; red exceeds the limit.

Source: Lung_Patient_6_clinical_metrics_full_resolution.json, full resolution. The curves and table come from the same solve.