Dose-volume histogram — Lung_Patient_33

Clinician plan vs. GA (180° excluded). A curve further left and lower means less dose to that structure.



Curves further left are better for organs; the PTV curve should stay high, then drop sharply.
Target curves should overlap; organ curves show the tradeoffs. Esophagus is on page 2.

DVH by structure — Lung_Patient_33

Same curves, one structure per panel, all plans, each annotated with that structure's protocol metrics.
Bold = clinically preferable: organs favor lower dose; PTV favors the value closest to goal within the limit. Grey = all plans are within 0.01 of each other. Amber misses a goal,



Clinical criteria — Lung_Patient_33

Organs: lower is better. PTV: closest to goal without exceeding the limit. Bold marks the best plan.

Criterion	Limit	Goal	Clinician	GA 180° excluded	
GTV max	69	66	62.89	**62.68**	Gy
PTV max	69	66	**64.97**	63.61	Gy
ESOPHAGUS max	66	–	11.84	**5.68**	Gy
ESOPHAGUS mean	34	21	1.31	**0.61**	Gy
ESOPHAGUS V60Gy	17	–	0.00	0.00	%
HEART max	66	–	0.00	0.00	Gy
HEART mean	27	20	0.00	0.00	Gy
HEART V30Gy	50	48	0.00	0.00	%
LUNG_L max	66	–	**12.58**	30.11	Gy
LUNG_R max	66	–	64.97	**63.61**	Gy
CORD max	50	48	14.38	**9.52**	Gy
SKIN max	60	–	40.93	**39.93**	Gy
LUNGS_NOT_GTV max	66	–	64.97	**63.61**	Gy
LUNGS_NOT_GTV mean	21	20	**3.19**	3.45	Gy
LUNGS_NOT_GTV V20Gy	37	–	5.90	**5.21**	%
PTV D95 (target coverage)			59.60	59.67	
Objective value (search signal)			51.09	43.04	
MOSEK solve time			10 s	10 s	

```
Beam angles (gantry°)   Clinician  30°, 0°, 330°, 300°, 225°, 205°, 185°
                        GA         300°, 15°, 75°, 165°, 240°, 270°, 345°
```

Grey rows: all plans agree within 0.01 — heart, lung and LUNGS_NOT_GTV max sit at 66 Gy because those structures abut the target. Amber misses the goal; red exceeds the limit.

Source: Lung_Patient_33_clinical_metrics_full_resolution.json, full resolution. The curves and table come from the same solve.