



Dose-volume histogram — Lung_Patient_39



Curves further left are better for organs; the PTV curve should stay high, then drop sharply. Target curves should overlap; organ curves show the tradeoffs. Esophagus is on page 2.

DVH by structure — Lung_Patient_39

Same curves, one structure per panel, all plans, each annotated with that structure's protocol metrics.
Bold = clinically preferable: organs favor lower dose; PTV favors the value closest to goal within the limit. Grey = all plans are within 0.01 of each other. Amber misses a goal,



Clinical criteria — Lung_Patient_39

Organs: lower is better. PTV: closest to goal without exceeding the limit. Bold marks the best plan.

Criterion	Limit	Goal	Clinician	GA 180° excluded	
GTV max	69	66	64.34	62.95	Gy
PTV max	69	66	66.00	63.62	Gy
ESOPHAGUS max	66	—	62.17	62.10	Gy
ESOPHAGUS mean	34	21	10.58	8.90	Gy
ESOPHAGUS V60Gy	17	—	3.41	4.35	%
HEART max	66	—	0.00	0.00	Gy
HEART mean	27	20	0.00	0.00	Gy
HEART V30Gy	50	48	0.00	0.00	%
LUNG_L max	66	—	66.00	63.62	Gy
LUNG_R max	66	—	54.23	55.46	Gy
CORD max	50	48	30.30	15.49	Gy
SKIN max	60	—	46.86	47.89	Gy
LUNGS_NOT_GTV max	66	—	66.00	63.62	Gy
LUNGS_NOT_GTV mean	21	20	7.30	7.99	Gy
LUNGS_NOT_GTV V20Gy	37	—	13.77	15.71	%
PTV D95 (target coverage)			59.49	59.71	
Objective value (search signal)			90.00	69.05	
MOSEK solve time			30 s	26 s	
Beam angles (gantry°)	Clinician	175°, 150°, 124°, 65°, 32°, 0°			
	GA	65°, 0°, 165°, 210°, 255°, 285°, 315°			

Grey rows: all plans agree within 0.01 — heart, lung and LUNGS_NOT_GTV max sit at 66 Gy because those structures abut the target. Amber misses the goal; red exceeds the limit.

Source: Lung_Patient_39_clinical_metrics_full_resolution.json, full resolution. The curves and table come from the same solve.