



Dose-volume histogram — Lung_Patient_22



Curves further left are better for organs; the PTV curve should stay high, then drop sharply. Target curves should overlap; organ curves show the tradeoffs. Esophagus is on page 2.

DVH by structure — Lung_Patient_22

Same curves, one structure per panel, all plans, each annotated with that structure's protocol metrics.
Bold = clinically preferable: organs favor lower dose; PTV favors the value closest to goal within the limit. Grey = all plans are within 0.01 of each other. Amber misses a goal,



Clinical criteria — Lung_Patient_22

Organs: lower is better. PTV: closest to goal without exceeding the limit. Bold marks the best plan.

Criterion	Limit	Goal	Clinician	GA 180° excluded	
GTV max	69	66	62.13	61.69	Gy
PTV max	69	66	65.68	64.33	Gy
ESOPHAGUS max	66	—	24.99	34.06	Gy
ESOPHAGUS mean	34	21	3.65	2.90	Gy
ESOPHAGUS V60Gy	17	—	0.00	0.00	%
HEART max	66	—	4.46	3.22	Gy
HEART mean	27	20	0.01	0.01	Gy
HEART V30Gy	50	48	0.00	0.00	%
LUNG_L max	66	—	66.00	65.42	Gy
LUNG_R max	66	—	59.57	59.34	Gy
CORD max	50	48	14.01	14.98	Gy
SKIN max	60	—	50.34	48.76	Gy
LUNGS_NOT_GTV max	66	—	66.00	65.42	Gy
LUNGS_NOT_GTV mean	21	20	7.76	8.13	Gy
LUNGS_NOT_GTV V20Gy	37	—	15.91	14.89	%
PTV D95 (target coverage)			59.75	59.80	
Objective value (search signal)			52.76	48.17	
MOSEK solve time			29 s	30 s	
Beam angles (gantry°)	Clinician	175°, 160°, 140°, 60°, 25°, 0°, 330°			
	GA	0°, 15°, 75°, 210°, 285°, 315°, 345°			

Grey rows: all plans agree within 0.01 — heart, lung and LUNGS_NOT_GTV max sit at 66 Gy because those structures abut the target. Amber misses the goal; red exceeds the limit.

Source: Lung_Patient_22_clinical_metrics_full_resolution.json, full resolution. The curves and table come from the same solve.