


Dose-volume histogram — Lung_Patient_20


Clinician plan vs. GA (180° excluded). A curve further left and lower means less dose to that structure.



Curves further left are better for organs; the PTV curve should stay high, then drop sharply. Target curves should overlap; organ curves show the tradeoffs. Esophagus is on page 2.

DVH by structure — Lung_Patient_20

Same curves, one structure per panel, all plans, each annotated with that structure's protocol metrics.
Bold = clinically preferable: organs favor lower dose; PTV favors the value closest to goal within the limit. Grey = all plans are within 0.01 of each other. Amber misses a goal,



Clinical criteria — Lung_Patient_20

Organs: lower is better. PTV: closest to goal without exceeding the limit. Bold marks the best plan.

Criterion	Limit	Goal	Clinician	GA 180° excluded	
GTV max	69	66	62.84	62.40	Gy
PTV max	69	66	64.60	63.77	Gy
ESOPHAGUS max	66	—	62.24	61.73	Gy
ESOPHAGUS mean	34	21	13.92	12.91	Gy
ESOPHAGUS V60Gy	17	—	10.09	7.82	%
HEART max	66	—	66.00	62.30	Gy
HEART mean	27	20	9.13	8.31	Gy
HEART V30Gy	50	48	9.21	7.90	%
LUNG_L max	66	—	31.50	47.85	Gy
LUNG_R max	66	—	65.46	64.19	Gy
CORD max	50	48	50.00	39.14	Gy
SKIN max	60	—	48.64	45.41	Gy
LUNGS_NOT_GTV max	66	—	65.46	64.19	Gy
LUNGS_NOT_GTV mean	21	20	9.71	10.58	Gy
LUNGS_NOT_GTV V20Gy	37	—	19.13	19.83	%
PTV D95 (target coverage)			59.63	59.69	
Objective value (search signal)			93.36	84.60	
Pipeline time (setup + solve)			28 s	937 s	
Beam angles (gantry°)	Clinician	0°, 335°, 306°, 238°, 212°, 185°			
	GA	185°, 15°, 45°, 75°, 120°, 270°, 330°			

Grey rows: all plans agree within 0.01 — heart, lung and LUNGS_NOT_GTV max sit at 66 Gy because those structures abut the target. Amber misses the goal; red exceeds the limit.

Source: Lung_Patient_20_clinical_metrics_full_resolution.json, full resolution. The curves and table come from the same solve.