



# Dose-volume histogram — Lung\_Patient\_8



Curves further left are better for organs; the PTV curve should stay high, then drop sharply. Target curves should overlap; organ curves show the tradeoffs. Esophagus is on page 2.

# DVH by structure — Lung\_Patient\_8

Same curves, one structure per panel, all plans, each annotated with that structure's protocol metrics.  
Bold = clinically preferable: organs favor lower dose; PTV favors the value closest to goal within the limit. Grey = all plans are within 0.01 of each other. Amber misses a goal,



# Clinical criteria — Lung\_Patient\_8

Organs: lower is better. PTV: closest to goal without exceeding the limit. Bold marks the best plan.

| Criterion                       | Limit     | Goal                                   | Clinician | GA<br>180° excluded |    |
|---------------------------------|-----------|----------------------------------------|-----------|---------------------|----|
| GTV max                         | 69        | 66                                     | 66.98     | 64.39               | Gy |
| PTV max                         | 69        | 66                                     | 66.98     | 65.45               | Gy |
| ESOPHAGUS max                   | 66        | —                                      | 66.00     | 63.91               | Gy |
| ESOPHAGUS mean                  | 34        | 21                                     | 25.22     | 23.89               | Gy |
| ESOPHAGUS V60Gy                 | 17        | —                                      | 20.05     | 14.33               | %  |
| HEART max_dose                  | —         | —                                      | n/a       | n/a                 |    |
| HEART mean_dose                 | —         | —                                      | n/a       | n/a                 |    |
| HEART dose_volume_V             | —         | —                                      | n/a       | n/a                 |    |
| HEART dose_volume_V             | —         | —                                      | n/a       | n/a                 |    |
| LUNG_L max                      | 66        | —                                      | 43.86     | 41.71               | Gy |
| LUNG_R max                      | 66        | —                                      | 64.69     | 64.42               | Gy |
| CORD max                        | 50        | 48                                     | 38.78     | 27.16               | Gy |
| SKIN max                        | 60        | —                                      | 48.84     | 48.56               | Gy |
| LUNGS_NOT_GTV max               | 66        | —                                      | 64.69     | 64.42               | Gy |
| LUNGS_NOT_GTV mean              | 21        | 20                                     | 14.78     | 15.68               | Gy |
| LUNGS_NOT_GTV V20Gy             | 37        | —                                      | 25.06     | 27.08               | %  |
| PTV D95 (target coverage)       |           |                                        | 59.48     | 59.57               |    |
| Objective value (search signal) |           |                                        | 120.12    | 107.63              |    |
| Pipeline time (setup + solve)   |           |                                        | 66 s      | 65 s                |    |
| Beam angles (gantry°)           | Clinician | 0°, 185°, 210°, 240°, 270°, 300°, 330° |           |                     |    |
|                                 | GA        | 15°, 90°, 165°, 225°, 240°, 285°, 315° |           |                     |    |

Grey rows: all plans agree within 0.01 — heart, lung and LUNGS\_NOT\_GTV max sit at 66 Gy because those structures abut the target. Amber misses the goal; red exceeds the limit.

Source: Lung\_Patient\_8\_clinical\_metrics\_full\_resolution.json, full resolution. The curves and table come from the same solve.