

Certificate No. 34116/DBBPAM
Date: September 30, 2019



Issuing Office:
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia
Phone/Facs: +62 21 88321176/88321166
Email: jum.cbt@sucofindo.co.id

CERTIFICATE OF CALIBRATION

CLIENT : SENTRAL TEHNOLOGY MANAGEMEN, PT
ADDRESS : CIKARANG SQUARE BLOK B NO. 11
JL. RAYA CIBARUSAH CIKARANG SELATAN
CIKARANG - JAWA BARAT

NAME OF EQUIPMENT : ANAK TIMBANGAN

SERIAL NO./CODE : - / CSM-015

MATERIALS : STAINLESS STEEL

WEIGHT/CLASS : 200 g / F1


MANUFACTURER : EXCELLENT

The Attachment available is an integral part of this certificate.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at www.sucofindo.co.id

CBT20295119-002



Husen Muhajir



3328099

SCI-2007A

Amendment No. AM 135/DBBPAM
Date: October 14, 2019



Issuing Office:
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia
Phone/Facs: +62 21 88321176/88321166
Email: jum.cbt@sucofindo.co.id

**AMENDMENT
TO CERTIFICATE NO. 34116/DBBPAM DATE: SEPTEMBER 30, 2019**


| WRITTEN | SHOULD READ |
|------------------------------------------|------------------------------------------|
| CLIENT : SENTRAL TEHNOLOGY MANAGEMEN, PT | CLIENT : SENTRAL TEHNOLOGI MANAGEMEN, PT |

All other matters in the certificate remain unchanged.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at www.sucofindo.co.id

CBT20295119-002



Husen Muhajir



3346106

SCI-2007A

CERTIFICATE OF CALIBRATION

TESTING CONDITION,
- TEMPERATURE : $21.6 \pm 0.3^{\circ}\text{C}$
- HUMIDITY : $55 \pm 1\%$
STANDARD & EQUIPMENT : CALIBRATED MASSES CLASS E2
Serial No. 35329151
Certificate of Calibration No. 108353
: THERMOHYGROMETER
Serial No. MS-2
Certificate of Calibration Cal. 3292019
*The measurement is traceable to International System of Units (SI)
through P2M-LIPI*
CALIBRATION METHOD : PO / MS / 05
RECEIVED ON : September 19, 2019
DATE OF CALIBRATION : September 20, 2019

RESULTS :

ANAK TIMBANGAN Serial No / Code : CSM 015
WEIGHT / CLASS : 200 g / F1

| NOMINAL | CONVENTIONAL MASS | UNCERTAINTY OF MEASUREMENT at 95% Confidence Level |
|---------|-------------------|-------------------------------------------------------|
| g | g | mg |
| 200 | 200.00010 | 0.33 |



1537707