



# <u>Sertifikat Kalibrasi</u>

Certificate of Calibration

### No. TM.0158.18

Nomor Order Order Number

SO-01-18-000923

Tanggal Diterbitkan : 02 Agustus 2018

Issued Date

**IDENTITAS PEMILIK** Owner Identification

Pemilik

PT. SENTRAL TEHNOLOGI MANAGEMEN

Owner

Alamat Address

MT Haryono Square of 03A/02 Jl. MT. Haryono Kav.10 Jatinegara Jakarta Timur

DKI Jakarta Raya -13330

**IDENTITAS ALAT** 

Instrument Identification

Nama Alat

Tachometer

Tipe Model SE 300

Instrument

Merk / Pabrik

Nomor Seri

16025700128

Manufacturer

Serial Number

No. Identitas

**ID Number** 

**KST 002** 

Sanwa

**METODE & TELUSURAN** Method(s) & Traceability

Metode Kalibrasi

IK-KAL-TM 02

Calibration Method(s)

Telusuran

: SI melalui NMI Puslit KIM LIPI (Indonesia)

Traceability

### **KETERANGAN UMUM**

#### General Remark

Sistem Manajemen Mutu Laboratorium Kalibrasi, yang mana kalibrasi ini dilakukan, sesuai dengan ISO/IEC 17025:2005 "Persyaratan Umum Kompetensi Laboratorium Pengujian dan Laboratorium Kalibrasi"

The Calibration Laboratory's Management System for Quality, under which this calibration was performed, meets the ISO/IEC 17025:2005 "General Requirements for the Competence of Testing and Calibration Laboratories".

Ketidakpastian pengukuran pada kalibrasi ini didapat dengan menggunakan evaluasi Tipe-A dan Tipe-B, sesuai dengan JCGM 100 : 2008 (Guide to the Expression of Uncertainty in Measurement), dan diestimasi pada tingkat kepercayaan

The uncertainty(ies) of measurement associated in this calibration was/were obtained by Type-A and Type-B evaluation, according to JCGM 100: 2008 (Guide to the Expression of Uncertainty in Measurement), and estimated at a level of confidence of 95%.

**Authorized Signatory** 



halaman 1 dari 2 Page

04383 SF-024





SF-024

Tanggal Kalibrasi Calibration Date

: 01 Agustus 2018

Nomor Sertifikat

: TM.0158.18

Certificate Number

: 23.1 °C ± 1 °C

Lokasi Kalibrasi Calibration Location

: AEQUUS Solutions, PT. Mandiri Transforma Global

Kondisi Lingkungan Environment

43 % ± 3 % rh

**ALAT STANDAR** Standard(s) Used

1. Digital Stroboscope, E 016

## HASIL KALIBRASI Result of Calibration

Parameter:	Tachometer					
Range	Standard Value	Instrument Reading	Correction	Uncertainty		k
Auto	50.180 rpm	50.1 rpm	+0.1 rpm	± 0.057	rpm	2
	100.000 rpm	100.0 rpm	0.0 rpm	± 0.059	rpm	
	200.020 rpm	200.0 rpm	0.0 rpm	± 0.059	rpm	
	500.090 rpm	500.0 rpm	+0.1 rpm	± 0.12	rpm	
	1000.10 rpm	1000 rpm	+0.1 rpm	± 0.11	rpm	
	2001.10 rpm	2001 rpm	+0.4 rpm	± 1.5	rpm	
	4006.20 rpm	4004 rpm	+2.2 rpm	± 0.11	rpm	
	6010.70 rpm	6011 rpm	-0.3 rpm	± 0.25	rpm	
	7991.50 rpm	7993 rpm	-1.5 rpm	± 0.36	rpm	
	9996.10 rpm	10000 rpm	-3.9 rpm	± 0.47	rpm	
	11980.10 rpm	11984 rpm	-3.9 rpm	± 0.16	rpm	