

Certificate No. 37322/ENBPAO
Date: September 28, 2021



Issuing Office:
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia
Phone/Facs: +62 21 88321176/021 88321166
Email: cs.cbt@sucofindo.co.id

CERTIFICATE OF CALIBRATION

CLIENT : PT. SENTRAL TEKNOLOGI MANAGEMEN
ADDRESS : CIKARANG SQUARE BLOK B NO. 11,
JL. CIBARUSAH CIKARANG SELATAN

NAME OF EQUIPMENT : ANAK TIMBANGAN

SERIAL NO./ CODE : -

TYPE / MATERIALS : -/SS

WEIGHT/ CLASS : 2000 g / F1

MANUFACTURER : EXCELLENT SCALE


The Attachment available is an integral part of this certificate.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at www.sucofindo.co.id

38010521001227

Sub Dept. Calibration



Husen Muhajir



3937722

SCI-2007A

CERTIFICATE OF CALIBRATION

TESTING CONDITION,
- TEMPERATURE : ($19,2 \pm 0,1$) °C
- HUMIDITY : (47 ± 2) %
STANDARD & EQUIPMENT : ANAK TIMBANGAN
Serial No. 15849-1
Certificate of Calibration No. Cal. 5712019
: THERMOHYGROMETER
Serial No. MS-03
Certificate of Calibration No. Cal. 5922020
*The measurement is traceable to International System of Units (SI) through
SNSU-BSN*
CALIBRATION METHOD : PO / MS / 05
RECEIVED ON : September 20, 2021
DATE OF CALIBRATION : September 21, 2021

RESULTS :

NOMINAL	CONVENTIONAL MASS	UNCERTAINTY OF MEASUREMENT at 95% Confidence Level	k =
g	g	mg	
2000	2000,013	3,3	1,97



1893390