

Certificate No. 37319/ENBPAO  
Date: September 28, 2021



Issuing Office:  
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia  
Phone/Facs: +62 21 88321176/021 88321166  
Email: cs.cbt@sucofindo.co.id

## CERTIFICATE OF CALIBRATION

CLIENT : PT. SENTRAL TEHNOLOGI MANAGEMEN  
ADDRESS : CIKARANG SQUARE BLOK B NO. 11,  
JL. CIBARUSAH CIKARANG SELATAN

NAME OF EQUIPMENT : ANAK TIMBANGAN

SERIAL NO./ CODE : -

TYPE/ MODEL : -

MATERIALS : -

WEIGHT/ CLASS : 200 g / F1

MANUFACTURER : EXCELLENT

The Attachment available is an integral part of this certificate.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at [www.sucofindo.co.id](http://www.sucofindo.co.id)

38010521001227

Sub Dept. Calibration  
  
Husen Muhajir



3937719

SCI-2007A



Attachment  
To Certificate No. 37319/ENBPAO  
Date: September 28, 2021



Issuing Office:  
Phone/Facs: +62 21 88321176/021 88321166  
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia  
Email: cs.cbt@sucofindo.co.id


Page 2 of 2

## CERTIFICATE OF CALIBRATION

TESTING CONDITION,  
- TEMPERATURE : (  $19,8 \pm 0,3$  ) °C  
- HUMIDITY : (  $59 \pm 2$  ) %  
STANDARD & EQUIPMENT : CALIBRATED MASSES CLASS E2  
Serial No. 35329151  
Certificate of Calibration No. Cal. 5032020  
: THERMOHYGROMETER  
Serial No. MS-03  
Certificate of Calibration No. Cal. 5922020  
*The measurement is traceable to International System of Units (SI)  
through SNSU-BSN*  
CALIBRATION METHOD : PO / MS / 05  
RECEIVED ON : September 20, 2021  
DATE OF CALIBRATION : September 21, 2021

### RESULTS :

NOMINAL	CONVENTIONAL MASS	UNCERTAINTY OF MEASUREMENT at 95% Confidence Level	k =
g	g	mg	
200	199,99995	0,33	2,00



1893387

SCI-2007P