





Sertifikat Kalibrasi

CALIBRATION CERTIFICATE

F-C.02-02 (00)

Nomor / No: 59305/KAL/KLS/III/22 Halaman / Pages : Page 1 of 2

Identitas Alat / Identity of Instrument

Nama / Item : Ring Gauge
Jangkauan / Range : 14.000 mm
Pabrik / Factory : NSK
Tipe / Type : Nomor Seri / Serial No : TL0063
Resolusi / Resolution : -

Identitas Pemilik / Identity of Owner

Nama / Name : PT. SENTRAL TEHNOLOGI MANAGEMEN

Alamat / Address : MT HARYONO SQUARE OF 03A/02 13330 JAKARTA TIMUR

Tgl Terima Alat / Received : 01-Mar-2022 Tanggal Kalibrasi / Calibrated : 01-Mar-2022

Tempat Kalibrasi / Venue : Laboratorium Kalibrasi PT. Kawan Lama Sejahtera Cikarang

Suhu / Temperature : $(20 \pm 0.4)^{\circ}$ C Kelembaban / Humidity : (50.5 ± 2.3) % RH Alat standard kalibrasi / Standard used : ULM DMS* Metode Kalibrasi / Calibration Method : JIS B7420:1997

Ketertelusuran / Traceability : Tertelusur ke satuan SI melalui (traceable to SI through) SNSU-BSN

Catatan / Note : 114115 - SPK : WOC-S302-2203-000003

Hasil Kalibrasi / Calibration Result

Ketidakpastian kalibrasi untuk penunjukkan alat dilaporkan pada halaman ke 2. Ketidakpastian yang dilaporkan menyatakan expanded uncertainty yang dinyatakan pada tingkat kepercayaan sekitar 95% menggunakan faktor cakupan k = 2

Sertifikat ini merupakan satu kesatuan dan penggandaannya dinyatakan sah bila dilakukan secara lengkap

This Certificate shall not be reproduced except in full

Cikarang, 10 Mar 2022 Approved by:

EGIDIUS KELLSON Calibration Manager

Kal

Jl. Industri Selatan Blok PP No.4 Kawasan Industri Jababeka II - Cikarang, Bekasi 17550 - INDONESIA Tel.: (62-21) 8983 7500 Ext. 1116, Fax.: (62-21) 8983 7599

Email: labkalibrasickr@kawanlama.com, svcmkt@kawanlama.com, msservckr3@kawanlama.com







Laboratorium Kalibrasi

HASIL KALIBRASI

F-C.02-02 (00)

Halaman / Pages : Page 2 of 2

Calibration Result
No. 59305/KAL/KLS/III/22

Sumbu X

No.	Standard Value (mm)	Instrument Reading (mm)	Correction (mm)	U95 (mm)
1	14.0001	14.0000	0,0001	0.0036

Sumbu Y

No.	Standard Value (mm)	Instrument Reading (mm)	Correction (mm)	U95 (mm)
1	14.0001	14.0000	0.0001	0.0036

- Maximum Correction: 0.0001 mm

Confidence level 95%, k = 2.

(Checked by)

(Agus Nur F) Calibration Coordinator

=== Akhir Sertifikat ===

JI. Industri Selatan Blok PP No.4
Kawasan Industri Jababeka II - Cikarang, Bekasi 17550 - INDONESIA
Tel.: (62-21) 8983 7500 Ext. 1116, Fax.: (62-21) 8983 7599
Email: labkalibrasickr@kawanlama.com, www.kawanlama.com