



Certificate No. 33378/ENBPAO  
Date: September 1, 2021



Issuing Office:  
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia  
Phone/Facs: +62 21 88321176/021 88321166  
Email: cs.cbt@sucofindo.co.id

## CERTIFICATE OF CALIBRATION

CLIENT : PT. SENTRAL TEHNOLOGI MANAGEMEN  
ADDRESS : CIKARANG SQUARE BLOK B NO. 11  
JL. RAYA CIBARUSAH CIKARANG SELATAN  
JAWA BARAT 17530

NAME OF EQUIPMENT : CALIPER CHECKER

SERIAL NO./CODE : 130785

TYPE/MODEL : -

RANGE/RESOLUTION ( mm ) : 0 ~ 300


MANUFACTURER : METROLOGY-TAIWAN

The Attachment available is an integral part of this certificate.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at [www.sucofindo.co.id](http://www.sucofindo.co.id)

38010521001019



3880626

SCI-2007A



## CERTIFICATE OF CALIBRATION

TESTING CONDITION,  
- TEMPERATURE : ( 20.0 ± 0,4 ) °C  
- HUMIDITY : ( 50 ± 3 ) %  
STANDARD & EQUIPMENT : UNIVERSAL HORIZONTAL METROSCOPE  
Model ULM Opal-600 (S/N : 99112)  
Certificate of Calibration No. S. 20-0549  
: GAUGE BLOCK GRADE 0  
Serial No. 978412  
Certificate of Calibration No. S. 054203  
: GAUGE BLOCK SET M8  
Serial No. 16157  
Certificate of Calibration No. 19-0067  
*The measurement is traceable to International System of Units (SI) through  
SNSU-BSN & BBLM*  
CALIBRATION METHOD : PO/DM/12  
RECEIVED ON : August 12, 2021  
DATE OF CALIBRATION : August 30, 2021

### RESULT

Nominal Size	Deviation		Parallelism		Uncertainty of Measurement (at 95% CL,k=2)	
	Inside mm	Outside mm	Inside mm	Outside mm	Inside µm	Outside µm
25	-0.00383	0.00036	0.00031	0.00021	1.2	1.2
50	0.00659	0.00552	0.00007	0.00033	1.2	1.2
75	0.00555	-0.00470	0.00014	0.00045	1.2	1.2
100	0.00456	-0.00421	0.00035	0.00053	1.2	1.2
150	0.00097	-0.00421	0.00038	0.00053	1.2	1.2
200	-0.00049	-0.00463	0.00046	0.00070	1.2	1.2
250	0.00525	-0.00445	0.00015	0.00051	1.2	1.2
300	0.00345	-0.00463	0.00022	0.00052	1.2	1.2



1857950