

# JINGSTONE

PRECISION MEASUREMENT & CALIBRATION TECHNOLOGY INTERNATIONAL CONSULTANT CO., LTD.

NO. 168, GUANG-WU RD, KAOHSIUNG, TAIWAN, R.O.C. TEL : 886-7-3855577 FAX : 886-7-3852838

<http://www.metrology.com.tw>

E-mail: [jingstone@metrology.com.tw](mailto:jingstone@metrology.com.tw)

NO : JS-RD-R

## REPORT OF CALIBRATION

PAGE : 1 OF 1

APPLICANT :

COUNTRY :

EQUIPMENT

MANUFACTORY

GRADE.

SERIAL NO.

Gauge Block (Individual)

METROLOGY

Grade : 1

—

Temperature :  $20 \pm 1^{\circ}\text{C}$  Humidity :  $50 \pm 10\%$

DATE : 2020/02/06

### STANDARDS & EQUIPMENT

DESCRIPTION	MANUFACTURER	TRACEABLE NO.	EFFECTIVE TRACEABLE DATE
GAUGE BLOCK	METROLOGY/TAIWAN	TAF ( 0455 ) NO : S10607007	2020/08/03
GAUGE BLOCK COMPARATOR	TESA/SWISS	N.C.	

UNCERTAINTY OF MEASURING : BASED UPON 95% CONFIDENCE LIMITS :  $0.08 \mu\text{m} \sim 0.23 \mu\text{m}$

### CALIBRATION STATEMENT

- JINGSTONE CALIBRATION CENTER (JCC) certifies that the above pending calibrated instrument has been calibrated strict by means of standard employed with direct/ indirect compared method. Above standard and testing equipment are traceable to National Standards Calibration Laboratory - Taiwan(NML), and International Standard Calibration Laboratory - USA(NIST) , England(NPL) , Germany(PTB) etc.
- The management and technical operation in JINGSTONE and subsidiary calibration laboratory (JCC) are according to and also comply with quality standard requirement of ISO 9001& ISO/IEC 17025
- Special claim: this report is only valid for the above pending-calibrated instruments. Partly summary or report copied is invalid.
- Taiwan Accreditation Foundation Accredited Length Calibration Laboratory TAF NO.0193 ILAC-MRA

### CALIBRATION RESULT

Nominal Size (mm)	Series No.	Error ( $\mu\text{m}$ )
50	199149	-0.18
50	199441	0.00

