

CERN Summer Program Report

Ali Fele Paranj

August 21, 2019



Contents

1 How ALPHA Works	3
1.1 Magnets	4
1.2 Detectors	4
1.3 Penning Trap	5
2 My Projects	6
2.1 COMSOL Simulation	6
2.1.1 Short Beamlne	7
2.1.2 Long Beamlne	7
2.1.3 Real Beamlne	8
2.1.4 Buncher Simulation	8
2.2 LabVIEW Interface	9
2.2.1 Positron Source	9
2.2.2 Experiment Control	10
2.2.3 Virtual Instrument	11
2.3 Compact Rio Upgrade	12
2.3.1 PCB and Front Panel Design	12
2.4 Full Simulation	14
2.4.1 Geometry	14
2.4.2 Running Simulation on HTCondor Provided by LXPLUS Linux Cluster	15
3 Lectures, Workshops and Visits	16
3.1 Classroom Courses	16
3.2 Online Courses	18
3.3 Workshops and Visits	18

1 How ALPHA Works

ALPHA is an international collaboration based at CERN, and which is working with trapped antihydrogen atoms, the antimatter counterpart of the simplest atom, hydrogen. By precise comparisons of hydrogen and antihydrogen, the experiment hopes to study fundamental symmetries between matter and antimatter.



Figure 1: Antimatter Factory (left) and ALPHA experiment hall (right)

ALPHA has some important components that make the research on antimatters possible. These are: the Penning trap, which holds the positrons and antiprotons before we use them to make antihydrogen, the Atom trap, which traps and holds the antihydrogen atoms, and the Annihilation vertex imaging detector, which detects the antihydrogen atoms when we allow them to annihilate and can find the point at which they annihilated. You can see the full setup of experiment in figure 2 .




Figure 2: Full Setup of ALPHA experiment

Note: For pictures in these sections I have used pictures used in the web site of ALPHA

1.1 Magnets

The ALPHA magnetic trap is a variant of a type of atom trap called an 'Ioffe trap'. This magnetic trap is used to trap antihydrogen atoms which are neutral and electric fields can not be used to trap the particles. Such traps work because most atoms interact with a magnetic field through a property called their magnetic dipole moment. If the atom is moving in a magnetic field, it will gain and lose energy as the strength of the magnetic field near the atom changes. Making a magnetic field that increases in all directions from a central minimum point means that some atoms will gain potential energy and lose kinetic energy if they move away from the minimum. Atoms that have low enough total energy will convert all of their kinetic energy to potential energy and be reflected from the higher magnetic field and be trapped. You can think of this like a marble rolling in a bowl – a slow-moving marble can't reach the edge of the bowl and will be 'trapped' in the bowl.




Figure 3: ALPHA Magnetic Trap

1.2 Detectors

The ALPHA neutral trap is surrounded by a complex particle detector, called the Silicon Vertex Detector (SVD). The SVD could be described as a four-megapixel 3D camera, it 'sees' inside the ALPHA -apparatus and is sensitive enough to tell us where and when a single annihilation event occurs. In ALPHA, the antihydrogen atoms annihilate mainly at the gold-coated trap walls, but occasionally the annihilation can take place in the vacuum with the tiny amount of residual gas always present in the vacuum systems. The annihilating particles in the ALPHA trap are positron and antiproton. Positron, being a lepton, annihilates with its counterpart, electron, and produces two gamma rays. The annihilation of antiproton is a more complex event, but during the annihilation process, several energetic charged particles called pions are emitted. The pions

penetrate through the ALPHA -apparatus as well as the SVD, during which a tiny amount of energy is deposited into the three thin silicon sensor layers forming the SVD. The SVD records the locations of these interactions and, using this information, constructs the pion track (helix). As there are several of these tracks, the intersection of the tracks then gives the annihilation spatial location (vertex). In addition to the annihilation events, there is also cosmic muon background the SVD records. The fingerprint of these events, however, is very different from the annihilations and they can be effectively rejected.




Figure 4: Silicon Vertex Detector

1.3 Penning Trap

It is a basic and unavoidable fact in the antimatter business that to produce antihydrogen, antiprotons and positrons must be mixed. So, ALPHA must have the ability to confine and manipulate charged plasmas with reasonable efficiency and at cryogenic temperatures to boot! This is accomplished in ALPHA through the use of Penning traps, a type of trap commonly used in plasma physics experiments to confine charged plasmas. The charge is, in fact, the difference, and indeed the dilemma faced when attempting to trap antihydrogen. Because antihydrogen is neutral, it cannot be held in a traditional Penning trap. This is where ALPHAs unique magnetic trap comes in (see 1.1).

As for positron, antiproton and electron plasmas: a Penning trap will certainly do the trick. In a Penning trap, charged plasmas are confined in a superposition of magnetic and electric fields. You can find a great summary of fields used to trap the particles in figure 5




Figure 5: Penning Trap

2 My Projects

2.1 COMSOL Simulation

When the particles are released from positron source, they trap at the potential well of the accumulator. Then they release from the trap to travel toward the penning trap, which traps the positrons and antiprotons before mixing and making antihydrogen. positrons, on their way to reach the penning trap, expand because of the space charge. This will make the particles beam larger. So the efficiency of trapping positrons at the limited length of penning trap will decrease.

First, we decided to see this expansion on simulation. So I set up a simulation with COMSOL, using "charged particles tracer" module, and after running the simulation and getting the results, we could see that the standard deviation of the Z component of the position of particles increases as they travel toward penning trap. because of the long beamline, and the limitation of computational resources I set the simulation in three steps: Short beamline, Long beamline, Real beamline. in all of the simulations in this section, I have connected the buncher to the ground which has zero electric potential. Because in the first step we were interested in seeing the shape of the beam without bunching them. also, the electric potential on the electrodes of the accumulator was the same as are in a real experiment. you can see the electric potential calculated by simulation on the central axis of the beamline cylinder in the figure 6




Figure 6: Electric Potential in the Central axis of the beamline cylinder

2.1.1 Short Beamline

In this simulation, I set the length of the beamline to be about 3 meters which only covers the accumulator and buncher. The electric potential on the electrodes of the accumulator was exactly same as the real values used in the experiment, but for simplicity, I connected the buncher to the ground in order to have zero electrical potential on it. I was interested in evaluating the beam shape when they leave the buncher.

The results were quite wired. Because although we could see the expansion because of the particle-particle interaction and space charge, the results were showing a strange decrease at the standard deviation of the z position of the particles. and the depth of the well in the std plot was getting significant as the number of particles was increasing. You can find these results in the figure 7. So I decided to run more simulations with longer beamlines.

2.1.2 Long Beamline

In this part of the simulation, I increased the length of the beamline. the size of the beamline for these simulations was about 7 meters.

The significant result of this change was a decrease in the depth of the well at the $\text{std}(z)$, and the more significant result was the time of simulation that increased from 18 hours to 31 hours. It was because of enlarging the simulation domain.




Figure 7: Trajectory and Standard deviation of beam with 6000 Particles

finally, we noticed that the well was because of the surface that was closing the end of the beamline. Because as the particles were reaching there, they were sticking to the surface (this was because of the option that I set in preparing the simulation), so they couldn't go further away. This was resulting in a decrease in the standard deviation of the z position of particles. if this guess was true, we should see less decrease in the std on simulations with more longer beam line.




Figure 8: Trajectory and Standard deviation of beam with 6000 Particles in a beam line 6 meters long

2.1.3 Real Beamline

this simulatio was very computationally costly, which took about 23 hours to simulate just five particles. The reults was as we expected. There was no decrease in $\text{std}(z)$ of particles. So this simulation was colser to the real setup than perivous ones. So I tryed to simulate the more particles in this setup.




Figure 9: Trajectory and Standard deviation of beam with 6000 Particles in a beam line 11 meters long

2.1.4 Buncher Simulation

As I mentioned earlier, the beam of positrons expand as they travel toward penning trap and this expansion can be suppressed using buncher. We need




Figure 10: Electric Potential in the central axis of beam line. In this simulation buncher has $V = 10\sin(2\pi ft)$ potential in which $f = 10MHz$

And the result of the effect of the sine potential of particles is figure.




Figure 11: Trajectory and Standard deviation of beam with 100 Particles in a beam line 11 meters long. In this simulation buncher has $V = 100\sin(2\pi ft)$ potential in which $f = 1MHz$

2.2 LabVIEW Interface

Advanced and complicated Experiments like most of the experiments at cern, are impossible without using the advanced electronics for both collecting and analyzing data and controlling the experiment apartures. For these advanced experiments there are tons of parameters to be controlled, which are done in the control room (figure 12). So having a safe and graphical interface for this purpose has a key rule. In this project I designed a safe and graphical interface to control the positron source.




Figure 12: Control Room and Experiment Hall

2.2.1 Positron Source

ALPHA derives its positrons from a radioactive beta-decay source containing an isotope of sodium, Na-22. This isotope, which has a conveniently long half-life of about 2.6 years, emits positrons with a large spread of kinetic energies up to about 545 keV. Such energetic positrons cannot easily be applied for antihydrogen production, so that ALPHA uses a well established technique to produce a low energy (eV) beam of positrons in vacuum.

Positrons implanted into solid material typically have a lifetime less than one nanosecond, a thousand millionth of a second. However, during that brief time most will slow down by a variety of energy loss processes to reach kinetic energies close to those characteristic of the temperature of the solid. This process is termed moderation, as the positrons kinetic energy is lowered, or moderated. Whilst most of the positrons penetrate deep into the bulk of the material and annihilate there, about 1 percent stop close enough to the surface that they can diffuse back to it before they annihilate. Incredibly, most of the positrons which reach the surface are emitted into vacuum at low energy, and can be readily formed into a beam and transported, typically using magnetic guiding fields. ALPHA uses a solid film of condensed neon as its moderator; this is one of the most efficient positron moderators.




Figure 13: Positron Source and Accumulator

2.2.2 Experiment Control

The computers that controls the different parameters of setup are held by [so-toons] held on the control platforms (figure 14).

All of this computers has special cards installed on their motherboards (for example NI PCI-6229 for analog inputs, NI PCI-6713 for analog outputs, NI PCI-8431 for RS 485 communications, NI PCI-8430 for RS 232 communication and many other cards). The user can control the experiment through these cards. The analog and digital signals which indicates the state of individual parts on experiment setup, are collected by the input gates. Then by evaluating the signals in control room proper signal is send by output gates to the experiment




Figure 14: Control Platform

components (like vacuum pumps, valves, current of magnets, electric potential on electrodes, etc).




Figure 15: PCI 8430 card (left), and Computers on control platform equipped by DAQ cards (right)

2.2.3 Virtual Instrument

The Virtual Interface of VI that I have designed will control the valves, magnets, and vacuum pumps that are connected to the positron source and accumulator. You can see the front panel or the GUI of VI[figure??].

The block diagram that controls this VI consists of three different sub VIs. I have designed block diagram to be modular and my supervisor can add other control and options after my departure. each of these sub VIs is for Valves,

Vacuum pumps and pass camra section. These VIs must feed by a number that indicates the state of valve or vacuum pump to be On or Off, and by a string that the Icon of the elnts in GUI are in it. The sub VI will brrows the proper Icon considering the numeric input and will send the picture as Output. all of this elements are in a event structure to reduce the amount of cpu uccpation. the reduction of CPU use is because the while loop around the event structure will run for 1 time, just when you changfe the values of the each event structure.

2.3 Compact Rio Upgrade

As mentioned before, handling advanced experiments requires advanced electronic setups. As the time passes the electronics and computers get more powerful. So one of the important 'Must do's for every experiment is upgrading the components of experiment to the last technologies. ALPHA was born at 2009[????] and since that day many updates had been done on setup and computers. A part of new update is transferring the old computers with DAQ cards to professional Compact Rio computers that are designed mainly for Experimental and data acquisition purposes. In more details, CompactRIO (or cRIO) is a real-time embedded industrial controller made by National Instruments for industrial control systems. The CompactRIO is a combination of a real-time controller, reconfigurable IO Modules (RIO), FPGA module and an Ethernet expansion chassis.[Wikipedia of cRIO](figure 16). In these computers we use DAQ modules instead of DAQ cards. This will make the control platform very compact and of course more professional. In some of the modules that we use with this computers we need to separate the data lines in order to use them individually. for this purpose we need to design a printed circuit board (PCB), that maps every pin on D-subminiature connector to individual lemo connectors.




Figure 16: Compact Rio

2.3.1 PCB and Front Panel Design

In this project I designed the PCB using "Altium design' software. these PCBs connect the D-subminator connector to individual Lemo Connectors. Throughout this report I will call these PCBs "Data Line Separator" or DLS. We had four different types of modules, so I needed to design four different PCBs. The modules were NI9264, NI9401, NI9403, NI9205. these modules are digital and analog modules that are connected to compact rio system to control the exper-



Figure 17: (left to right) NI 9205, NI 9264, NI 9401, NI 9403

iment aparues. For example one can control valvs, vacuum pumps, mass flow rate controlers, etc. These are done by using LabVIEW interface which I worked with and designed an interface in other project that I will descripe later. You can find mentioned modules in figure ??

As I said before, for complicated experiments there is a control platform that contols the parameters of experiment. For example figure ??? is the control platform of APLHA experiment. These platforms are where that the Compact Rio will sit at. Each compact rio will be held at chassis and these chassises will be held horizontally on top of each other by [sotoon]. so to keep every thing not messy, we need to design fron panels that can mount on chassies, and screw the PCBs to them. I did so for each of the modules and you can see the front panel of the NI9205 module in figure ??. You can find full documents at my github repository for cern as well.





Figure 18: First: DLS designed for NI9205 module, Second: 3D model of DLS, Third: Front Panel Designed for NI9205 DLS

2.4 Full Simulation

Simple simulations are very useful when you want to evaluate your stimates using the minimum computation resources. But after finding the rough simulations satisfactory, you need to improve you simulated model and add more details to it. In the simple simualtion of buncher we find out that buncher works pretty well in bunching the expanding beam. So we decided to add more details to our simulation and simulate the full setup of magnets and beam line.

2.4.1 Geometry

To make it possible, we needed to make a CAD model of magents in full setup. The data of magnets was in a file with .cond type which was designed to feed the data to the OPERA software to calculate the magnetic field. After finding the meaning of those numbers on .cond file with the help of my supervisor ,First I wrote a python code to transfer those data to a more clear data set of magnets[figure of excel]. Then I used the Inventor software to design all of the 82 magnets that contributes in the magnetic field of beam line. Then I added the Inventor model of buncher and Accumulator(that was previuosly created by ALPHA group).



```

Activities Terminal Tue 15:25
afelepar@lxplus774:~$ ssh -XY afelepar@lxplus
Warning: Permanently added the ECDSA host key for IP address '188.184.30.79' to the list of known hosts.
Password:
* ****
* Welcome to lxplus774.cern.ch, CentOS, 7.6.1810
* Archiving: /var/www/html/lxplus/doc/soft-archive
* Remember: you have agreed to the CERN
* Computing Rules, see https://cern.ch/computing-rules
* the measures necessary to ensure compliance.
* https://cern.ch/ComputingRules
* Puppet environment: production, Roger state: production
* Foreman hostgroup: lxplus/nodes/login
* Available login zones: lxplus
* lxplus Public login Service: https://cern.ch/lxplusdoc
* lxplus alias switch to C7 on Apr 2nd 2019 .. http://cern.ch/gK7lq
* ****
afelepar@lxplus774 ~]$ condor_q

-- Schedd: blgbird09.cern.ch : <188.185.71.142:9618>... @ 08/20/19 15:20:58
OWNER BATCH_NAME SUBMITTED DONE RUN IDLE TOTAL JOB_IDS
afelepar CMD: my_script 8/20 12:35 - 1 - 1 1928188.0

1 jobs; 0 completed, 0 removed, 0 idle, 1 running, 0 held, 0 suspended
afelepar@lxplus774 ~]$ condor_status

```

Figure 19: SSH tunnel to lxplus cluster through linux terminal on my laptop

2.4.2 Running Simulation on HTCondor Provided by LXPLUS Linux Cluster

Althogh I was using two computers in parallel to generate the results faster, but Since Simulating the full setup was very computational expensive, So it impossible to simulate the whole setup with 32 GB of RAM. The time of simulation was very long as well, so we needed to run our simulation on HTcondor. HTcondor stands for High Throughput Computing and is a specialized workload management system for compute-intensive jobs. Like other full-featured batch systems, HTCondor provides a job queueing mechanism, scheduling policy, priority scheme, resource monitoring, and resource management. Users submit their serial or parallel jobs to HTCondor, HTCondor places them into a queue, chooses when and where to run the jobs based upon a policy, carefully monitors their progress, and ultimately informs the user upon completion[<https://research.cs.wisc.edu/htcondor/description.html>]. linux clusters of cern provides the HTcondor computation and you can access big memory nodes and higher number of cpu cores. But since I was here yat cern for 53 days, it was impossible to run all of the simulations on clusters and get results. So in this part of my project I set up the full simulation file with the real initial values used in apartues .since I was familiar with Linux I could set up the environment and submit some sample simularions [figure lxplus] and my supervisor will run the simulation on cluster after my departure. This simulation contains a parametric sweep to search the optimum values for "amplitude", "Frequency" and "phase" of the the sine potential that is applied on the buncher. The real Quantitive results will be generated with this simulation and it will help us to tune the paramters of bunher on order to bunch the positron beam.

3 Lectures, Workshops and Visits

During my stay at cern I attended almost all of the Lectures. This lecture series was one of the best ones I attended during my life. The content of lectures was very interesting, the presentation of the lecturers was excellent and the lectures was very up to date. Here is the Courses that I attended and the "Physics and Medical Applications" I think was the best among them which was the most relevant subject to my field of studies.

3.1 Classroom Courses

1. Physics and Medical Applications by Manjy Dosanjih
2. Particle World by Tara Shears
3. Detectors by Werner Riegler
4. Foundation of Statistics by Nicolas Berger
5. Electronics DAQ and Trigger
6. Theoretical Concepts in Particle Physics by Andrew Cohen
7. From Raw Data to Physics Results by Paul James
8. Experimental Physics at Hadron Colliders by Marumi Kado
9. The Standard Model
10. Astroparticle Physics
11. Heavy Ions
12. Introduction to Cosmology
13. Beyond Standard Model
14. Nuclear Physics at CERN
15. What is String Theory
16. Future High-Energy Collider Projects
17. Antimatter at Lab

3.2 Online Courses

I attended some Online safety courses which were as following :

1. Computer Security
2. Emergency Evacuation
3. Radiation Protection
4. Electrical Safety Fundamentals
5. Electrical Safety Facilities
6. Cryogenic Safety Awareness
7. Chemical Safety Awareness

3.3 Workshops and Visits

During the program there were many cool workshops and visits. Here is a list of items that I attended

1. Open Data in Educational Activities Workshop
2. Basic Cloud Chamber Workshop
3. Advanced Cloud Chamber Workshop
4. Data Acquisition / Trigger Workshop
5. Root Summer Student Workshop
6. Silicon Sensors
7. CMS visit