Supplement Material 4. A summary of the parameters used for speech intelligibility testing

Studies, and acclimatization periods	Prescription formulae	Condition	Stimulus	Presentation level	Scoring technique
Boymans and Dreschler (2012), acclimatized for 6 weeks	Different prescriptions	Speech in quiet	VU 98 sentences Uttered by female speaker	At the participants' hearing thresholds	--
		Speech in noise	VU 98 sentences Uttered by male and female speakers Continuous noise, time-reversed speech spatially combined and separated	Noise was fixed at 65 dB A	--
Karimi et al. (2016), No acclimatization	Same prescription	Speech in quiet	Words (CVC)	At 70 dB SPL	Word scoring
Chang et al. (2018), No acclimatization	Same prescription	Speech in quiet	Sentences	At the participants' hearing thresholds	Sentence
		Speech in noise	Sentences	Noise presented from the same angle at 65 dB A	Sentence
Valente et al. (2018), , Participants acclimatized for 4 weeks	Different prescriptions	Speech in quiet	CNC words	At 50, 65 and 80 dB SPL	Word and phoneme scoring
		Speech in noise	Sentences	Noise presented from eight loudspeakers at 65 dB A	Sentence
Denys et al. (2019), Participants acclimatized for 2 weeks	Same prescription	Speech in quiet	NVA word list (CVC) Uttered by a male speaker	At the participants' hearing thresholds	Phoneme scoring
		Speech in noise	Sentences Uttered by a female speaker	Noise was fixed at 65 dB A	Key word scoring

Note: CNC = Consonant-Nucleus-Consonant; dB = decibel; SPL = sound pressure level; dBA = A-weighted; CVC = consonant vowel consonant.