Supplementary Table 1. Meta-Regression for DBT for the Diagnosis of Breast Cancer

Covariate	No. of Studies	Pooled Summary Estimates		
		Sensitivity (95% CI)	Specificity (95% CI)	*P**
Modality				0.190
DBT alone	11	0.93 (0.89–0.96)	0.90 (0.84–0.97)	
DBT combined with DM	9	0.86 (0.81–0.92)	0.89 (0.82–0.97)	
Country				0.400
Asian	10	0.92 (0.88–0.96)	0.92 (0.86–0.97)	
Non-Asian	10	0.88 (0.83–0.93)	0.88 (0.79–0.96)	
Existence of calcifications				0.950
Yes	15	0.90 (0.86–0.93)	0.90 (0.85–0.96)	
No	5	0.90 (0.81–0.99)	0.90 (0.79–1.00)	
Breast density†				0.900
≥ 2 or b	5	0.90 (0.80–1.00)	0.92 (0.84–1.00)	
All (1–4 or a–d)	15	0.90 (0.86–0.93)	0.89 (0.83–0.95)	
BI-RADS threshold				0.010
≥ 4	15	0.91 (0.87–0.94)	0.90 (0.84–0.96)	
≥ 3	3	0.90 (0.82–0.94)	0.95 (0.87–1.00)	
Study design				0.760
Prospective	4	0.92 (0.86–0.98)	0.89 (0.78–1.00)	
Retrospective	16	0.89 (0.85–0.93)	0.90 (0.85–0.96)	
Sampling				0.920
Consecutive	9	0.89 (0.84–0.94)	0.89 (0.82–0.97)	
Non-consecutive	11	0.90 (0.86–0.95)	0.89 (0.82–0.97)	
Sample sizes				0.990
≥ 200	10	0.90 (0.85–0.94)	0.90 (0.83–0.97)	
< 200	10	0.90 (0.86–0.95)	0.90 (0.83–0.97)	
Indications for diagnostic examination				0.230
Symptom only	3	0.93 (0.87–0.99)	0.95 (0.89–1.00)	
Symptom or screening recall	17	0.89 (0.85–0.93)	0.89 (0.83–0.95)	
No. of readers				0.160
≥ 5	7	0.87 (0.82–0.92)	0.83 (0.71–0.95)	
< 5	13	0.91 (0.87–0.95)	0.93 (0.88–0.97)	

**p* value in sensitivity and/or specificity between subgroup, †Breast density patterns were described in either way of using composition 1, 2, 3, 4 or a, b, c, d. BI-RADS = Breast Imaging Reporting and Data System, CI = confidence interval, DBT = digital breast tomosynthesis, DM = digital mammography